



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

March 30, 2016

Medtronic, Inc.  
Cheryl Swanson  
Sr. Principal Regulatory Affairs Specialist  
8200 Coral Sea Street NE  
Mounds View, Minnesota 55112

Re: K153160

Trade/Device Name: AVIVO™ Mobile Patient Management (MPM) System  
Regulation Number: 21 CFR 870.1025  
Regulation Name: Arrhythmia Detector and Alarm (Including ST-Segment Measurement and Alarm)  
Regulatory Class: Class II  
Product Code: MHX, DSB, DSI  
Dated: February 29, 2016  
Received: March 1, 2016

Dear Cheryl Swanson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman", is written over a faint, light-colored FDA logo watermark.

for Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: January 31, 2017  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)  
K153160

Device Name  
AVIVO™ ® Mobile Patient Management (MPM) System

### Indications for Use (Describe)

The AVIVO™ Mobile Patient Management (MPM) System is intended to continuously measure, record and periodically transmit physiological data. The AVIVO™ MPM System is indicated for patients:

- With fluid management problems
- Taking diuretic medication
- Living with heart failure
- Living with end-stage renal disease
- Suffering from recurrent dehydration
- Who require monitoring for the detection of non-lethal cardiac arrhythmias such as, but not limited to, supraventricular tachycardias (e.g. atrial fibrillation, atrial flutter, paroxysmal SVTs), ventricular ectopy, bradyarrhythmias and conduction disorders.

The AVIVO™ MPM System monitors, derives and displays:

- ECG
- Heart Rate (including HR variability)
- Activity
- Posture
- Respiration Rate (including RR variability)
- Body Fluid Status

The system has not been tested specifically for pediatric use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) SUMMARY

The following table provides background information regarding this Special 510(k) submission:

<b>Date Prepared:</b>	October 30, 2015
<b>510(k) Owner / Address:</b>	Medtronic, Inc. Cardiac Rhythm and Heart Failure 8200 Coral Sea Street NE Mounds View, MN 55112
<b>Contact Person:</b>	Cheryl Swanson Senior Principal Regulatory Affairs Specialist <a href="mailto:cheryl.swanson@medtronic.com">cheryl.swanson@medtronic.com</a> Office: (763) 514-0088 Cell: (651) 242-3506  <u>Secondary Contact:</u> Ryan Calabrese Regulatory Affairs Director <a href="mailto:ryan.s.calabrese@medtronic.com">ryan.s.calabrese@medtronic.com</a> Office: 763-526-3515
<b>Submission Type:</b>	Special 510(k): Device Modification
<b>Device Trade Name:</b>	AVIVO™ Mobile Patient Management (MPM) System
<b>Device Common Name:</b>	Mobile Patient Management (MPM) System
<b>Product Code and Classification Regulation Name:</b>	DSB: Plethysmograph, Impedence DSI: Detector and Alarm, Arrhythmia MHX: Monitor, Physiological, Patient (With Arrhythmia Detection or Alarms)
<b>Predicate Devices:</b>	<b><u>Primary</u></b> K113187, cleared January 4, 2012 (AVIVO™ MPM) <b><u>Secondary</u></b> K113372, cleared March 7, 2012 (NUVANT™ MCT) K133701, cleared February 6, 2014 (NUVANT™ MCT) - <b>rebranded as SEEQ™ MCT</b>

## Device Description

The AVIVO™ MPM System is a wearable, wireless physiological monitoring and arrhythmia detection system that is used by patients to aid clinicians in the identification, diagnosis and management of various clinical conditions, events and/or trends. It consists primarily of the Wearable Sensor (monitoring device) and the Transmitter (portable data transmission device). In combination with interpretation services provided by Medtronic Monitoring, Inc.'s Monitoring Center, as well as secure online review of data by healthcare providers, the AVIVO™ MPM System enables patient- and physician-friendly physiological monitoring and arrhythmia detection for extended periods of time.

## Significant Physical and Performance Characteristics of the Device

The AVIVO™ MPM System is comprised of the following non-sterile components:

- The Wearable Sensor, which adheres to the patient's torso, contains electrodes and sensors for recording patient information.
- The Transmitter, which is the patient hand-held transceiver, receives information from the Wearable Sensor and transmits it to the Medtronic Server.
- The secure Server, which receives information from the Wearable Sensor via the Transmitter, and among other things, derives, calculates and displays the patient's physiological parameters using the data collected by the Wearable Sensor.

**This remains unchanged from the predicate device.**

## Indications for Use Statement

The AVIVO™ Mobile Patient Management (MPM) System is intended to continuously measure, record and periodically transmit physiological data. The AVIVO™ MPM System is indicated for patients:

- With fluid management problems
- Taking diuretic medication
- Living with heart failure
- Living with end-stage renal disease
- Suffering from recurrent dehydration
- Who require monitoring for the detection of non-lethal cardiac arrhythmias such as, but not limited to, supraventricular tachycardias (e.g. atrial fibrillation, atrial flutter, paroxysmal SVTs), ventricular ectopy, bradyarrhythmias and conduction disorders.

The AVIVO™ MPM System monitors, derives and displays:

- ECG
- Heart Rate (including HR variability)
- Activity
- Posture

- Respiration Rate (including RR variability)
- Body Fluid Status

The system has not been tested specifically for pediatric use.

## Intended Use of the Device

The AVIVO™ MPM System is intended for the ambulatory recording and monitoring of physiological parameter(s).

**This remains unchanged from the predicate device.**

## Comparison of Fundamental Scientific Technology with the Predicate Devices

The subject AVIVO™ MPM System operates the same as the predicates, based on the following fundamental scientific technology:

- The collection of physiological parameters by a multi-sensor patient-worn device (Wearable Sensor);
- The transmission of these parameters to a remote Server through a transceiver (Transmitter); and
- The receipt of the parameters by the Server and subsequent derivation into appropriate useful values for display.

**This remains unchanged from the predicate device.**

## Summary of Testing and Performance Data to Demonstrate Substantial Equivalence

The AVIVO™ MPM System is supported by the successful completion of the following tests to demonstrate substantial equivalence.

**Table 4: Summary of Completed Testing to Support this Submission**

Section Description	Modification 1	Modification 2	Modification 3
<b>Sterilization and Shelf Life</b>	N/A – no changes to sterilization or shelf life related to this modification	Testing for this change already provided in K133701	N/A – no changes to sterilization or shelf life related to this modification
<b>Biocompatibility</b>	N/A – no changes to biocompatibility related to this modification	Although there were no changes to the patient-contacting materials related to this modification, biocompatibility testing was repeated	N/A – no changes to biocompatibility related to this modification

Section Description	Modification 1	Modification 2	Modification 3
		and results provided in K133701	
<b>Software</b>	Testing for this modification already provided in K113372	Testing for this modification already provided in K133701	Discussed in the <b>Software Section</b>
<b>EMC Testing</b>	N/A – no EMC changes associated with this modification	Testing for this modification already provided in K133701	N/A – no EMC changes associated with this modification
<b>Performance Testing – Bench Mechanical Electrical</b>	N/A – no mechanical or electrical changes associated with this modification	Testing for this change already provided in K133701	N/A – no mechanical or electrical changes associated with this modification
<b>Performance Testing – Animal</b>	N/A – no animal testing required for this modification	N/A – no animal testing required for this modification	N/A – no animal testing required for this modification
<b>Performance Testing – Clinical</b>	N/A – no clinical testing required for this modification	N/A – no clinical testing required for this modification	N/A – no clinical testing required for this modification
<b>Cybersecurity</b>	N/A – no cybersecurity changes associated with this modification	N/A – no cybersecurity changes associated with this modification	N/A – no cybersecurity changes associated with this modification.  Note: This section is included due to the release of the final FDA guidance document since the last submission. See the <b>Cybersecurity Section</b> for details.
<b>RF &amp; Wireless Technology</b>	N/A – no RF & Wireless changes associated with this modification.	N/A – no RF or Wireless changes associated with this modification.  Note: following this submission, the Transmitter's cellular technology was updated. As such, RF & Wireless	N/A – no RF & Wireless changes associated with this modification.

Section Description	Modification 1	Modification 2	Modification 3
		Technology is discussed in the <b>RF &amp; Wireless Technology Section</b>	

System validation testing was also successfully completed to demonstrate that the modified AVIVO™ MPM System functions as expected and that the system meets user needs and intended use.

### Guidance Used to Demonstrate Substantial Equivalence

Guidance documents used/considered for these device modifications include, but are not limited to, the following:

- Draft Guidance Document titled Use of International Standard ISO 10993, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing, issued on April 23, 2013
- Design Considerations for Devices Intended for Home Use, issued on November 24, 2014
- Radio Frequency Wireless Technology in Medical Devices, issued on August 13, 2013
- Content of Premarket Submissions for Management of Cybersecurity in Medical Devices, issued on October 2, 2014
- Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices, issued May 11, 2005
- Class II Special Controls Guidance Document: Arrhythmia Detector and Alarm dated October 28, 2003

### Standards Used to Demonstrate Substantial Equivalence

Table 5 identifies the standards that were also used, in whole or in part, to demonstrate substantial equivalence:

**Table 5: Standards to Demonstrate Substantial Equivalence**

Standards Organization / Number	Standards Title	Date / Version
IEC 60601-1	Medical electrical equipment – Part 1: General requirements for basic safety and essential performance	2005 + CORR 1 (2006) + CORR 2 (2007) / 3 <sup>rd</sup> Edition

Standards Organization / Number	Standards Title	Date / Version
IEC 60601-1-2	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	2007 / 3 <sup>rd</sup> Edition
IEC 60601-1-6	Medical electrical equipment – Part 1-6: General Requirements for Basic Safety and Essential Performance – Collateral Standard: Usability	2010-01 / 3 <sup>rd</sup> Edition
IEC 60601-1-11	Medical electrical equipment – Part 1-11: General Requirements for Basic Safety and Essential Performance – Collateral Standard Requirements for Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment	2010-04 / 1 <sup>st</sup> Edition
ANSI/AAMI/IEC 60601-2-47	Medical electrical equipment - Part 2-47: Particular requirements for the safety and essential performance of ambulatory electrocardiographic systems	2012 / 2 <sup>nd</sup> Edition
ANSI/AAMI/IEC 62304	Medical Device Software – Software life cycle processes (Edition 1)	2006-005
AAMI/ANSI EC 12	Disposable ECG electrodes	2000 (R) 2010
AAMI/ANSI EC 57	Testing and reporting performance results of cardiac rhythm and ST-segment measurement algorithms	2012
AAMI/ANSI/ISO 10993-1	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process	2009

Standards Organization / Number	Standards Title	Date / Version
AAMI/ANSI/ISO 10993-5	Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity	2009
AAMI/ANSI/ISO 10993-10	Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization	2010
ISO 14971	Medical Devices – Application of Risk Management to Medical Devices (Edition 2)	2007-03-01; 2012

## Conclusion

Medtronic has demonstrated that the AVIVO™ MPM System, as described in this submission, results in a substantially equivalent device because the fundamental scientific technology, operating principles and intended use are unchanged from the predicate devices. Summary data has been provided to demonstrate reasonable assurance of safety and effectiveness of the AVIVO™ MPM System and to demonstrate substantial equivalence to its predicates. As supported by the descriptive information, verification, validation and standards testing, the modified AVIVO™ MPM System is substantially equivalent to the predicate devices.