



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

Tyber Medical LLC  
Mr. Mark F. Schenk  
Director of QA/RA  
89 Headquarters Plaza North, #1464  
Morristown, New Jersey 07960

November 25, 2015

Re: K153180

Trade/Device Name: Tyber Medical MST Trauma Screw  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or threaded metallic bone fixation fastener  
Regulatory Class: Class II  
Product Code: HWC  
Dated: November 3, 2015  
Received: November 3, 2015

Dear Mr. Schenk:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director,  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

## Indications for Use

Form Approved: OMB No. 0910-0120

Expiration Date: January 31, 2017

See PRA Statement below.

510(k) Number (if known)

K153180

Device Name

Tyber Medical MST Trauma Screw

Indications for Use (Describe)

The Tyber Medical MST Trauma Screws are indicated for use in bone reconstruction, osteotomy, arthrodesis, joint fusion, fracture repair, and fracture fixation of bones appropriate for the size of the device. Screws are intended for single use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Special 510(k) Summary**

as required by section 807.92(c).

**Tyber Medical  
MST Trauma Screw****K 153180**

Updated	11/23/15
Submitter:	Tyber Medical LLC 89 Headquarters Plaza North, #1464 Morristown, New Jersey 07960
Contact Person	Mark F Schenk Director of QA/RA Phone: (610) 507-8255 Fax: (866) 889-9914 Email: mschenk@tybermed.com
Trade Name	Tyber Medical MST Trauma Screw
Common Name	Bone Compression Screw
Device Class	Class II
Classification Name and Number	Smooth or threaded metallic bone fixation fastener 21 CFR 888.3040
Classification Panel:	Orthopedic
Product Code	HWC
Predicate Devices	Tyber Medical Trauma Screw – K133842
Device Description	General trauma screw for compression and fixation of bone with modified surface treatment (MST). The screws will be provided sterile in both solid and cannulated form, made of from titanium.

Intended Use	A trauma screw designed to apply compression and fixation between two adjacent segments of cortical and/or calcaneus bone.
Indications for Use	The Tyber Medical MST Trauma Screws are indicated for use in bone reconstruction, osteotomy, arthrodesis, joint fusion, fracture repair, and fracture fixation of bones appropriate for the size of the device. Screws are intended for single use only.

Statement of Technological Comparison and Fundamental Scientific Technology	Tyber Medical MST Trauma Screw and its predicate device have the same indications for use, similar design, similar materials, technology principles of operation and test results.
---	--

Nonclinical Testing Summary	<p>The following tests were performed to demonstrate that the Tyber Medical MST Trauma Screw is substantially equivalent to other predicate device.</p> <ul style="list-style-type: none"> <li>• Pull out Test per ASTM F543</li> <li>• Static Torsion Test per ASTM F543</li> <li>• Biocompatibility evaluation including Cytotoxicity testing per ISO 10993-5.</li> <li>• X-ray photoelectron spectroscopy (XPS) was performed to verify the MST Trauma Screw have the same elemental composition as the predicate.</li> </ul> <p>The results of these studies showed that the Tyber Medical MST Trauma Screw met the acceptance criteria.</p>
Clinical Test Summary	n/a

Conclusion	Tyber Medical MST Trauma Screw and its predicate device have the same indications for use, similar design, and test results. Both devices are manufactured using materials with a long history of use in orthopedic implants.
------------	---