



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

December 11, 2015

Becton, Dickinson and Company  
Matthew Trachtenberg  
Staff Regulatory Affairs Specialist  
1 Becton Drive  
Franklin Lakes, New Jersey 07666

Re: K153257

Trade/Device Name: BD FlowSmart™ Set/MiniMed™ Pro-Set™

Regulation Number: 21 CFR 880.5440

Regulation Name: Intravascular Administration Set

Regulatory Class: Class II

Product Code: FPA, FPK

Dated: November 9, 2015

Received: November 10, 2015

Dear Mr. Matthew Trachtenberg:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Erin I. Keith -S**

Erin I. Keith, M.S.

Director

Division of Anesthesiology,

General Hospital, Respiratory,

Infection Control, and Dental Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K153257

Device Name

BD FlowSmart™ Set/MiniMed™ Pro-Set™

Indications for Use (Describe)

The BD FlowSmart™/MiniMed™ Pro-Set™ infusion set is intended for the subcutaneous infusion of medication, including insulin, from an external infusion pump.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **510(k) Summary**

K153257

**Submitted By:** Matthew S. Trachtenberg  
Staff Regulatory Affairs Specialist, BD Medical  
1 Becton Drive  
Franklin Lakes, NJ 07417  
Tel: 201 847 6337  
Fax: 201 847 5307

**Date Prepared:** December 4, 2015

**Device Name:** Trade Name: BD FlowSmart™ Set/MiniMed™ Pro-Set™  
Common Name: Subcutaneous Infusion Set  
Classification: Class II device; 21 CFR 880.5440,  
Intravascular Administration set  
Product Code: FPA (intravascular administration set)  
FPK (tubing, intravascular administration set)

### **Legally marketed predicate devices to which substantial equivalence is being claimed:**

K150059 – BD FlowSmart™ Set

### **Device Description:**

The BD FlowSmart™/MiniMed™ Pro-Set™ infusion set is a subcutaneous administration set intended to interface with commercially available infusion pumps with suitable connections. The infusion set features a flexible perforated catheter perpendicular to an adhesive patch and detachable tubing. The tubing is connected on one end to the medication reservoir of an infusion pump and on the other end to the patient, attached to the skin by an adhesive base. It is a single-use sterile device.

### **Intended Use:**

The BD FlowSmart™/MiniMed™ Pro-Set™ infusion set is intended for the subcutaneous infusion of medication, including insulin, from an external infusion pump.

### **Comparison with Predicate Devices:**

The BD FlowSmart™ Set/MiniMed™ Pro-Set™ infusion set has the same intended use as its predicate device for the subcutaneous delivery of fluids and medication, including insulin, from an external infusion pump. Additionally, the fundamental scientific technology of the proposed device is unchanged from the legally marketed device. The technological characteristics include an adhesive base, catheter, and tubing connection components. The needle insertion method, 90 degree angle of insertion, and tubing connection types are identical features to its predicate device. The modifications made include the incorporation of an alternate adhesive material and a minor modification to the tubing configuration. Non-clinical testing supports substantial equivalence of the subject device despite these technological differences. The table below provides a side by side comparison of the subject device compared to its predicate.

<b>Feature</b>	<b>Subject Device: BD FlowSmart™ Set/ MiniMed™ Pro-Set™</b>	<b>Predicate Device: BD FlowSmart™ Set</b>
<i>510(k) Number</i>	K153257	K150059
<i>Manufacturer</i>	BD	BD
<i>Intended Use</i>	For the subcutaneous infusion of medication, including insulin, from an external infusion pump.	For the subcutaneous infusion of medication, including insulin, from an external infusion pump.
<i>System Components</i>	Infusion set + tubing	Infusion set + tubing
<i>Needle Insertion Method</i>	Manual or Automatic	Manual or Automatic
<i>Insertion Needle Gauge</i>	30G	30G
<i>Angle of Insertion</i>	Straight 90 degrees	Straight 90 degrees
<i>Plastic Cannula Length</i>	6mm	6mm
<i>Plastic Cannula Gauge</i>	28G	28G
<i>Tubing Connection Type</i>	Paradigm® and Luer Lock	Paradigm® and Luer Lock
<i>Tubing Length (cm)</i>	61 and 107	61 and 107
<i>Connection Positions</i>	Multiple	Multiple
<i>Replacement Frequency</i>	Disposable, replaced every 72 hours	Disposable, replaced every 72 hours
<i>Provided Sterile</i>	YES	YES

**Testing:**

BD has performed non-clinical testing to demonstrate substantial equivalence. This testing included assessments related to device material biocompatibility (ISO 10993: Biological evaluation of medical devices), and device functional performance (ISO 10555-1: Sterile single use intravascular catheters, ISO ID26: Medical Electrical Equipment. Part 2: Particular requirements for the safety of infusion pumps, ISO 11135-1: Sterilization of healthcare products). Results of these tests were found to be acceptable and demonstrated that the subject infusion set met requirements for its intended use and is substantially equivalent to its predicate device.

**Conclusion:**

The analysis and testing performed demonstrate the BD FlowSmart™ Set/MiniMed™ Pro-Set™ device is substantially equivalent to its predicate device.