



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Medicrea International S.A.
Mr. David Ryan
VP Product Development and Marketing
14 Porte Du Grand Lyon
01700 Neyron
FRANCE

April 22, 2016

Re: K153413
Trade/Device Name: PASS LP Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, OSH, MNH, MNI, KWP
Dated: April 8, 2016
Received: April 11, 2016

Dear Mr. Ryan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K153413

Device Name
PASS LP Spinal System

Indications for Use (Describe)

The PASS LP Spinal System is a pedicle screw fixation system intended for immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (e.g., fracture or dislocation), deformity or curvature (e.g., scoliosis, kyphosis, and/or lordosis), tumor, spinal stenosis, pseudarthrosis, or failed previous fusion.

Except for rod plates, when used for posterior noncervical pedicle screw fixation in pediatric patients, the PASS LP Spinal System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. Additionally, the system is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/spondylolysis and fracture caused by tumor and/or trauma. The PASS LP Spinal System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

1. **DEVICE SUBMITTER**

MEDICREA® INTERNATIONAL S.A.

14 Porte Du Grand Lyon

01700 NEYRON- France

Phone: +33 4 72 01 87 87

Fax: +33 4 72 01 87 88

Contact Person:

David RYAN

VP Product Development and Marketing

dryan@medicrea.com

Date Prepared: April 8, 2016

2. **DEVICE**

Name of Device: PASS LP Spinal System

Common or Usual Name: Pedicle Screw Spinal Lumbar System

Classification Name:

- ✓ Orthosis, Spinal Pedicle Fixation per MNI 888.3070
- ✓ Orthosis, Spondylolisthesis Spinal Fixation per MNH 888.3070
- ✓ Appliance, Fixation, Spinal Interlaminar per KWP 888.3050
- ✓ Pedicle Screw Spinal System, Adolescent Idiopathic Scoliosis per OSH 888.3070
- ✓ Orthosis, Spinal Pedicle Fixation, For Degenerative Disc Disease per NKB 888.3070

Regulatory Class: III

Classification Product Code: NKB

Subsequent Product Code: OSH, MNH, MNI, and KWP

3. **PREDICATE DEVICE**

Primary Predicate:

- ✓ PASS LP Spinal System, K141398.

This predicate has not been subject to a design-related recall.

Additional Predicate:

- ✓ PASS LP Spinal System, K112493

No reference device was used in this submission

4. DEVICE DESCRIPTION

The PASS LP Spinal System is designed to contribute to correction and surgical stabilization of the thoracic, lumbar and sacral spine.

The system consists of pedicle screws, hooks, sacral plates, iliac screws, clamps, rods, nuts, rod plates and crosslink components. It can be used for single or multiple level fixations. Components are manufactured from titanium alloy (Ti-6Al-4V ELI) that conforms to ASTM F136 or cobalt-chromium molybdenum alloy Co-Cr28Mo6 that conforms to ISO 5832-12 and ASTM F1537.

A subset of PASS LP Spinal System components may be used for posterior pedicle screw fixation in pediatrics cases. These constructs may be comprised of a variety of shapes and sizes of rods, hooks, sacral plates, iliac screws, clamps, nuts and crosslink components. The PASS LP components can be rigidly locked into a variety of configurations, with each construct being tailored made for the individual case.

The purpose of this submission is to extend to the PASS LP Spinal System, with the addition of new components: Polyaxial Cannulated Pedicle Screws.

5. INDICATIONS FOR USE

The PASS LP Spinal System is a pedicle screw fixation system intended for immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (e.g., fracture or dislocation), deformity or curvature (e.g., scoliosis, kyphosis, and/or lordosis), tumor, spinal stenosis, pseudarthrosis, or failed previous fusion.

Except for rod plates, when used for posterior non-cervical pedicle screw fixation in pediatric patients, the PASS LP Spinal System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. Additionally, the system is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/spondylolysis and fracture caused by tumor and/or trauma. The PASS LP Spinal System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

The Indications For Use statement for the PASS LP Spinal System is identical to the predicate device. Both the subject and predicate device have the same intended use for the treatment of acute and chronic instabilities or deformities of thoracic, lumbar, and sacral spine.

6. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

The table below compares the features and characteristics of the PASS LP Spinal System to its predicate device.

Device	PASS LP Spinal System – New Components	PASS LP Spinal System	PASS LP Spinal System
510(k) number	K153413	K141398	K112493
Intended use			
Thoracic, Lumbar spine	Yes	Yes	Yes
Posterior Approach	Yes	Yes	Yes
Design			
Screw Diameters	Ø4.5mm	Ø4.5mm to Ø7.5mm	Ø5.0mm to Ø7.5mm
Screw Lengths	From 25mm to 60mm with 5mm increment	From 25mm to 60mm with 5mm increment	From 25mm to 60mm with 5mm increment
Cannulated screws	Yes	No	Yes
Color Coded	Yes	Yes	Yes
Shape	Conical thread Spherical head for connection with other PASS LP components Breakable threaded part for Nut tightening	Conical thread Spherical head for connection with other PASS LP components Breakable threaded part for Nut tightening	Conical thread Spherical head for connection with other PASS LP components Breakable threaded part for Nut tightening
Threaded Extension	Long threaded extension	Long threaded extension	Long threaded extension
Materials			
PASS LP Additional Components	Titanium alloy (Ti-6Al-4V ELI, following standards ASTM F136 and ISO 5832-3)	Titanium alloy (Ti-6Al-4V ELI, following standards ASTM F136 and ISO 5832-3)	Titanium alloy (Ti-6Al-4V ELI, following standards ASTM F136 and ISO 5832-3)

Material composition is identical to other MEDICREA® INTERNATIONAL products that have been cleared via the 510(k) process.

7. PERFORMANCE DATA

Biocompatibility Testing

The biocompatibility evaluation for the PASS LP system was conducted in accordance with the FDA blue book Memorandum #G95-1 "Use of International Standard ISO-10993, 'Biological Evaluation of Medical Devices Part 1: Evaluation and Testing,'" May 1, 1995, and International Standard ISO 10993-1 "Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing Within a Risk Management Process," as recognized by FDA. The battery of testing included the following tests:

- ✓ Cytotoxicity
- ✓ Sensitization
- ✓ Irritation
- ✓ Systemic toxicity
- ✓ Pyrogen Testing

According to the standard **ISO 10993-1**, the PASS LP Spinal System is defined as implantable device in contact with tissue and bone and as a permanent contact with the patient.

For chemical composition, the material conform to Ti-6Al-4V ELI, following standards ASTM F136 and ISO 5832-3.

Mechanical testing

When applicable, the tests performed on the additional components (pull out according to ASTM F543; static torsion and static bending according to ASTM F2193) indicate that the products are as mechanically sound as other devices commercially available.

Clinical study

No clinical studies were performed.

Animal study

No animal studies were performed.

8. CONCLUSION

MEDICREA® INTERNATIONAL S.A. PASS LP Spinal System is substantially equivalent to its predicate device in terms of indications for use, design, material and function.