



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

March 2, 2016

Medtronic Sofamor Danek USA, Incorporated  
Ankit K. Shah  
Senior Regulatory Affairs Specialist  
1800 Pyramid Place  
Memphis, Tennessee 38132

Re: K153438

Trade/Device Name: PERIMETER<sup>®</sup> C (Titanium) Spinal System  
Regulation Number: 21 CFR 888.3080  
Regulation Name: Intervertebral body fusion device  
Regulatory Class: Class II  
Product Code: ODP  
Dated: February 8, 2016  
Received: February 10, 2016

Dear Ankit Shah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Lori A. Wiggins -S**

for  
Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K153438

Device Name

PERIMETER® C (Titanium) Spinal System

Indications for Use (Describe)

The PERIMETER® C Spinal System is intended to be used for anterior cervical interbody fusion procedures in skeletally mature patients with cervical disc disease at one level from the C2-C3 disc to the C7-T1 disc who have had six weeks of non-operative treatment. Cervical disc disease is defined as intractable radiculopathy and/or myelopathy with herniated disc and/or osteophyte formation on posterior vertebral endplates producing symptomatic nerve root and/or spinal cord compression confirmed by radiographic studies. Additionally, the PERIMETER® C Spinal System implants are to be used with autogenous bone graft and supplemental fixation and implanted via an open, anterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**PERIMETER® C (Titanium) Spinal System**  
**510(k) Summary**  
**January 2016**

- I. Submitter:** Medtronic Sofamor Danek USA, Inc.  
1800 Pyramid Place  
Memphis, TN 38132  
Telephone: (901)396-3133  
Fax: (901) 346-9738
- Contact:** Ankit K. Shah  
Senior Regulatory Affairs Specialist
- Date Prepared:** January 6, 2016
- II. Device**
- Name of Device:** PERIMETER® C (Titanium) Spinal System
- Common or Usual Name:** Intervertebral body fusion device
- Classification Name:** Intervertebral fusion device with bone graft,  
cervical: 21 CFR 888.3080
- Regulatory Class:** Class II
- Product Code:** ODP
- III. Predicate Device**
- K132584 (S.E. 12/04/2013) PERIMETER® C Spinal System – Primary Predicate
- K133645 (S.E. 01/03/2014) – PERIMETER® C Spinal System - Additional Predicate
- The predicates device have not been subject to a design related recall.*
- IV. Device Description:**
- The PERIMETER® C (Titanium) Spinal System consists of spacers/cages of various widths and heights, which can be inserted between two cervical vertebral bodies to give support and correction during cervical interbody fusion procedures. Additionally, this implant has six degrees of lordosis and the superior and inferior surfaces of the implant

are designed with teeth which interact with the surface of the vertebral endplates to aid in resisting expulsion. The hollow geometry of the implants allows them to be packed with autogenous bone graft and is to be used with supplemental fixation in all procedures.

The purpose of this submission is to include additional interbody cages manufactured from medical grade titanium alloy (Ti-6Al-4V ELI) that are packaged sterile via gamma irradiation.

**V. Indications for Use:**

The PERIMETER® C Spinal System is intended to be used for anterior cervical interbody fusion procedures in skeletally mature patients with cervical disc disease at one level from the C2-C3 disc to the C7-T1 disc who have had six weeks of non-operative treatment. Cervical disc disease is defined as intractable radiculopathy and/or myelopathy with herniated disc and/or osteophyte formation on posterior vertebral endplates producing symptomatic nerve root and/or spinal cord compression confirmed by radiographic studies. Additionally, the PERIMETER® C Spinal System implants are to be used with autogenous bone graft and supplemental fixation and implanted via an open, anterior approach.

**VI. Comparison of Technological Characteristics with Predicate Device**

The subject PERIMETER® C (Titanium) Spinal System Devices have identical: indications for use, intended use, design, material (Titanium Alloy per ASTM F136), sterilization methods, levels of attachment and fundamental scientific technology as the PERIMETER® C Spinal System predicate previously cleared by the FDA in K132584 (S.E. 12/04/2013). Additionally, the new PERIMETER® C (Titanium) Spinal System Devices are provided sterile identical to that of the predicate 2 devices found in K133645 (S.E. 01/03/2014).

**VII. Performance Data**

The following performance data are provided in support of the substantial equivalence determination.

**Biocompatibility**

The biocompatibility evaluation for the PERIMETER® C (Titanium) Spinal System devices was conducted in accordance with FDA's Draft Guidance for Industry and FDA Staff "Use of International Standard ISO-10993, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing" issued April, 23, 2013.

The subject PERIMETER® C (Titanium) Spinal System intervertebral body fusion devices are permanent implants and will be classified as permanent, >30 day body contact according to with FDA's Draft Guidance for Industry and FDA Staff "Use of International Standard ISO-10993, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing". The subject intervertebral body fusion devices are manufactured from identical materials as the predicate devices, in accordance with the following ASTM standard:

**ASTM F136** – Standard Specification for Wrought Titanium-6 Aluminum-4 Vanadium ELI (Extra Low Interstitial) Alloy for Surgical Implant Applications

The titanium alloy material used for the subject PERIMETER® C (Titanium) Spinal System intervertebral body fusion devices has a long clinical history of safe and effective use in similar medical devices. Therefore, no additional biocompatibility testing is required.

**Mechanical Testing**

In accordance with, Guidance for Industry and FDA Staff – Spinal System 510(k)'s", Medtronic has evaluated the subject devices to demonstrate substantial equivalence to the predicate devices. It was determined that subject devices do not represent a new worst case. An engineering rationale was used to demonstrate substantial equivalence. As a new worst case has not been indicated and an engineering rationale was deemed adequate to prove equivalence to the predicate device, no additional mechanical testing is required.

## **VIII. Conclusion**

Based on a risk analysis, engineering rationale, and additional supporting documentation provided in the pre-market notification, the subject PERIMETER® C (Titanium) Spinal System intervertebral body fusion devices are as safe and effective as the following predicates: K132584 (S.E. 12/04/2013) – Primary Predicate and K133645 (S.E. 01/03/2014) - Predicate 2.