



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

May 13, 2016

Dental Direkt Gmbh  
% Ms. Vera Buescher  
Quality & Regulatory Manager  
Dental DireK153490/S001kt Gmbh (former Dental Direkt Of Amerika Ug)  
Industriezentrum 106-108  
Spenge, 32139 DE

Re: K153490

Trade/Device Name: Dd Medical Polymers  
Regulation Number: 21 CFR 872.3770  
Regulation Name: Temporary Crown And Bridge Resin  
Regulatory Class: Class II  
Product Code: EBG  
Dated: April 11, 2016  
Received: April 14, 2016

Dear Ms. Buescher:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Susan Runno DDS, MA". The signature is written in a cursive style. In the background, there is a faint, semi-transparent logo of the FDA (U.S. Food and Drug Administration).

Erin I. Keith, M.S.  
Director  
Division of Anesthesiology,  
General Hospital, Respiratory,  
Infection Control, and Dental Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

**Indications for Use**

**510(k) Number:** K153490

**Device Name:** DD medical polymers:  
- DD tempMED  
- DD Bio Splint P

**Indications for Use:**

DD medical polymers are indicated for temporary ( $\leq 12$  months) crowns, bridges and bite splints. Applications include both anterior and posterior structures.

Prescription Use   X    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



510(k) Number: **K153490**

## **510(k) Summary**

<b>Submitter of 510(k)</b>	Dental Direkt GmbH Industriezentrum 106-108 32139 Spenge / Germany
<b>Contact Persons</b>	Mr. Uwe Greitens, Authorized Representative Ms. Vera Büscher, Quality & Regulatory Manager Phone: +49 5225 86319-0 Fax: +49 5225 86319-99
<b>Establishment Registration Number</b>	3008347275
<b>Date Prepared</b>	2016/05/12
<b>Trade Name of Device</b>	DD medical polymers: - DD tempMED - DD Bio Splint P
<b>Common Name</b>	Crown and Bridge, Temporary, Resin
<b>Classification Name</b>	Temporary crown and bridge resin
<b>Regulation Number</b>	21 CFR 872.3770
<b>Classification Product Code</b>	EBG, MQC
<b>Panel</b>	Dental
<b>Classification</b>	Class II
<b>Predicate Device</b>	K150432: Union Dental S.A. Idodentine (Dental Polymer Blank)

<b>Indications for Use</b>	DD medical polymers are indicated for temporary ( $\leq 12$ months) crowns, bridges and bite splints. Applications include both anterior and posterior structures.
<b>Device Description</b>	DD medical polymers, made from medical grade, are industrially polymerized, pre-colored or clear dental milling blanks designed for milled fabrication of temporary anterior and posterior crowns and bridges (tooth-colored variants) or bite splints (clear variant) on dental CAD/CAM systems.
<b>Technological characteristics</b>	The technologic characteristics are highly similar as demonstrated in performance testing and in chemical composition; both devices are composed primarily of polymethylmethacrylate, while the amount and percentage of color oxides in the submission device varies from the predicate.
<b>Comparison of Required Technology Characteristics</b>	The following table shows a summary of the technological characteristics of DD medical polymers compared to the predicate device.

<b>Comparison of Required Technology Characteristics</b>		
<b>Feature</b>	<b>Submission device</b>	<b>Predicate device</b>
<b>Trade name</b>	<i>DD medical polymers</i>	<i>Idodentine Disc</i>
<b>510(k)</b>	K153490	K150432
<b>Product code</b>	EBG, MQC	EBG
<b>Regulatory class</b>	Class II	Class II
<b>Manufacturer</b>	Dental Direkt GmbH	Union Dental S.A.
<b>Intended use</b>	DD medical polymers are indicated for temporary ( $\leq 12$ months) crowns, bridges and bite splints. Applications include both anterior and posterior structures.	Acrylic polymer blank particularly suitable for making removable or temporary dental structures such as crowns and bridges using milling technology using CAD/CAM. Indications - Temporary anterior and posterior crowns - Temporary anterior and posterior bridges with up to two adjacent pontics Maximum recommended usage period: 12 months - Removable structures for dentures (denture bases) - Removable structures for therapeutic restorations (bite splints or occlusal splints)
<b>Max. application time</b>	12 months	12 months
<b>Technology</b>	Blank for dental CAD/CAM machining	Blank for dental CAD/CAM machining
<b>Shape</b>	Disc	Disc or block
<b>Shade</b>	VITA-shades, clear	VITA-shades, clear, pink
<b>Comparison of Required Technology Characteristics</b>		
<b>Raw material</b>	PMMA	PMMA
<b>Chemical composition [Units]</b>		
<b>Material base</b>	> 99.0 [wt%] (Polymethyl methacrylate)	Polymethyl methacrylate
<b>Coloring oxides</b>	$\leq 1.0$ [wt%]	Not specified
<b>Physical characteristics [Units]</b>		
<b>Performance testing</b>	Tested according to DIN EN ISO 20795-1 and ISO 10477	Tested according to DIN EN ISO 20795-1 and ISO 10477
<b>Biocompatibility</b>	EN ISO 10933-1, -5	EN ISO 10933-1, -5,

## Discussion of Tests Performed

### Clinical Tests

Dental Direkt GmbH did not conduct, nor rely upon, clinical tests to determine substantial equivalence.

### Non Clinical Tests

Non-clinical testing according to the following standards was performed in order to validate the product against the company's specified design requirements:

- EN ISO 10993-1 (biological compatibility) and EN ISO 10993-5 (cytotoxicity):

Finished *DD medical polymer* products in all variants were tested by an accredited testing laboratory (GLP-certified) with respect to biocompatibility. The laboratory certified that "the insolubility is in compliance with the requirements of the standard. There is no evidence that effects hazardous to the patient will arise by release of leachable ingredients/contaminants".

- DIN EN ISO 20795-1:2013: Dentistry - Base polymers - Part 1: Denture base polymers
- ISO 10477:2004: Dentistry – Polymer-based Crown and Bridge Materials

Requirement	Requirements acc. to DIN EN ISO 20795-1 / ISO 10477:				
	Required Value (ISO 20795-1)	Required Value (ISO 10477)	Value Subm. Device	Value Predicate Device	Passed / Failed
Flexural strength [MPa]	≥ 65	≥ 50	≥ 75	90	Passed
Flexural modulus [MPa]	≥ 2000	-	2800 (± 200)	not specified	Passed
Residual MMA monomer [%]	≤ 2.2	-	0.4	1.4	Passed
Water sorption [ $\mu\text{g}/\text{mm}^3$ ]	≤ 32	≤ 40	≤ 23	23	Passed
Solubility [ $\mu\text{g}/\text{mm}^3$ ]	≤ 1.6	≤ 7.5	0.2	0.0	Passed

**Substantial  
Equivalence  
Conclusion**

The small difference in the indication of the predicate device is due to a bigger range of shades available. The missing pink variant by Dental Direkt is the only reason why the application field of denture bases is not contained in the indication of DD medical polymers. The restriction regarding the maximum amount of pontics for bridge constructions is listed as contraindication in the IFU of DD medical polymers. In summary neither the fundamental intended use nor the safety and effectiveness of the product is influenced by the deviation of the indication of submission device and predicate device. Based on comparison of technology, including composition and performance testing, biocompatibility testing, and highly similar indications for use, DD medical polymers are substantially equivalent to Idodentine (K150432).