



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

June 1, 2016

Roche Molecular Systems, Inc.
Jintao Chen, Ph.D.
Director, Regulatory Affairs
4300 Hacienda Drive
Pleasanton, CA 94588-2722

Re: P150047

Trade/Device Name: **cobas®** EGFR Mutation Test v2

Filed: December 9, 2015

Amended: December 9, 2015; December 22, 2015; January 14, 2016; January 15, 2016;
February 1, 2016; May 20, 2016

Product Code: OWD

Dear Dr. Jintao Chen:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) for the **cobas®** EGFR Mutation Test v2. This device is indicated for:

The **cobas®** EGFR Mutation Test v2 is a real-time PCR test for the qualitative detection of defined mutations of the epidermal growth factor receptor (EGFR) gene in non-small cell lung cancer (NSCLC) patients. Defined EGFR mutations are detected using DNA isolated from formalin-fixed paraffin-embedded tumor tissue (FFPET) or circulating-free tumor DNA (cfDNA) from plasma derived from EDTA anti-coagulated peripheral whole blood.

The test is indicated as a companion diagnostic to aid in selecting NSCLC patients for treatment with the targeted therapies listed in Table 1 below in accordance with the approved therapeutic product labeling.

Table 1

Drug	FFPET	Plasma
TARCEVA® (erlotinib)	Exon 19 deletions and L858R	Exon 19 deletions and L858R
TAGRISSTO™ (osimertinib)	T790M	

Patients with positive **cobas®** EGFR Mutation Test v2 test results using plasma specimens for the presence of EGFR exon 19 deletions or L858R mutations are eligible for treatment with TARCEVA® (erlotinib). Patients who are negative for these mutations by this test

should be reflexed to routine biopsy and testing for EGFR mutations with the FFPET sample type.

Drug safety and efficacy have not been established for the EGFR mutations listed in Table 2 below that are also detected by the cobas® EGFR Mutation Test v2.

Table 2

Drug	FFPET	Plasma
TARCEVA® (erlotinib)	G719X, exon 20 insertions, T790M, S768I and L861Q	G719X, exon 20 insertions, T790M, S768I and L861Q
TAGRISSTO™ (osimertinib)	G719X, exon 19 deletions, L858R, exon 20 insertions, S768I, and L861Q	G719X, exon 19 deletions, L858R, exon 20 insertions, T790M, S768I, and L861Q

For manual sample preparation, FFPET specimens are processed using the cobas® DNA Sample Preparation Kit and plasma specimens are processed using the cobas® cfDNA Sample Preparation Kit. The cobas z 480 analyzer is used for automated amplification and detection.

We are pleased to inform you that the PMA is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). FDA has determined that this restriction on sale and distribution is necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 18 months when stored at 2-8°C. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84. This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final UDI rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). Additionally, 21 CFR 814.84 (b)(4) requires PMA annual reports submitted after September 24,

2014, to identify each device identifier currently in use for the subject device, and the device identifiers for devices that have been discontinued since the previous periodic report. It is not necessary to identify any device identifier discontinued prior to December 23, 2013. For more information on these requirements, please see the UDI website, <http://www.fda.gov/udi>.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

You have agreed to provide the following data as part of a future supplement:

1. Limited specimen handling and specimen storage stability information was included in the PMA application. Additional testing is needed to determine the robustness and consistency of assay performance in regard to both pre-analytical and storage conditions. Successful completion of the agreed upon study protocol is needed to validate and support claims regarding whole blood and plasma sample handling and storage stability for the **cobas**[®] EGFR Mutation Test v2 using clinical plasma samples. The final study data, study conclusions, and labeling revisions should be submitted within 2 years of PMA approval date.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process" <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm>.

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at <http://www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm>.

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA CDRH Internet HomePage located at <http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in 6 copies, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing.

U.S. Food and Drug Administration
Center for Devices and Radiological Health
PMA Document Control Center - WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Karen Bijwaard at 301-796-6162 or Karen.Bijwaard@fda.hhs.gov.

Sincerely yours,

Reena Philip -S

Reena Philip, Ph.D.

Director

Division of Molecular Genetics and Pathology

Office of *In Vitro* Diagnostics and Radiological

Health

Center for Devices and Radiological Health