



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

October 19, 2016

Diabetic Supply of Suncoast, Inc.  
c/o Matt Hedlund  
510k Consultant  
3924 NE 157th Place  
Lake Forest Park, Washington 98155

Re: K160199  
Trade/Device Name: Advocate Insulin Pen Needles  
Regulation Number: 21 CFR 880.5570  
Regulation Name: Hypodermic Single Lumen Needle  
Regulatory Class: II  
Product Code: FMI  
Dated: September 21, 2016  
Received: September 22, 2016

Dear Matt Hedlund:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

 Tina Kiang  
-S

Tina Kiang, Ph.D.  
Acting Director  
Division of Anesthesiology,  
General Hospital, Respiratory,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K160199

Device Name

Advocate Insulin Pen Needles

Indications for Use (Describe)

These disposable sterile insulin pen needles are intended for subcutaneous injection of insulin in the treatment of diabetes.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRAStaff@fda.hhs.gov](mailto:PRAStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary for K160199

Diabetic Supply of Suncoast, Inc.  
PO Box 2102  
Vega Alta, PR 00692  
1-866-373-2824  
Contact: Victoria Thuss  
Date Prepared: October 11, 2016

1. Subject Device  
Trade Name: Advocate Insulin Pen Needles  
Common Name: Pen Needle  
Regulation Number: 21 CFR 880.5570  
Regulation Name: Hypodermic single lumen needle  
Regulatory Class: Class II  
Product Code: FMI
  
2. Predicate Device  
Trade Name: Comfort EZ Pen Needle  
510(k) Number: K121632  
Common Name: Pen Needle  
Regulation Number: 21 CFR 880.5570  
Regulation Name: Hypodermic single lumen needle  
Regulatory Class: Class II  
Product Code: FMI
  
3. Description of Device: The Advocate Insulin Pen Needle consists of a polypropylene 'outer cap' enclosure. This polypropylene cap contains an opening at one end to allow the needle or cannula to exit. The other end of the outer cap is the 'hub' that can be connected to various pens. Contained within the outer cap is the cannula assembly made up of the stainless steel cannula that is contained within an 'inner cap' polypropylene shell that is located on top of the hub. Medical grade blister paper covers and seals the hub. The Advocate Insulin Pen Needles when sealed with the blister paper are EO sterilized and provide a sterile fluid path during use. The Advocate Insulin Pen Needles when connected to the pen injector, is intended for subcutaneous injection of insulin.
  
4. Indications for Use: *These disposable sterile insulin pen needles are intended for subcutaneous injection of insulin in the treatment of diabetes.* This is the same indication for use statement as the statement for the predicate device, K121632.

5. Technological characteristics: The following is a comparison of the technological characteristics of the subject device with the predicate device:

Technological Characteristic	Subject Device Advocate Insulin Pen Needle	Predicate Device Comfort EZ Pen Needle - K121632
Size	31G 5 mm; 31G 6 mm; 31G 8 mm; 32G 4 mm; 33 G 4 mm; 29G 12mm	31 G 5 mm; 31 G 6 mm; 31 G 8 mm; 32 G 4 mm; 32 G 5 mm; 32 G 6 mm
Design	Outer Cap; Inner Cap; Cannula; Hub; Paper Seal	Outer Cap; Inner Cap; Cannula; Hub; Paper Seal
Material	Polypropylene; Stainless Steel; Dialyzing Paper; Glue; Silicone Oil	Polypropylene; Stainless Steel; Dialyzing Paper; Glue; Silicone Oil
Sterilization	EO Sterilization  SAL of 10 <sup>-6</sup>	EO Sterilization  SAL of 10 <sup>-6</sup>
Labeling	Primary Packaging Label- Size; EO Sterile; Use By Date; Lot; Single Use; Manufacturer Information; Warnings; Enclosed Package Insert.	Primary Packaging Label- Size; EO Sterile; Use By Date; Lot; Single Use; Manufacturer Information; Warnings.
Performance Testing	Biocompatibility Pen Injector (NIS) Compatibility Mechanical Testing Sterility Validation EO Residue Shelf Life Testing	Biocompatibility Pen Injector (NIS)Compatibility Mechanical Testing Sterility Validation EO Residue Shelf Life Testing

Conclusion: The technological characteristics of the Advocate Insulin Pen Needles are the same as the Predicate Device. The differences of certain pen sizes (33 G 4 mm; 29G 12mm) with the Predicate Device do not raise new questions of safety and/or effectiveness as these two pen sizes have the same intended use, material, design, sterilization, labeling, and performance testing as the Predicate Device.

6. Performance testing: The following table defines the nonclinical performance testing submitted. All pre-determined acceptance criteria were met:

Performance Test	Normative References/ Description
Accelerated Shelf Life	ASTM F 1980-07 ISO 11607-1 ISO 11608-2 ASTM F 1929 ASTM F88
EO Sterilization Validation	The validation process used a standard half cycle method to demonstrate the efficacy of the EO sterilization process.
EO Residual	Sterilant residual limit for limited exposure met of EO<4mg and ECH <9mg.
Biocompatibility Testing	Cytotoxicity – ISO 10993-5 Sensitization – ISO 10993-10 Intracutaneous reactivity – ISO 10993-10 Systemic Toxicity - ISO 10993-11 Haemocompatibilitiy – ISO 10993-4
Performance Testing	ISO 11608 – 2 Includes mechanical testing and determination of functional compatibility with specific pen injector devices (NIS).

Conclusion: The performance testing demonstrates that the Advocate Insulin Pen Needles submitted under this 510(k) are substantially equivalent to the Predicate Device.