



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

June 9, 2016

Medos International SARL  
% Ms. Kristine Christo  
Director, Regulatory Affairs  
Depuy Mitek Inc., a Johnson & Johnson Company  
325 Paramount Drive  
Raynham, Massachusetts 02767

Re: K161001

Trade/Device Name: MILAGRO<sup>®</sup> ADVANCE PEEK Interference Screw  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or threaded metallic bone fixation fastener.  
Regulatory Class: Class II  
Product Code: MBI  
Dated: April 7, 2016  
Received: April 11, 2016

Dear Ms. Christo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K161001

Device Name

MILAGRO® ADVANCE PEEK Interference Screw

Indications for Use (Describe)

The MILAGRO® ADVANCE PEEK Interference Screw is intended for attachment of soft tissue grafts or bone-tendon-bone grafts to the tibia and/or femur during cruciate ligament reconstruction procedures.

Additionally, the 7, 8 and 9 mm x 23 mm screws are indicated for: medial and lateral collateral ligament repair, medial patellofemoral ligament reconstruction (femur fixation) of the knee, proximal bicep tenodesis in the shoulder and distal bicep tenodesis in the elbow.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASaff@fda.hhs.gov](mailto:PRASaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**SECTION 2 - 510(k) SUMMARY****MILAGRO® ADVANCE PEEK Interference Screw**


---

**Submitter's Name and Address** DePuy Mitek  
a Johnson & Johnson company  
325 Paramount Drive  
Raynham, MA 02767

Date Prepared: March 25, 2016

---

**Contact Person** Kristine Christo  
Director, Regulatory Affairs Telephone: 508-828-3359  
DePuy Mitek, Inc. Facsimile: 508-977-6911  
a Johnson & Johnson company e-mail: [kchristo@its.jnj.com](mailto:kchristo@its.jnj.com)  
325 Paramount Drive  
Raynham, MA 02767, USA

---

**Name of Medical Device** Proprietary Name: MILAGRO® ADVANCE PEEK Interference Screw  
Classification Name: Smooth or threaded metallic bone fixation fasteners  
Common Name: Fastener, fixation, nondegradable, soft tissue

---

**Substantial Equivalence** The MILAGRO® ADVANCE PEEK Interference Screw is substantially equivalent to:  

- K123362, K143660, MILAGRO ADVANCE™ Interference Screw

Reference device:  

- K130539 Healix Advance™ Knotless PEEK Anchor

---

**Device Classification** Smooth or threaded metallic bone fixation fastener, classified as Class II, product code MBI, regulated under 21 CFR 888.3040.

---

**Device Description** The MILAGRO® ADVANCE PEEK Interference Screw is a non-absorbable, tapered, cannulated, threaded fastener for use in interference fixation of soft tissue grafts or bone-tendon-bone grafts. The Interference Screw is made from Polyetheretherkeytone (PEEK). The MILAGRO® ADVANCE PEEK Interference Screw is provided sterile and is for single patient use only.

---

---

<b>Technological Characteristics</b>	The proposed MILAGRO® ADVANCE PEEK Interference Screw is similar to the predicate MILAGRO ADVANCE™ Interference Screw (K123362, K143660) in that they share the same indication for use, screw design, sterilization method, and shelf life. The proposed MILAGRO® ADVANCE PEEK Interference Screw is similar to the reference device, Healix Advance™ Knotless PEEK Anchor (K130539), in that they share the same PEEK material.
--------------------------------------	---

---

<b>Indications for Use</b>	The MILAGRO® ADVANCE PEEK Interference Screw is intended for attachment of soft tissue grafts or bone-tendon-bone grafts to the tibia and/or femur during cruciate ligament reconstruction procedures.
----------------------------	--

Additionally, the 7, 8 and 9 mm x 23 mm screws will be indicated for: medial and lateral collateral ligament repair, medial patellofemoral ligament reconstruction (femur fixation) of the knee, proximal bicep tenodesis in the shoulder and distal bicep tenodesis in the elbow.

---

<b>Non clinical Testing</b>	Verification activities were performed on the implant and / or its predicate. Testing assessments include pull out testing, and insertion / failure torque.
-----------------------------	---

<b>Safety and Performance</b>	Results of performance testing have demonstrated that the proposed devices are suitable for their intended use.
-------------------------------	---

The proposed device also met requirement of bacterial endotoxin testing.

Based on similarities in the indications for use, technological characteristics, and performance in comparison to the predicate devices, the proposed MILAGRO® ADVANCE PEEK Interference Screw has shown to be substantially equivalent to the predicate device under the Federal Food, Drug and Cosmetic Act.

---

