

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center – WO66-G609 Silver Spring, MD 20993-0002

August 12, 2016

Medtronic Sofamor Danek Mr. Ankit Shah Senior Regulatory Affairs Specialist 1800 Pyramid Place Memphis, Tennessee 38132

Re: K161210

 Trade/Device Name:
 RIALTO[™] SI Fusion System, MEDTRONIC RESUSABLE INSTRUMENTS FOR USE WITH IPC[™] POWEREASE[™] SYSTEM, MEDTRONIC NAVIGATED MANUAK REUSABLE INSTRUMENTS FOR USE WITH THE STEALTHSTATION[™] AND IPC[™] POWEREASE[™]SYSTEMS

 Regulation Number:
 21 CFR 888.3040

 Regulation Name:
 Smooth or threaded metallic bone fixation fastener

 Regulatory Class:
 Class II

 Product Code:
 OUR, OLO, HWE

 Dated:
 July 29, 2016

 Received:
 August 1, 2016

Dear Mr. Shah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing

(21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<u>http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm</u> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson Director Division of Orthopedic Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

<pre>#-The-Counter Use (21 CFR 801 Subpart C) #-The-Counter Use (21 CFR 801 Subpart C) # Work Reduction Act of 1995. * STAFF EMAIL ADDRESS BELOW.* # Staff # Burden, to: # an Services #</pre>	r both, as applicable) ption Use (Part 21 CFR 801 Subpart D) CONTINUE ON A SEPARATE PAGE Instruction applies only to requirements of the Paperw r SEND YOUR COMPLETED FORM TO THE PRA SI re for this collection of information is estimated to aver instructions, search existing data sources, gather and collection, including suggestions for reducing this to pepartment of Health and Human Food and Drug Administration Office of Chief Information Office Paperwork Reduction Act (PRA) PRAStaff@fda.hhs.gov roy may not conduct or sponsor, and a person is not re	Type of Use (Select one or I ☐ Prescripti ★DO NOT : The burden time time to review the c of this informatio <i>"An agency</i>
In-The-Counter Use (21 CFR 801 Subpart C) IF NEEDED. Work Reduction Act of 1995. STAFF EMAIL ADDRESS BELOW.* Staff Emaintain the data needed and complete this burden estimate or any other aspect s burden, to: an Services	ar both, as applicable) ption Use (Part 21 CFR 801 Subpart D) continue on Section applies only to requirements of the Paperw This section applies only to requirements of the Paperw re for this collection of information is estimated to aver instructions, search existing data sources, gather and collection, including suggestions for reducing this to Department of Health and Human Food and Drug Administration Office of Chief Information Office Paperwork Reduction Act (PRA) PRAStaff@da.hhs.gov	Type of Use (Select one or I ⊠ Prescripti *DO NOT s The burden time time to review in: and review the c of this informatio
r-The-Counter Use (21 CFR 801 Subpart C) IF NEEDED.	ar both, as applicable) ption Use (Part 21 CFR 801 Subpart D) Over- CONTINUE ON A SEPARATE PAGE I	Type of Use (Select one or I
ar-The-Counter Use (21 CFR 801 Subpart C)	ar both, as applicable) ption Use (Part 21 CFR 801 Subpart D)	Type of Use <i>(Select one or i</i> ⊠ Prescripti
		Tubo Af Lino AGalactera
or conditions including sacroiliac joint	stem <i>ibe)</i> ion System is intended for sacroiliac joint fusion for rative sacroiliitis.	K161210 Device Name RIALTO TM SI Fusion Syste Indications for Use (<i>Describ</i> The RIALTO TM SI Fusio disruptions and degenera
See PRA Statement below.	Indications for Use	
Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017	RTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration	DEPAR

PSC Publishing Services (301) 443-6740 EF	4	4) Page 1 of	FORM FDA 3881 (8/1
oond to, a collection of er."	rson is not required to resp currently valid OMB numbe	gency may not conduct or sponsor, and a per information unless it displays a c	"An a
ADDRESS BELOW.* per response, including the data needed and complete data needed and complete timate or any other aspect	THE PRA STAFF EMAIL , nated to average 79 hours j , gather and maintain the d s regarding this burden esti educing this burden, to: educing this burden, to: h and Human Services inistration mation Officer n Act (PRA) Staff pov	NOT SEND YOUR COMPLETED FORM TO 1 time for this collection of information is estim lew instructions, search existing data sources. the collection of information. Send comments the collection, including suggestions for re Department of Health Food and Drug Adm Office of Chief Inform Paperwork Reductio PRAStaff@fda.hhs.g	*DO I The burder time to revi and review of this infor
n Act of 1995.	ATE PAGE IF NEEDED.	CONTINUE ON A SEPARA This section applies only to requirements o	
Jse (21 CFR 801 Subpart C)	Over-The-Counter U	ne or both, as applicable) scription Use (Part 21 CFR 801 Subpart D)	Type of Use (Select o
SYSTEM and hard tissue and bone, and , Sternotomy, and General surgical 's and working end attachments I in placement	E IPC TM POWEREASE TM S illing and sawing of soft a dic, Arthroscopic, Spinal, procedures. It is also used	ABLE INSTRUMENTS FOR USE WITH THI escribe) dicated for the incision/cutting, removal, dri rosurgical (Cranial, Craniofacial), Orthoped LEASE TM System is indicated for drilling, tr y, including open and minimally invasive p , posts and rods.	510(k) Number (<i>if kno</i> K161210 Device Name MEDTRONIC REUS IPC TM System is ind biomaterials in Neu procedures. The IPC TM POWER during spinal surger or cutting of screws
ee PRA Statement below.	Se	Indications for Use	
orm Approved: OMB No. 0910-0120 xpiration Date: January 31, 2017	ICES Fo	EPARTMENT OF HEALTH AND HUMAN SERV Food and Drug Administration	D
Page 2 of 3	-		

PSC Publishing Services (301) 443-6740 EF	+	14) Page Fof	FORM FDA 3881 (8
respond to, a collection of umber."	son is not required to urrently valid OMB nu	agency may not conduct or sponsor, and a per information unless it displays a c	"An
ίλ	n and Human Service nistration lation Officer Act (PRA) Staff ov	Department of Health Food and Drug Admi Office of Chief Inform Paperwork Reductior PRAStaff@fda.hhs.g	
AL ADDRESS BELOW." ours per response, including the the data needed and complete n estimate or any other aspect p:	the PKA STAFF EM ated to average 79 ho gather and maintain regarding this burden, to ducing this burden, to	in time for this collection of information is estim view instructions, search existing data sources, v the collection of information. Send comments rmation collection, including suggestions for re	The burde time to re and revie of this inf
ALL ADDRESS RELOW *	the Paperwork Redu	This section applies only to requirements of	* DO
ĒD.	TE PAGE IF NEED	CONTINUE ON A SEPARA	
nter Use (21 CFR 801 Subpart C)	Over-The-Cour	one or both, as applicable) escription Use (Part 21 CFR 801 Subpart D)	Type of Use (Select
THE STEALTHSTATIONTM AND IPCTM aration and placement of Medtronic structures in either open, or minimally esigned for use with the use of stereotactic surgery may be ong bone, or vertebra can be identified f the anatomy. Medtronic Navigated	VTS FOR USE WITH sed during the prepa ocating anatomical s its are specifically de ndition in which the such as a skull, a lo gitized landmarks of EREASE TM System	own) IGATED MANUAL REUSABLE INSTRUMEN STEMS Describe) ed Reusable Instruments are intended to be u al surgery to assist the surgeon in precisely le ss. Medtronic Navigated Reusable Instrumen ystem, which is indicated for any medical co here reference to a rigid anatomical structure MR-based model, fluoroscopy images, or dij nts are also compatible with the IPC TM POW	510(k) Number (<i>if k</i> r K161210 Device Name MEDTRONIC NAV POWEREASETM SV Screws during spir invasive, procedur StealthStation TM S appropriate, and w relative to a CT or Reusable Instrume
See PRA Statement below.		Indications for Use	
Form Approved: OMB No. 0910-0120	CES	EPARTMENT OF HEALTH AND HUMAN SERVI Food and Drug Administration	
Page 3 of 3			

RIALTO[™] SI Fusion System 510(k) SUMMARY July 2016

I. Submitter Medtronic Sofamor Danek USA, Inc. 1800 Pyramid Place Memphis, TN 38132 (901)396-3133

> Contact Ankit K. Shah Senior Regulatory Affairs Specialist

Date Prepared July 28, 2016

- II. Device
 - Name of Device RIALTOTM SI Fusion System, Medtronic Reusable Instruments for Use with the IPCTM PowereaseTM System, and Medtronic Navigated Manual Reusable Instruments for use with the StealthStationTM and IPCTM PowereaseTM Systems
 - Common Name Smooth or threaded metallic bone fixation fastener (Implant) Taps, Drivers and Drills (Instrument)
 - Classification Name Sacroiliac Joint Fixation OUR (For Implant)

Stereotaxic Instrument - OLO (For Navigated Instruments)

Surgical Instrument Motors and Accessories/Attachments - HWE

	(For Instruments Compatible with the IPC®
	POWEREASE® System)
Classification	Class II (Implant and Instruments)
Product Codes	OUR (Implants) 21 CFR 888.3040 OLO (Navigated Instruments) 21 CFR 882.4560
	HWE (IPC® POWEREASE® Compatible Instruments) 21 CFR 878.4820
Predicates	There are 6 Predicates.
	MSB Sacroiliac Joint Fusion Device K110472, S.E. 05/29/2012 (Primary Predicate)
	<u>iFuse Implant System</u> K131405, S.E. 10/16/2013
	<u>IPC® POWEREASE® System</u> K111520, S.E. 10/26/2011
	Navigated CD Horizon Solera Screwdriver/Taps K140454, S.E. 05/22/2014
	Navigated CD HORIZON® SOLERA TM Screwdrivers, CD HORIZON® SOLERA TM Taps, CD HORIZON® SOLERA TM Iliac Taps, CD HORIZON® LEGACY TM Taps K124004, S.E. 03/22/2013
	Navigated Disc Prep Instruments and CAPSTONE Trials K150231, S.E. 06/16/2015
	The predicates have not been subject to a design related recall.

III. Product Description

RIALTOTM SI Fusion System

The subject RIALTO[™] SI Fusion System consists of cannulated devices available in various lengths, used to provide stabilization when fusion of the sacroiliac joint is desired. Autograft and/or allograft may be placed in conjunction with the RIALTO[™] SI Fusion System. The RIALTO[™] implant are made using Titanium Alloy and are 40mm-60mm in length with a diameter of 12mm. This device may be implanted via a minimally invasive approach using fluoroscopy or navigated instruments compatible with Medtronic StealthStation® and IPC® POWEREASE®.

Medtronic Reusable Instruments Only Compatible with the IPC® POWEREASE® System

The subject Medtronic Reusable Taps, Drivers and Drills are spine preparation instruments made of high grade stainless steel and are designed specifically for subject RIALTOTM SI Fusion System. The subject Taps, Drivers and Drills are compatible with Medtronic's IPC® POWEREASE® System may be connected to the IPC® POWEREASE® handpiece. The subject Taps, Drivers and Drills can be used manually using existing Medtronic Class I Exempt quick connect handles in place of the IPC® POWEREASE® handpiece.

<u>Medtronic Reusable Instruments Compatible with the STEALTHSTATION® System</u> and IPC® POWEREASE® System

The subject Medtronic Navigated Reusable Taps, Drivers and Drills are spine preparation instruments made of high grade stainless steel. These instruments were specifically designed for use with subject RIALTOTM SI Fusion System in procedures where the use of stereotactic surgery may be appropriate. The subject Taps, Drivers and Drills are also compatible with Medtronic's IPC® POWEREASETM System when connected to the POWEREASETM handpiece or may be used manually with existing Medtronic Class I Exempt quick connect handles in place of the IPC® POWEREASE® handpiece or the NavLockTM tracker.

IV. Indications for Use:

RIALTOTM SI Fusion System

The RIALTO[™] SI Fusion System is intended for sacroiliac joint fusion for conditions including sacroiliac joint disruptions and degenerative sacroiliitis.

Medtronic Reusable Instruments For Use with the IPC® POWEREASE® System

IPC® System is indicated for the incision/cutting, removal, drilling and sawing of soft and hard tissue and bone, and biomaterials in Neurosurgical (Cranial, Craniofacial), Orthopedic, Arthroscopic, Spinal, Sternotomy, and General surgical procedures.

The IPC® POWEREASE® System is indicated for drilling, tapping and driving screws and working end attachments during spinal surgery, including open and minimally invasive procedures. It is also used in placement or cutting of screws, posts and rods.

Medtronic Navigated Reusable Instruments For Use with STEALTHSTATION® and IPC® POWEREASETM Systems

Medtronic Navigated Reusable Instruments are intended to be used during the preparation and placement of Medtronic screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open, or minimally invasive, procedures. Medtronic Navigated Reusable Instruments are specifically designed for use with the StealthStation[®] System, which is indicated for any medical condition in which stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR-based model, fluoroscopy images, or digitized landmarks of the anatomy. Medtronic Navigated Reusable Instruments are also compatible with the IPC[®] POWEREASE[™] System.

V. Comparison of Technological Characteristics

The primary predicate for the RIALTO[™] SI Fusion System is the predicate SI-FIX Sacroiliac Joint Fusion System K110472, S.E. 05/29/2012 (Primary Predicate).

The subject RIALTO[™] SI Fusion System implants have the same or similar intended use, fundamental scientific technology, material, and indications as the following FDA cleared predicates K110472 (S.E. 05/29/2012) and K131405 (S.E. 10/16/2013).

The subject RIALTOTM SI Fusion System instruments have the same or similar intended use, fundamental scientific technology, material, and indications as the following FDA cleared predicates K111520 (S.E. 10/26/2011), K110472 (S.E. 05/29/2012), K140454 (S.E. 05/22/2014), K124004 (S.E. 03/22/2013) and K150231 (S.E. 06/16/2015)

VI. Performance Data

The following performance data are provided in support of the substantial equivalence determination.

Biocompatibility

Identical to the primary predicate devices the implants in the subject RIALTO[™] SI Fusion System are made using Titanium alloy.

The non-sterile instruments are manufactured using stainless steel and are identical to the materials used for the instruments cleared under predicate submissions.

The titanium alloy and stainless steel material used for the subject RIALTO[™] SI Fusion System implants and instruments have a long clinical history of safe and effective use in similar commercially available medical devices. Therefore, no additional biocompatibility testing is required.

Mechanical Testing

In accordance with the *Guidance for Industry and FDA Staff - Spinal System* 510(k)'s, Medtronic has evaluated the subject devices to demonstrate substantial equivalence to the predicate devices.

Design verification testing for the subjectimplants was completed in accordance with

- ASTM F543-13 Standard Specification and Test Methods for Metallic Medical Bone Screws
- ASTM F2193-02 (2007) Standard Specifications and Test Methods for Components Used in the Surgical Fixation of the Spinal Skeletal System

The tests completed were:

- Push-Out
- Torque–to-Fail
- Four Point Bend
- Shear Testing

The verification testing on the instruments included Navigation accuracy on the instruments compatible with StealthStationTM and IPCTM PowereaseTM.

The subject devices met the pre-determined acceptance criteria for all tests. Therefore, design verification testing determined that the subject screws and instruments are substantially equivalent to the predicate devices.

Design validation testing was performed that demonstrated that the subject instruments performed as intended and to validate the surgical approach.

Non-Pyrogenicity Endotoxin Testing

The bacterial endotoxin test, also known as Limulus Amebocyte Lysate (LAL) test, was performed utilizing worst case subject RIALTOTM SI Fusion implants to verify that the subject implants meet the 20 endotoxin units (EU)/device pyrogen limit specification. Testing was successfully performed and it was confirmed that the subject implants meet the 20 EU/device testing limit for general medical devices that are implanted as outlined in ANSI/AAMI ST72, Bacterial endotoxins – Test methods, routine monitoring, and

alternatives to batch testing and USP <161>, Transfusion and Infusion Assemblies and Similar Medical Devices.

VII. Conclusions

Based on the test results and additional supporting information provided in this premarket notification, the subject devices demonstrated substantial equivalence to the legally marketed predicate devices.