



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

September 22, 2016

VR Medical Technology Company  
% Korina A. Akhondzadeh  
Regulatory Consultant to VR Medical Technology  
KARA & Associates  
6965 El Camino Real, Suite 105-428  
Carlsbad, CA 92009

Re: K162188  
Trade/Device Name: Enteral Syringe with ENFit® Connector (10 ml to 60 ml),  
Enteral Syringe with ENFit® Low Dose Tip (1 ml To 5 ml)  
Regulation Number: 21 CFR§ 876.5980  
Regulation Name: Gastrointestinal Tube and Accessories  
Regulatory Class: II  
Product Code: PNR  
Dated: August 1, 2016  
Received: August 4, 2016

Dear Korina A. Akhondzadeh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Joyce M. Whang -S**

for Benjamin R. Fisher, Ph.D.  
Director  
Division of Reproductive, Gastro-Renal,  
and Urological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K162188

Device Name

Enteral Syringe with ENFit® Connector (10 ml to 60 ml)

Enteral Syringe with ENFit® Low Dose Tip (1 ml to 5 ml)

Indications for Use (Describe)

The device is indicated for use as a dispenser, a measuring device and a fluid transfer device. It is used to deliver fluids into the body enterally. It is intended to be used in clinical or home care settings by users ranging from clinicians to laypersons (under the supervision of a clinician) for patients in all age groups.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

This 510(k) summary is being submitted in accordance with the requirements of 21 CFR 807.92.

### I. SUBMITTER

VR Medical Technology  
90 Gao Xin Road, ZhouZhuang  
Kunshan 215325  
CHINA  
Tel: +86 512 5720 0660 (China)

Contact: Lee Pan, Ph.D., CEO  
Email: [lpn@vrmedical.net](mailto:lpn@vrmedical.net)

### Official Correspondent

Korina A. Akhondzadeh  
KARA & Associates - Regulatory Consultant to VR  
Medical Technology  
6965 El Camino Real, Suite 105-428  
Carlsbad, CA 92009  
Tel: 760-798-9642  
Fax: 760-798-9643

Email [korina@karaandassociates.com](mailto:korina@karaandassociates.com)

**Date Prepared:** August 1, 2016

### II. DEVICE

Device Name: Enteral Syringe with ENFit Connector (10 ml to 60 ml) and Enteral Syringe with ENFit Low Dose Tip (1 ml to 5 ml)  
Common/Usual Name: Enteral Syringe  
Classification Name: Gastrointestinal tube and accessories  
Product Code: PNR  
Class: II

### III. PREDICATE DEVICE

#### Predicate Device

Neoconnect Oral/Enteral Syringes with ENFit Connector (12 mL to 100 mL) and NeoConnect Low Dose Tip Oral/Enteral Syringes with ENFit Connector (0.5 mL to 6 mL), K161039

This predicate device has not been subject to a recall.

#### **IV. DEVICE DESCRIPTION**

The Enteral Syringe with ENFit Connector (10 ml to 60 ml) and the Enteral Syringe with ENFit Low Dose Tip (1 ml to 5 ml) are both sterile or non-sterile, standard piston style enteral syringes with an integrated female ENFit compatible connector compliant to the requirements of ISO DIS 80369-3, First Edition, Small-bore Connectors for Liquids and Gases in Healthcare Applications – Part 3: Connectors for Enteral Applications and designed to be compatible with male ENFit connectors.

The Enteral Syringe with ENFit Connector has capacities of 10, 20, 30 and 60 ml and is available in two different connector configurations: Centerline and Offset.

The Enteral Syringe with ENFit Low Dose Tip is a low dose syringe and has capacities of 1, 3 and 5 ml. The ENFit Low Dose tip is a modification of the ENFit connector designed for the delivery of low doses of fluids for enteral feeding.

#### **V. INDICATIONS FOR USE**

The device is indicated for use as a dispenser, a measuring device and a fluid transfer device. It is used to deliver fluids into the body enterally. It is intended to be used in clinical or home care settings by users ranging from clinicians to laypersons (under the supervision of a clinician) for patients in all age groups.

#### **VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE**

Both the subject Enteral Syringe with ENFit Connector (10 ml to 60 ml) and Enteral Syringe with ENFit Low Dose Tip (1 ml to 5 ml) devices and predicate Oral/Enteral Syringes with ENFit Connector (12 ml to 100 ml) and Low Dose Tip Oral/Enteral Syringes with ENFit Connector (0.5 ml to 6 ml) devices are a standard piston-style syringe with an integral female ENFit Connector. The ENFit Connector for both the VR Medical and NeoMed devices were designed to be compliant with ISO DIS 80369-3, First Edition, Small-bore Connectors for Liquids and Gases in Healthcare Applications – Part 3: Connectors for Enteral Applications.

Both the Enteral Syringe with ENFit Low Dose Tip (1 ml to 5 ml) and predicate Low Dose Tip Oral/Enteral Syringes with ENFit Connector (0.5 ml to 6 ml) incorporate the same ENFit low dose tip design to ensure accuracy when dispensing small volumes.

The ENFit connectors on both the subject and predicate devices are compatible only with enteral access devices or accessories having ENFit compliant or compatible male connectors to prevent misconnections to non-enteral devices. Both the subject and predicate devices are available in both sterile and non-sterile versions.

## VII. PERFORMANCE DATA

The following performance testing was conducted on the Enteral Syringe with ENFit Connector (10 ml to 60 ml) and the Enteral Syringe with ENFit Low Dose Tip (1 ml to 5 ml).

### Biocompatibility Testing

- Cytotoxicity per ISO 10993-5
- Sensitization per ISO 10993-10
- Irritation per ISO 10993-10

### Performance Testing

Tests per ISO 7886-1:

- Cleanliness
- Lubricant
- Position of scale
- Tolerance on graduated capacity
- Dead space
- Fit of piston in barrel
- Slide force test
- Freedom from air and liquid leakage
- Limits for acidity or alkalinity
- Limits of extractable metals

Tests per ISO DIS 80369-3:

- Dimensional verification
- Positive pressure liquid leakage
- Stress cracking
- Resistance to separation from axial load
- Resistance to separation from unscrewing
- Resistance to overriding
- Disconnection by unscrewing

Printing Firmness

For the Enteral Syringe with ENFit Connector

- Risk Assessment per ISO 14971

For the Enteral Syringe with ENFit Low Dose Tip

- Low Dose Tip Misconnection Risk Management Reports
- Usability Study for Low Dose Tip Design Feature

## VIII. CONCLUSIONS

The Enteral Syringe with ENFit Connector (10 to 60 ml) and the Enteral Syringe with ENFit Low Dose Tip (1 to 5 ml) are substantially equivalent to the predicate Oral/Enteral Syringes with ENFit Connector (12 ml to 100 ml) and Low Dose Tip Oral/Enteral Syringes with ENFit Connector (0.5 ml to 6 ml) devices.