



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

October 18, 2016

Covidien LLC  
Ms. Trang Huynh  
Principal Regulatory Affairs Specialist  
60 Middletown Avenue  
North Haven, Connecticut 06473

Re: K162584  
Trade/Device Name: VersaOne™ Bladeless Trocar  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and Accessories  
Regulatory Class: Class II  
Product Code: GCJ  
Dated: October 11, 2016  
Received: October 12, 2016

Dear Ms. Huynh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Jennifer R. Stevenson -A**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.

Director

Division of Surgical Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K162584

Device Name

VersaOne™ Bladeless Trocar

Indications for Use (Describe)

The VersaOne™ bladeless trocars are intended for use in a variety of gynecologic, general, thoracic and urologic endoscopic procedures to create and maintain a port of entry.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*



## 510(k) Summary

This 510(k) summary information is submitted in accordance with the requirements of the Safe Medical Devices Act (SMDA) of 1990. The contents of the 510(k) summary have been provided in conformance with 21 CFR §807.92.

SUBMITTER:	Covidien llc 60 Middletown Avenue North Haven, CT 06473 USA
CONTACT PERSON:	Trang Huynh Principal Regulatory Affairs Specialist Phone: (203) 492-7473 Fax: (203) 492-5029 Email: <a href="mailto:trang.huynh27@medtronic.com">trang.huynh27@medtronic.com</a>
DATE PREPARED:	September 16, 2016
PRODUCT CODE:	GCJ
REGULATION NUMBER:	78
CLASS CODE:	Class II, Pursuant to 21 CFR 876.1500
CLASSIFICATION NAME:	Endoscope and Accessories
TRADE /PROPRIETARY NAME:	VersaOne™ Bladeless Trocar
COMMON/USUAL NAME:	Surgical Trocar
PREDICATE DEVICES:	VersaOne™ Bladeless Trocar (K151548) Autosuture™ Versaport™ Plus Bladeless Trocar (K081169)
INTENDED USE:	The VersaOne™ bladeless trocars are intended for use in a variety of gynecologic, general, thoracic and urologic endoscopic procedures to create and maintain a port of entry.



DEVICE  
DESCRIPTION:

VersaOne™ Bladeless Trocar 8mm is available in two lengths (long (150mm) and standard (100mm)); both are 8mm in diameter and contain transparent fixation cannulas. VersaOne™ Bladeless Trocar 8mm consists of an obturator component and a cannula assembly. The obturator has a dolphin nose (conical) shaped bladeless tip. The cannula assembly is composed of a seal system, trocar housing, transparent cannula sleeve, and a stopcock. The seal system prevents loss of pneumoperitoneum when instruments are inserted or withdrawn. The seal system accommodates instruments ranging from 5mm to 8mm in diameter. The 8mm trocar housing is removable for specimen retrieval. The cannula sleeve has a bevel shape at the distal end to reduce penetration forces during insertion and transparent for better visualization. The stopcock valve is for insufflation and desufflation. The device is supplied sterile single-use. The device is sterilized via Ethylene Oxide (EO) sterilization method. The packaging configuration is one device packed in a sealed Nylon / Tyvek breather pouch, six pouches per display box, and both pouch and display box have appropriate labeling.

SUMMARY  
COMPARING THE  
TECHNOLOGICAL  
CHARACTERISTIC  
S OF THE  
SUBJECT AND  
PREDICATE  
DEVICES:

Modifications to the design of the current VersaOne™ Bladeless Trocar (K151548) include diameter change to the cannula and obturator components to accommodate an 8mm instrument.

Sterilization assessment was performed to confirm EO sterilization equivalency to the predicate device with respect to sterilization processes per ISO 11135 requirements and EO residuals per ISO 10993-7 requirements.

Shelf life assessment was performed per Covidien standard operating procedures to confirm product performance over time.

Biocompatibility testing was performed to confirm that all components are comprised of materials that are in accordance with ISO 10993-1 for their intended patient contact profile.

Performance testing including bench top and animal was performed to compare the functional performance of the subject device and the predicate device.

- Bench top testing that supports the intended use of this device includes:
  - Leak rate
  - Insertion force & Removal force
  - Penetration force & Fixation force
- Animal testing that supports the intended use of the device includes:
  - Penetration force & Fixation force

CONCLUSION: The results of testing demonstrated the modified VersaOne™ Bladeless Trocar 8mm is substantially equivalent to the legally marketed device.