



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

Accuray Incorporated  
% Mr. Keith Picker  
Senior Regulatory Affairs Specialist  
1240 Deming Way  
Madison, WI 53717

July 25, 2017

Re: K171837

Trade/Device Name: TomoTherapy Treatment Delivery System with iDMS  
Regulation Number: 21 CFR 892.5050  
Regulation Name: Medical charged-particle radiation therapy system  
Regulatory Class: Class II  
Product Code: IYE  
Dated: June 19, 2017  
Received: June 20, 2017

Dear Keith Picker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

A handwritten signature in blue ink that reads "Michael D. O'Hara". The signature is written in a cursive style. To the right of the signature, the word "For" is printed in a small, black, sans-serif font.

Robert A. Ochs, Ph.D.

Director

Division of Radiological Health

Office of In Vitro Diagnostics

and Radiological Health

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K171837

Device Name

TomoTherapy Treatment Delivery System with iDMS

Indications for Use (Describe)

The TomoTherapy Treatment Delivery System with iDMS is indicated for the delivery of radiation therapy, stereotactic radiotherapy or stereotactic radiosurgery to tumors or other targeted tissues anywhere in the body under the direction of a licensed medical practitioner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

---

## Section 8 510(k) Summary

### Submitter

Accuray Incorporated  
1209 Deming Way  
Madison, WI 53717-1954  
Phone: 608-824-2800  
Fax: 608-824-2996

Contact: Keith Picker  
Date Prepared: June 19, 2017

### Device Identification

Device Name: TomoTherapy Treatment Delivery System with iDMS  
Trade & Brand Names: TomoTherapy Treatment Delivery System with iDMS  
Common Name: Radiation Therapy System  
Regulation Number: 21 CFR 892.5050  
Regulation Name: Medical charged-particle radiation therapy system  
Regulatory Class: Class II  
Product Code: IYE

### Predicate Device

TomoTherapy Treatment System (K121934)

### Device Description

The TomoTherapy Treatment Delivery System with iDMS is a radiation therapy delivery system that provides Image Guided Radiation Therapy (IGRT) using integral megavoltage CT imaging capabilities and delivers helical (rotational) and fixed-angle (non-rotational) radiation therapy to tumors and other targeted tissues.

The TomoTherapy Treatment Delivery System with iDMS is a modified version of the predicate TomoTherapy Treatment System (last cleared on K121934). The TomoTherapy Treatment Delivery System with iDMS is the same as the predicate device minus the Data Management System and Treatment Planning Station, but substituting the Treatment Delivery Console (TDC) from the Radixact Treatment Delivery System (last cleared on K161146) in place of the predicate Operator Station. The TomoTherapy Treatment Delivery System with iDMS is designed to deliver radiation therapy treatment plans generated on planning systems such as the Accuray Precision™ Treatment Planning System (last cleared on K171086) and stored on data management systems such as the Accuray iDMS™ Integrated Data Management System (last cleared on K161144).

The TomoTherapy Treatment Delivery System with iDMS is a prescription device that delivers radiation in accordance with a physician approved plan. As with the predicate, the TomoTherapy Treatment Delivery System with iDMS does not diagnose disease, recommend treatment regimens or quantify treatment effectiveness. Accordingly, it is not intended for diagnostic use.

### **Intended Use**

The TomoTherapy Treatment Delivery System with iDMS is intended to be used for the delivery of radiation therapy, stereotactic radiotherapy or stereotactic radiosurgery to tumors or other targeted tissues. The megavoltage x-ray radiation is delivered using rotational, non-rotational, intensity modulated (IMRT), or non-modulated (non-IMRT/three dimensional conformal) treatment techniques and using image-guided (IGRT) or non-image-guided workflows in accordance with the physician approved plan.

### **Indications for Use**

The TomoTherapy Treatment Delivery System with iDMS is indicated for the delivery of radiation therapy, stereotactic radiotherapy or stereotactic radiosurgery to tumors or other targeted tissues anywhere in the body under the direction of a licensed medical practitioner.

While the indications for use of the TomoTherapy Treatment Delivery System with iDMS are stated slightly differently from those of the predicate, the differences are not critical to the intended use of the devices nor do they affect the safety and/or effectiveness of the new device as compared to the predicate device. Both devices provide the same types of radiation therapy. In fact, the intended use and indications for use statements for the TomoTherapy Treatment Delivery System with iDMS (shown above) are the same as for the Radixact Treatment Delivery System (last cleared on K161146), except for the substitution of the device name.

### **Technological Characteristics**

The TomoTherapy Treatment Delivery System with iDMS has imaging and treatment capabilities equivalent to those of the predicate TomoTherapy Treatment System. In fact, the TomoTherapy Treatment Delivery System with iDMS uses the same CT style gantry, patient couch and power distribution unit (PDU) as the predicate. Further, the clinical workflow is the same as that of the predicate TomoTherapy Treatment System.

The TomoTherapy Treatment Delivery System with iDMS and the predicate device employ the same fundamental scientific principles, and have substantially equivalent intended uses, principles of operation, performance specifications and technological characteristics.

The main differences between the predicate and the TomoTherapy Treatment Delivery System with iDMS are:

- 1) The predicate includes integral treatment planning and data management subsystems, whereas the TomoTherapy Treatment Delivery System with iDMS is strictly a radiation treatment delivery device.
- 2) The TomoTherapy Treatment Delivery System with iDMS uses the Treatment Delivery Console (TDC) of the Radixact Treatment Delivery System (last cleared on K161146) in place of the Operator Station (OS) from the predicate, to which the TDC is functionally equivalent.
- 3) The TomoTherapy Treatment Delivery System with iDMS includes software changes (due to the use of the TDC from the Radixact Treatment Delivery System) to allow it to connect and work compatibly with the Accuray Precision™ Treatment Planning System (last cleared on K171086) and iDMS™ Integrated Data Management System (K161144).

Where there are technological differences between the TomoTherapy Treatment Delivery System with iDMS and the predicate device, those differences do not raise different questions of safety or effectiveness.

A table comparing the predicate cleared on K121934 and the TomoTherapy Treatment Delivery System with iDMS is presented below:

<b>General Characteristics</b>	<b><u>Predicate Device</u> TomoTherapy Treatment System (K121934)</b>	<b><u>Subject Device</u> TomoTherapy Treatment Delivery System with iDMS</b>
<b>System Configuration</b>	Radiation delivery system integrated with data management system and planning system	Stand alone radiation delivery system (does not include data management system or planning system)

<b>General Characteristics</b>	<b><u>Predicate Device</u> TomoTherapy Treatment System (K121934)</b>	<b><u>Subject Device</u> TomoTherapy Treatment Delivery System with iDMS</b>
<b>Vault</b>  Min. Room Dimensions Height* Width* Length*  Device Dimensions (gantry and couch) Height* Width* Length* Weight  Environment Line Voltage Ambient Room Temperature Relative Humidity	270 cm 462 cm 596 cm  252 cm 280 cm 466 cm 4,943 kg  380-480V, 3-Phase  68-75 °F (20-24 °C) 30%-60%, non-condensing	Identical to predicate** Identical to predicate ** Identical to predicate **  Identical to predicate ** Identical to predicate ** Identical to predicate ** Identical to predicate **  Identical to predicate  Identical to predicate Identical to predicate
<b>Gantry Mechanical Features</b>  Degrees of Rotation  Direction of Rotation  Bore Size  Speed of Rotation Treatment Imaging	Continuous rotation around Y-axis (axes per IEC 61217)  Clockwise (as viewed from the foot of the patient couch)  85 cm diameter  1 to 5 RPM 6 RPM	Identical to predicate  Identical to predicate  Identical to predicate  Identical to predicate Identical to predicate
<b>Radiation Delivery Modes</b>	Helical	Identical to predicate
	Direct	Identical to predicate

<b>General Characteristics</b>	<b><u>Predicate Device</u> TomoTherapy Treatment System (K121934)</b>	<b><u>Subject Device</u> TomoTherapy Treatment Delivery System with iDMS</b>
<b>Photon Beam</b>  Accelerator Type  RF Source  Nominal Energy  Fixed Field Size  Dynamic Field Size  Dose Rate	Standing wave  Magnetron  6 MV (single energy)  1.0 cm x 40 cm 2.5 cm x 40 cm 5.0 cm x 40 cm  1.0 - 2.5 cm x 40 cm 1.0 - 5.0 cm x 40 cm  850 cGy/min	Identical to predicate  Identical to predicate  Identical to predicate  Identical to predicate Identical to predicate Identical to predicate  Identical to predicate Identical to predicate
<b>Collimation</b>	Primary collimation, jaws and multi-leaf collimator (MLC)	Identical to predicate
<b>Imaging</b>  Field of View  Dose per MVCT image (typical)  Slice spacing  Spatial Resolution	39 cm diameter  0.5 - 3.0 cGy  1, 2, 3, 4 and 6 mm reconstruction intervals  1.6 mm spatial resolution	Identical to predicate  Identical to predicate  Identical to predicate  Identical to predicate

<b>General Characteristics</b>	<b><u>Predicate Device</u> TomoTherapy Treatment System (K121934)</b>	<b><u>Subject Device</u> TomoTherapy Treatment Delivery System with iDMS</b>
<p><b>Laser System</b></p> <p>Stationary</p> <p>Moveable (for patient positioning and registration)</p>	<p>Green lasers, identify virtual and actual isocenter</p> <p>Red lasers, offset from virtual isocenter</p>	<p>Identical to predicate</p> <p>Identical to predicate</p>
<p><b>Patient Couch</b></p> <p>Biocompatibility</p> <p>Motion</p> <p>X-axis</p> <p>Y-axis</p> <p>Z-axis</p>	<p>Carbon-fiber top</p> <p>Independent of other axes</p> <p>Coupled with Z-axis (via couch Cobra motion)</p> <p>Coupled with Y-axis (via couch Cobra motion)</p>	<p>Identical to predicate</p> <p>Identical to predicate</p> <p>Identical to predicate</p> <p>Identical to predicate</p>
<p><b>Power Distribution</b></p> <p>Isolation</p> <p>UPS for Data Back-up</p>	<p>Transformer</p> <p>Provided</p>	<p>Identical to predicate</p> <p>Identical to predicate</p>
<p><b>Operator Station</b></p>	<p>User interface to system functions (i.e., patient and procedure selection, and procedure delivery)</p>	<p>Uses the Treatment Delivery Console (TDC), cleared as part of the Radixact Treatment Delivery System (K161146), which is functionally equivalent to the predicate Operator Station</p>

<b>General Characteristics</b>	<b><u>Predicate Device</u> TomoTherapy Treatment System (K121934)</b>	<b><u>Subject Device</u> TomoTherapy Treatment Delivery System with iDMS</b>
<b>Machine Control Software</b>  Data Interfaces Operator Station	Controls radiation delivery and positioning systems (referred to as the RDS – Radiation Delivery System)  Provides measurements and status during operation	Functionally equivalent to predicate (referred to as the ECS – Embedded Controls Subsystem)  Functionally equivalent to predicate
<b>Database</b>	Integrated database used for gathering operational data and storage of procedure data	Works with functionally-equivalent external database such as the Accuray iDMS™ Integrated Data Management System (K161144)
<b>Safety Features</b>	Interlock Subsystems  Data integrity checking	Functionally equivalent to predicate  Functionally equivalent to predicate

\* Dimensions are rounded to the nearest centimeter.

\*\* Information Source: TomoTherapy Treatment Delivery System with iDMS Technical Specifications

### **Performance Data**

Results of verification and validation testing confirm that the TomoTherapy Treatment Delivery System with iDMS conforms to design specifications and meets the needs of the intended users. Additionally, testing performed by an independent certified testing laboratory demonstrates the device complies with the requirements of applicable FDA recognized consensus safety standards for radiation therapy medical devices. These test results demonstrate that the TomoTherapy Treatment Delivery System with iDMS is as safe and effective, and performs as well as the predicate device.

No animal or clinical tests were required to establish substantial equivalence with the predicate device.

## **Conclusion**

The TomoTherapy Treatment Delivery System with iDMS is substantially equivalent to the predicate device. The intended use, principles of operation, performance specifications and technological characteristics of the TomoTherapy Treatment Delivery System with iDMS are substantially equivalent to those of the predicate device. Minor differences do not raise different questions of safety and effectiveness of the TomoTherapy Treatment Delivery System with iDMS in comparison to the cleared predicate device. Further, performance data demonstrate that the TomoTherapy Treatment Delivery System with iDMS has substantially equivalent safety and performance characteristics in comparison to the predicate device. Accordingly, the TomoTherapy Treatment Delivery System with iDMS is substantially equivalent to the predicate device.