



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Zimmer Biomet Spine, Inc.
Ms. Kelly Stratton
Regulatory Affairs Specialist
10225 Westmoor Drive
Westminster, Colorado 80021

July 14, 2017

Re: K171907

Trade/Device Name: Vitality® Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral pedicle screw system
Regulatory Class: Class II
Product Code: NKB, KWP, KWQ
Dated: June 23, 2017
Received: June 26, 2017

Dear Ms. Stratton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K171907

Device Name

Vitality® Spinal Fixation System

Indications for Use (Describe)

The Vitality® Spinal Fixation System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients and for adolescent patients.

These devices are indicated as an adjunct to fusion for all of the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic sites), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e. scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, pseudoarthrosis and/or failed previous fusion. When used as an adjunct to fusion, the Vitality Spinal Fixation System is intended to be used with autograft and/or allograft.

In addition the Vitality® Spinal Fixation System is intended for treatment of severe spondylolisthesis (Grade 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and or ilium with removal of the implant after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/ilium.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the Vitality® System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The Vitality® System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Assembly of a spinal fixation construct using the Vitality® Spinal Fixation System may include additionally any Instinct Java® System hook, any APEX® System hook, or the Universal Clamp® (titanium). These direct-connect single point of fixation implants act in concert with the Vitality® System rods and implants to support its surgical indications for use.

In order to achieve additional levels of fixation, the Vitality® Spinal Fixation System rods may be connected to the Instinct Java® System (Ø5.5mm rod) or the Virage® OCT Spinal Fixation System (Ø3.5mm rod) with the corresponding Vitality® rod connectors. Refer to the Instinct Java® System and Virage® OCT Spinal Fixation System package inserts for instructions and indications for use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR § 807.92.

Preparation Date	July 12, 2017
Applicant/Sponsor	Zimmer Biomet Spine, Inc. 10225 Westmoor Dr. Westminster, CO 80021
Contact Person	Kelly Stratton Regulatory Affairs Specialist Phone: 303-501-8386 Fax: 303-501-8444
Trade Name	Vitality® Spinal Fixation System
Common Name	Pedicle Screw Spinal System
Device Class	Class II
Classification Name	Thoracolumbosacral Pedicle Screw System (NKB) Class II per 21 CFR §888.3070 Appliance, Fixation, Spinal Interlaminar (KWP) Class II per 21 CFR §888.3050 Appliance, Fixation, Spinal Intervertebral Body (KWQ) Class II per 21 CFR §888.3060
Device Panel	Orthopedic

Device Description & Technological Characteristics:

The Vitality® Spinal Fixation System is a thoracolumbar and sacroiliac fixation system designed to aid in the surgical correction of several types of spinal conditions. The system consists of a variety of spinal rods, pedicle screws, and connectors intended only to provide temporary stabilization during the development of a solid fusion of the spine with bone graft. The system can be rigidly locked into a variety of configurations, with each construct being customized to the patient's anatomy. All implants are single use only and should not be reused under any circumstances. The implant system is intended to be removed after solid fusion has occurred.

The system also includes instrumentation for insertion and removal and securing of the implants. All implants are made from medical grade titanium alloy; select rods are also available in medical grade cobalt chromium alloy. Implants made from medical grade titanium, medical grade titanium

alloy, and medical grade cobalt chromium may be used together. Never use titanium, titanium alloy, and/or cobalt chromium with stainless steel in the same implant construct.

The Vitality Spinal Fixation System is compatible with the Virage® OCT Spinal Fixation System Rods, Instinct® Java™ Spinal Fixation System Rods and Hooks, and Universal Clamp® Spinal Fixation System.

Intended Use / Indications for Use:

The Vitality® Spinal Fixation System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients and for adolescent patients.

These devices are indicated as an adjunct to fusion for all of the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic sites), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e. scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, pseudoarthrosis and/or failed previous fusion. When used as an adjunct to fusion, the Vitality® Spinal Fixation System is intended to be used with autograft and/or allograft.

In addition the Vitality® Spinal Fixation System is intended for treatment of severe spondylolisthesis (Grade 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and or ilium with removal of the implant after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/ilium.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the Vitality® System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The Vitality® System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Assembly of a spinal fixation construct using the Vitality® Spinal Fixation System may include additionally any Instinct® Java™ System hook, any APEX® System hook, or the Universal Clamp® (titanium). These direct-connect single point of fixation implants act in concert with the Vitality® System rods and implants to support its surgical indications for use.

In order to achieve additional levels of fixation, the Vitality® Spinal Fixation System rods may be connected to the Instinct® Java™ System (Ø5.5mm rod) or the Virage® OCT Spinal Fixation System (Ø3.5mm rod) with the corresponding Vitality® rod connectors. Refer to the Instinct® Java™ System and Virage® OCT Spinal Fixation System package inserts for instructions and indications for use.

Summary of Technologies:

The technological characteristics of the subject Vitality® Spinal Fixation System components remain the same as, or similar to, the predicate Vitality® Spinal Fixation System (K150896) in regards to intended use, indications for use, design, manufacturing methods, fundamental technology, and

operational principles. The purpose of this submission is to seek clearance for minor modifications to the screw design, the addition of hooks, and the addition of associated instruments.

Performance Data:

To support substantial equivalence, mechanical testing of the modified screw implants of the subject Vitality® Spinal Fixation System were assessed and tested appropriately in accordance with ASTM standards. Performance testing included tests per ASTM F1798 (axial grip strength, torsional grip strength, and flexion/extension moment grip strength) and per ASTM F1717 (static compression bending, dynamic compression bending, and static torsion). In all instances, the modified device functioned as intended and demonstrated substantial equivalence to the predicate device(s).

Substantial Equivalence:

The subject Vitality® Spinal Fixation System is substantially equivalent to the Vitality® Spinal Fixation System (K150896).

Conclusion:

The Vitality® Spinal Fixation System is substantially equivalent to the predicate system as a spinal fixation device in regards to intended use, indications for use, fundamental technology including design, materials, manufacturing methods, sterility, and operational principles. Furthermore, mechanical testing and other supporting information sufficiently demonstrate the substantial equivalence of the subject components to the Vitality® Spinal Fixation System, which has been cleared for non-cervical spinal fixation. Based on this information, the subject device does not raise any new issues regarding the safety or efficacy when compared to its predicates.