



February 23, 2018

NeuroStructures, Inc.
% Kenneth Maxwell II
Regulatory and Quality Specialist
Empirical Testing Corp.
4628 Northpark Drive
Colorado Springs, California 80918

Re: K173077

Trade/Device Name: Cavetto®-SA Cervical Cage System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: OVE
Dated: January 19, 2018
Received: January 22, 2018

Dear Mr. Maxwell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good

manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration
Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number (if known)

K173077

Device Name

Cavetto®-SA Cervical Cage System

Indications for Use (Describe)

The Cavetto®-SA Cervical Cage System is intended for spinal fusion procedures at one level (C2 to T1) in skeletally mature patients with degenerative disc disease (defined as pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies) of the cervical spine. Implants are intended to be implanted via an open, anterior approach and packed with autograft bone and/or allogenic bone graft composed of cancellous and/or corticocancellous bone graft. This cervical device is to be used in patients who have had six weeks of non-operative treatment. The Cavetto®-SA Cervical Cage System should be used with the provided bone screws and requires no additional supplementary fixation systems.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(K) SUMMARY

Submitter's Name:	Neurostructures, Inc.
Submitter's Address:	16 Technology Drive, Suite 165 Irvine, CA 92618
Submitter's Telephone:	Kathleen Wong kw@neurostructures.com 949.370.4497
Contact Person:	Kenneth C. Maxwell II Empirical Testing Corp. 719.291.6874 kmaxwell@empiricalconsulting.com
Date Summary was Prepared:	20 February 2018
Trade or Proprietary Name:	Cavetto®-SA Cervical Cage System
Common or Usual Name:	Intervertebral Fusion Device With Integrated Fixation, Cervical
Classification:	Class II per 21 CFR §888.3080
Product Code:	OVE
Classification Panel:	Division of Orthopedic Devices

DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

The Cavetto®-SA Cervical Cage System is an intervertebral fusion device made from medical grade PEEK per ASTM F2026, titanium per ASTM F560, and tantalum markers per ASTM F560. The subject device is offered in a variety of footprints, styles, and sizes to accommodate various patient anatomies.

INDICATIONS FOR USE

The Cavetto®-SA Cervical Cage System is intended for spinal fusion procedures at one level (C2 to T1) in skeletally mature patients with degenerative disc disease (defined as pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies) of the cervical spine. Implants are intended to be implanted via an open, anterior approach and packed with autograft bone and/or allogenic bone graft composed of cancellous and/or corticocancellous bone graft. This cervical device is to be used in patients who have had six weeks of non-operative treatment. The Cavetto®-SA Cervical Cage System should be used with the provided bone screws and requires no additional supplementary fixation systems.

TECHNOLOGICAL CHARACTERISTICS

The Cavetto®-SA Cervical Cage System and predicate devices have nearly identical technological characteristics and the minor differences do not raise any new issues of safety and

effectiveness. Specifically the following characteristics are identical between the subject and predicates:

- Indications for Use
- Principle of operation
- Implant material PEEK
- Implant material titanium
- Screw implant diameter and length
- Structural support mechanism
- Cage implant sizes

Table 5-1: Predicate Devices

510k Number	Trade or Proprietary or Model Name	Manufacturer	Predicate Type
K132894	Optio-C™	Zimmer Spine	Primary
K153250	Tesera	Renovis	Additional
K150053	STALIF C®	Centinel Spine	Additional
K152515	TOMCAT™	Choice Spine	Additional
K102606	AVS Anchor-C	Stryker Spine	Additional

PERFORMANCE DATA

The Cavetto®-SA Cervical Cage System has been tested in the following test modes:

- Static axial compression per ASTM F2077
- Static compressive shear per ASTM F2077
- Static torsion per ASTM F2077
- Dynamic axial compression per ASTM F2077
- Dynamic compressive shear per ASTM F2077
- Dynamic torsion per ASTM F2077
- Static subsidence per ASTM F2267
- Expulsion per ASTM F-04.25.02.02

The results of this non-clinical testing show that the strength of the Cavetto®-SA Cervical Cage System is sufficient for its intended use and is substantially equivalent to legally marketed predicate devices.

CONCLUSION

The overall technology characteristics and mechanical performance data lead to the conclusion that the Cavetto®-SA Cervical Cage System is substantially equivalent to the predicate device.