November 17, 2017

Monteris Medical
David Mueller
Senior Regulatory Affairs Specialist
14755 27th Avenue North
Suite C
Plymouth, Minnesota 55447

Re: K173305
Trade/Device Name: NeuroBlate System
Regulation Number: 21 CFR 878.4810
Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In Dermatology
Regulatory Class: Class II
Product Code: GEX, HAW
Dated: October 17, 2017
Received: October 18, 2017

Dear Mr. Mueller:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/) and CDRH Learn (http://www.fda.gov/Training/CDRHLearn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (http://www.fda.gov/DICE) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Michael J. Hoffmann -S

for Carlos L. Peña, PhD, MS
Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number *(if known)*
K173305

Device Name
Monteris Medical NeuroBlate™ System

Indications for Use *(Describe)*
The Monteris Medical NeuroBlate™ System is indicated for use to ablate, necrotize, or coagulate intracranial soft tissue, including brain structures, through interstitial irradiation or thermal therapy in medicine and surgery in the discipline of neurosurgery with 1064 nm lasers.

The Monteris Medical NeuroBlate™ System is intended for planning and monitoring thermal therapies under MRI visualization. It provides MRI based trajectory planning assistance for the stereotaxic placement of MRI compatible (conditional) NeuroBlate™ Laser Delivery Probes. It also provides real time thermographic analysis of selected MRI images.

When interpreted by a trained physician, this System provides information that may be useful in the determination or assessment of thermal therapy. Patient management decisions should not be made solely on the basis of the NeuroBlate™ System analysis.

Type of Use *(Select one or both, as applicable)*

☒ Prescription Use (Part 21 CFR 801 Subpart D) ☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

MONTERIS MEDICAL

Section 1: 510(k) Summary

5a. Device Information:

Category	Comments
Sponsor:	Monteris Medical Corp. 14755 27th Avenue North Suite C Plymouth, MN 55446 763-253-4710 Fax: 763-746-0084 www.monteris.com
Correspondent Contact Information:	David H. Mueller Senior Principal Regulatory Affairs Specialist Monteris Medical, TEL: 763-253-4710 x2732 FAX: 763-746-0084 Email: DMueller@Monteris.com
Device Common Name:	Magnetic Resonance Image Guided Laser Thermal Therapy System
Device Classification Number:	21 CFR 878.4810 Laser surgical instrument for use in general and plastic surgery and in dermatology 21 CFR 882.4560 Stereotaxic instrument
Device Classification & Product Code:	Class II, GEX Class II, HAW
Device Proprietary Name:	Monteris Medical NeuroBlate™ System

Predicate Device Information:

Predicate Device:	NeuroBlate™ System
Predicate Device Manufacturer:	Monteris Medical
Predicate Device Common Name:	Monteris NeuroBlate™ System
Predicate Device Premarket Notification #	K172881
Predicate Device Regulation:	21 CFR 878.4810 Laser surgical instrument for use in general and plastic surgery and in dermatology 21 CFR 882.4560 Stereotaxic instrument
Predicate Device Classification & Product Code:	Class II, GEX Class II, HAW

5b. Date Summary Prepared

30 October 2017

5c. Description of Device

The Monteris NeuroBlate™ System is a collection of MRI-compatible laser devices and accessories that create an MRI guided intracranial delivery of precision thermal therapy in the practice of neurosurgery.

As previously described in K172881, the NeuroBlate™ System is typically used for the minimally invasive ablation of target tissue (tumors, epileptic foci) in the brain.

The NeuroBlate™ System components consist of:

- Families of gas-cooled Laser Delivery Probe (Probe) (SideFire & FullFire) to deliver controlled energy to a target zone.
- Probe Drivers (Advanced Probe Driver, Robotic Probe Driver) which allow the surgeon to precisely position, stabilize and manipulate a probe, endoscope or other device within the target zone.
- An Interface Platform, which attaches to the MRI system patient table and provides supporting electronics for the Probe Drivers and interconnections for the Laser Delivery Probes (e.g., Connector Module);
- A System Electronics Rack and Components, which includes the laser and necessary umbilicals, cables, penetration panels, and small hardware for system mechanical, electrical, and electronic operation,
- A Control Workstation including the *M-Vision*™ and *M-Vision Pro*™ software, which includes a user interface for procedure planning, interactive monitoring of NeuroBlate™ procedures, and interfaces to the MRI and hardware subsystems.

The NeuroBlate™ System is utilized with stereotaxic frames and patient stabilization systems, such as:

- The Axiiis stereotaxic mini-frame and the Monteris Cranial Bolt and Mini-Bolt fixation components, and
- The AtamA Stabilization System and MRI receive-only head coil, as well as, other optional accessories, including: drill bits, bolts, thumbscrews, instrument adaptors, accessory host adaptors, MRI trajectory wands, cranial screws, fiducial markers, bone screws, stereotactic manual driver with mandrel and T-handle, and other manual accessory instruments and tools).

Thus, there is no change to entire system, with the exception of an additional (alternative) NeuroBlate M-Vision/ M-Vision Pro Software Package (V3.13). While the modified software package includes several modifications which do not meet the FDA's "significant change" submission criteria, there is at least one proposed change which modifies an existing risk control measure for a hazardous situation and thus, per FDA 510(k) guidance, requires a new 510(k) submission.

MONTERIS MEDICAL

5d. Indications for Use

There is no change to the indications for use, i.e., they remain:

> *The Monteris Medical NeuroBlate™ System is indicated for use to ablate, necrotize, or coagulate intracranial soft tissue, including brain structures, through interstitial irradiation or thermal therapy in medicine and surgery in the discipline of neurosurgery with 1064 nm lasers.*
>
> *The Monteris Medical NeuroBlate™ System is intended for planning and monitoring thermal therapies under MRI visualization. It provides MRI based trajectory planning assistance for the stereotaxic placement of MRI compatible (conditional) NeuroBlate™ Laser Delivery Probes. It also provides real-time thermographic analysis of selected MRI images.*
>
> *When interpreted by a trained physician, this System provides information that may be useful in the determination or assessment of thermal therapy. Patient management decisions should not be made solely on the basis of the NeuroBlate™ System analysis.*

5e. Comparison to Predicate Device

There is no change to entire system, with the exception of an additional (alternative) NeuroBlate M-Vision/ M-Vision Pro Software Package (V3.13). While the modified software package includes several modifications which do not meet the FDA's "significant change" submission criteria, there is at least one proposed change which modifies an existing risk control measure for a hazardous situation and thus, per FDA 510(k) guidance, requires a new 510(k) submission.

The application for the Monteris Medical NeuroBlate™ System with the M-Vision/ M-Vision Pro is substantially equivalent to the predicate Monteris NeuroBlate™ System (K172881) in intended use, technology, design and physician use.

As the modifications presented in the current device do not change the intended use, operating principles, or raise any unaddressed safety concerns, it can be concluded the application NeuroBlate™ System with the alternative M-Vision/ M-Vision Pro (V3.13) Software is substantially equivalent to the predicate NeuroBlate™ System.

5f. Summary of Supporting Data

The updated M-Vision/ M-Vision Pro (V3.13) Software development process followed Monteris' documented Quality System and incorporated a design verification and design validation process. This process included an overarching Design Verification and Design Validation Master Plan. This plan describes the design verification and the design validation of the user needs for the M-Vision Pro (V3.13) Software when used within the NeuroBlate System.

The Design Verification process utilized protocols to detail the associated tests. The corresponding test reports which confirmed (and documented) the design output met the design

input for the requirements. Each verification test protocol incorporated clearly defined acceptance criteria.

The Design Validation process utilized protocols to detail the associated tests. The corresponding test reports which confirmed (and documented) the modified NeuroBlate™ System met the user needs and intended use. Each validation protocol described the objective, test method and acceptance criteria.

Thus, the application for the Monteris Medical NeuroBlate™ System with the M-Vision Pro (V3.13) is substantially equivalent to the predicate Monteris NeuroBlate™ System (K172881) in intended use, technology, design and physician use.