

FEB 14 2002

## Special 510(k): Device Modification for the Osteonics® Spinal System Summary of Safety and Effectiveness

### Submission Information

Name and Address of the Sponsor of the 510(k) Submission:	Howmedica Osteonics Corp 59 Route 17 Allendale, NJ 07401-1677
Contact Person:	Karen Ariemma Regulatory Affairs Specialist
Date of Summary Preparation:	January 25, 2002

### Device Identification

Proprietary Name:	Osteonics® Spinal System
Common Name:	Spinal Fixation Appliances
Classification Name and Reference:	Spinal Interlaminar Fixation Orthosis 21 CFR 888.3050 Pedicle Screw Spinal System 21 CFR 888.3070

### Predicate Device Identification

The features of the subject device are substantially equivalent to features of the Osteonics® Spinal System (OSS) Combination Screw Ring/Blocker which was determined substantially equivalent via 510(k) K990158 and to the OSS Screw Blocker and Cap which were determined substantially equivalent via 510(k) K951725. The modified device shall be referred to here in as the Diapason Combo. The Xia 6 mm diameter rods were determined substantially equivalent via K984251.

### Description of Device Modification

The design change involves modifying the ring height, blocker height and assembly process of the predicate OSS Combination Screw Ring/Blocker. In addition, the submission covers use of predicate Xia 6 mm diameter Rods with the Osteonics Spinal System. The submission involves no change to the Xia rods themselves.

### Intended Use

The subject Diapason Combo, like the predicate OSS Combination Screw Ring/Blocker, are intended for use only with the components of the commercially available Osteonics® Spinal System/ Diapason System. The Rods from the Xia Spinal System are intended to be used with the components of the Osteonics® Spinal System. The uses for the legally marketed Osteonics® Spinal System/Diapason System are as follows:

**As a posterior, non-pedicle screw system of the T4-S2 spine, the Osteonics® Spinal System is indicated for:**

- Long and short curve scoliosis
- Vertebral fracture or dislocation
- Spondylolisthesis
- Degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies).
- Previously failed fusion
- Spinal tumor

**Pedicular Use:**

- When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the Osteonics® Spinal System is indicated for one or more of the following: degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor, and failed previous fusion (pseudarthrosis).
- In addition, the Osteonics® Spinal System is indicated for pedicle screw fixation in skeletally mature patients with severe spondylolisthesis (Grades 3 and 4) at the L5-S1 joint, having fusions with autogenous bone graft, having the device fixed or attached to the lumbar and sacral spine (with pedicle placement at L3 and below) with removal of the implants after the development of a solid fusion mass.

**Statement of Technological Comparison:**

Static and fatigue testing demonstrates the comparable mechanical properties of the subject Diapason Combo to the predicate OSS Screw Blocker and Cap.

Static and fatigue testing demonstrates the acceptable mechanical properties of the subject Xia rods with the predicate Osteonics® Spinal System/Diapason System.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

**FEB 14 2002**

Ms. Karen Ariemma  
Regulatory Affairs Specialist  
Howmedica Osteonics Corp.  
59 Route 17  
Allendale, New Jersey 07401

Re: K020293

Trade Name: Osteonics® Spinal System  
Regulatory Number: 21 CFR 888.3070 and 888.3050  
Regulation Name: Pedicle Screw Spinal System, Spondylolisthesis Spinal Fixation  
Device System, Spinal Interlaminar Fixation Orthosis,  
Regulatory Class: II  
Product Code: MNH, MNI, KWP  
Dated: January 25, 2002  
Received: January 28, 2002

Dear Ms. Ariemma:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Ms. Karen Ariemma

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or at (301) 443-6597, or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

  
for Celia M. Witten, Ph.D., M.D.  
Director  
Division of General Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K020293

Device Name: Modification to the Osteonics® Spinal System

The subject component, a design modification to the Osteonics® Spinal System Combination Screw Ring/Blocker, is a single-use device which is sold non-sterile and is intended for use only with the other components of the commercially available Osteonics® Spinal System. The Rods from the Xia Spinal System are intended to be used with the components of the Osteonics® Spinal System.

The uses for the legally marketed Osteonics® Spinal System are as follows:

**As a posterior, non-pedicle screw system of the T4-S2 spine, the Osteonics® Spinal System is indicated for:**

- Long and short curve scoliosis
- Vertebral fracture or dislocation
- Spondylolisthesis
- Degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies).
- Previously failed fusion
- Spinal tumor

**Pedicular Use:**

- When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the Osteonics® Spinal System is indicated for one or more of the following: degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor, and failed previous fusion (pseudarthrosis).
- In addition, the Osteonics® Spinal system is indicated for pedicle screw fixation in skeletally mature patients with severe spondylolisthesis (Grades 3 and 4) at the L5-S1 joint, having fusions with autogenous bone graft, having the device fixed or attached to the lumbar and sacral spine (with pedicle placement at L3 and below) with removal of the implants after the development of a solid fusion mass.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓

(Per 21 CFR 801.109)

Miriam C. Provost  
(Division Sign-Off)

Division of General, Restorative  
and Neurological Devices (Optional Format 1-2-96)

Over-The-Counter Use \_\_\_\_\_

510(k) Number K020293