



August 7, 2020

Streck, Inc.
Deborah Kipp
Regulatory Affairs Manager
7002 S. 109th Street
La Vista, NE 68128

Re: DEN200001

Trade/Device Name: Cell-Free DNA BCT
Regulation Number: 21 CFR 862.1676
Regulation Name: Blood collection device for cell-free nucleic acids
Regulatory Class: Class II
Product Code: QMA
Dated: January 9, 2020
Received: January 10, 2020

Dear Deborah Kipp:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your De Novo request for classification of the Cell-Free DNA BCT, a prescription device with the following indications for use:

Cell-Free DNA BCT is a direct-draw venous whole blood collection device intended for the collection, stabilization, and transport of venous whole blood samples for use in conjunction with cell-free DNA next-generation sequencing liquid biopsy assays that have been cleared or approved for use with samples collected in the Cell-Free DNA BCT device.

Although this letter refers to your product as a device, please be aware that some granted products may instead be combination products. If you have questions on whether your product is a combination product, contact CDRHProductJurisdiction@fda.hhs.gov. FDA concludes that this device should be classified into Class II. This order, therefore, classifies the Cell-Free DNA BCT, and substantially equivalent devices of this generic type, into Class II under the generic name Blood collection device for cell-free nucleic acids.

FDA identifies this generic type of device as:

Blood collection device for cell-free nucleic acids. A blood collection device for cell-free nucleic acids is a device intended for medical purposes to collect, store, transport, and handle blood specimens and to stabilize and isolate cell-free nucleic acid components prior to further testing.

Section 513(f)(2) of the Food, Drug and Cosmetic Act (the FD&C Act) was amended by section 607 of the Food and Drug Administration Safety and Innovation Act (FDASIA) on July 9, 2012. This law provides two options for De Novo classification. First, any person who receives a "not substantially equivalent" (NSE) determination in response to a 510(k) for a device that has not been previously classified under the Act may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act. On December 13, 2016, the 21st Century Cures Act removed a requirement that a De Novo request be submitted within 30 days of receiving an NSE determination. Alternatively, any person who determines that there is no legally marketed device upon which to base a determination of substantial equivalence may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act without first submitting a 510(k). FDA shall, within 120 days of receiving such a request, classify the device. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a notice in the Federal Register announcing the classification.

On January 10, 2020, FDA received your De Novo requesting classification of the Cell-Free DNA BCT. The request was submitted under section 513(f)(2) of the FD&C Act. In order to classify the Cell-Free DNA BCT into class I or II, it is necessary that the proposed class have sufficient regulatory controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use. After review of the information submitted in the De Novo request, FDA has determined that, for the previously stated indications for use, the Cell-Free DNA BCT can be classified in class II with the establishment of special controls for class II. FDA believes that class II (special) controls provide reasonable assurance of the safety and effectiveness of the device type. The identified risks and mitigation measures associated with the device type are summarized in the following table:

Identified Risk	Mitigation Measures
Blood pathogen exposure/Injury	Certain design verification and validation
Failure to collect and transport sample	Certain design verification and validation
Insufficient sample quantity and quality	Certain design verification and validation

In combination with the general controls of the FD&C Act, the blood collection device for cell-free nucleic acids is subject to the following special controls:

1. Design verification and validation documentation must include appropriate design inputs and design outputs that are essential for the proper functioning of the device for its intended use, including all of its indications for use, and must include the following:
 - i. Documentation demonstrating that appropriate, as determined by Food and Drug Administration, measures are in place (e.g., validated device design features and specifications) to ensure that users of blood collection device for cell-free nucleic acids devices are not exposed to undue risk of blood-borne pathogen exposure and operator injury during use of the device, including blood collection, transportation, and centrifugation processes.

- ii. Documentation demonstrating that appropriate, as determined by Food and Drug Administration, measures are in place (e.g., validated device design features and specifications) to ensure that the device reproducibly and reliably collects, transports, stabilizes, and isolates cell-free nucleic acids of sufficient yield and quality suitable for downstream applications as appropriate for its intended use. At a minimum, these measures must include:
 - A. Data demonstrating that blood samples collected in the device have reproducible cell-free nucleic acid yields that are suitable, as determined by Food and Drug Administration, for downstream testing as appropriate for the intended use, including estimates of within-lot, within-device, and lot-to-lot variability.
 - B. Data demonstrating that cell-free nucleic acid yields isolated from blood specimens collected into the device do not add clinically significant bias to test results obtained using the downstream application(s) described in the intended use. For devices indicated for use with multiple downstream applications, data demonstrating acceptable performance for each type of claimed use or, alternatively, an appropriate, as determined by Food and Drug Administration, clinical justification for why such data are not needed.
 - C. Data demonstrating that the device appropriately stabilizes cell-free nucleic acids after sample collection, during storage, and during transport over the claimed shelf life of the device.
 - D. Data demonstrating that samples collected in the device have minimal levels of contamination with other types of nucleic acids present in cells or cellular components, and that these levels of contamination do not interfere with downstream testing.
 - E. Data from analytical or clinical studies that demonstrate that, when used as intended, the device consistently draws a blood sample volume that is within the indicated fill range.
 - F. Data from analytical or clinical studies that demonstrate that, when used as intended, cell-free nucleic acid yield, stability, and quality are not significantly impacted by interference due to other parts of the device (such as reduced or excess active ingredient) or specimen collection and processing procedures (such as hemolysis, centrifugation, or mixing of blood with anticoagulant or additives).
 - G. Data from analytical studies that demonstrate that the device is suitable for its intended use across all storage and sample handling conditions described in the device labeling, including device shelf life and shipping conditions (e.g., temperature, humidity, duration).
- iii. A protocol, reviewed and determined acceptable by the Food and Drug Administration, that specifies the verification and validation activities that will be performed for anticipated device modifications to reevaluate performance claims or performance specifications. This protocol must include a process for assessing whether a modification to technology, engineering, performance, materials, specifications, or indications for use, or any combination thereof, could significantly affect the safety or effectiveness of the device. The protocol must include assessment metrics, acceptance criteria, and analytical methods for the performance testing of changes.

Section 510(m) of the FD&C Act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the FD&C Act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device type. FDA has determined premarket notification is necessary to provide reasonable assurance of the safety and effectiveness of the device type and, therefore, the device is not exempt from the premarket notification requirements of the FD&C Act. Thus, persons who intend to market this device type must submit a premarket notification containing information on the blood collection device for cell-free nucleic acids they intend to market prior to marketing the device.

Please be advised that FDA's decision to grant this De Novo request does not mean that FDA has made a determination that your device complies with other requirements of the FD&C Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the FD & C Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and if applicable, the electronic product radiation control provisions (Sections 531-542 of the FD & C Act); 21 CFR 1000-1050.

A notice announcing this classification order will be published in the Federal Register. A copy of this order and supporting documentation are on file in the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Room 1061, Rockville, MD 20852 and are available for inspection between 9 a.m. and 4 p.m., Monday through Friday.

As a result of this order, you may immediately market your device as described in the De Novo request, subject to the general control provisions of the FD&C Act and the special controls identified in this order.

For comprehensive regulatory information about medical devices and radiation-emitting products, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

If you have any questions concerning the contents of the letter, please contact Lindsey Lund at 240-402-5267.

Sincerely,

Kellie B. Kelm, Ph.D.
Director
Division of Chemistry and Toxicology Devices
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health