



DEC 21 2012

Biocoat Inc.
c/o James Johnston, Ph.D.
Principal Investigator
455 Pennsylvania Avenue, Suite 120
Ft. Washington, Pennsylvania 19034

Re: k032874

Trade/Device Name: HBA Sperm-Hyaluronan Binding Assay
Regulation Number: 21 CFR § 864.5220
Regulation Name: Automated differential cell counter
Regulatory Class: Class II
Product Code: GKZ
Date: September 9, 2003
Received: September 17, 2003

Dear Dr. Johnston:

This letter corrects our substantially equivalent letter of November 24, 2003.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of


Page 2 – James Johnston, Ph.D.

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Maria M. Chan, Ph.D.
Director
Division of Immunology and Hematology Devices
Office of *In Vitro* Diagnostics and Radiological Health
Center for Devices and Radiological Health

Enclosure

SECTION 10


Indications for use statement

510(k) Number : K 03 2874

Device Name: HBA Sperm-Hyaluronan Binding Assay

Indications For Use:

1. As a component of the standard analysis of semen in the diagnosis of suspected male infertility,
2. As a component of analyses for determining the proper course of IVF treatment of infertility.



Division Sign-Off

**Office of In Vitro Diagnostic Device
Evaluation and Safety**

510(k) K03 2874