

JUL 22 2005

Non-Confidential Summary of Safety and Effectiveness

Page 1 of 2
April 29, 2005

Axon Medical Inc.
2355 South 1070 West, Suite D
Salt Lake City, Utah 84119

Tel – (801) 484-3820
Fax – (801) 977-0363

Official Contact: Joseph Orr – President

Proprietary or Trade Name: AneFin 100

Common/Usual Name: Rebreathing / Absorber

Primary Classification Name: Gas Scavenging Apparatus

Primary Classification Code: CBN

Secondary Classification Name: Rebreathing Device

Secondary Classification Code: BYW

Primary Predicate Devices: RFS Vacuum gauge scavenging circuit,
Accutron - K033503
“Protect-OR” filter, Charcoal based scavenging device,
Foregger - Pre-Amendment

Secondary Predicate: Model A100 CO₂ absorber with bypass valve
Penlon, 510(k) exempt
Non invasive cardiac output monitor, NICO
(Product code: CCK)
Novamatrix - K030886

Device Description: The AneFin combines an anesthetic gas absorber to remove anesthetic gas from inhaled air and a rebreathing hose which increases inspired CO₂ amounts which allows increased patient ventilation while preventing hypoxia during emergence from volatile inhaled anesthesia.

Intended Use:

The AneFin 100 is intended to speed emergence from volatile inhaled anesthetics by removing unwanted anesthetic gases and increasing spontaneous breathing through partial rebreathing.

Environment of Use: Operating room, surgical suite, anywhere inhaled volatile anesthetics are administered.

Non-Confidential Summary of Safety and Effectiveness

Page 2 of 2
 April 29, 2005

Comparison to Predicate Devices to demonstrate substantial equivalence

Attribute	Proposed Device	RFS Vacuum scavenging K033503	Protect-OR scavenging device Pre-amendment	Model A100 CO ₂ absorber bypass valve, exempt	Non-invasive cardiac output monitor, NICO K030886
Intended Use	Speed emergence from inhaled volatile anesthetics	Remove anesthetic agent from operating room	Remove anesthetic agent from operating room	Remove CO ₂ and allow rapid build-up of CO ₂ by rebreathing	Measure cardiac output by partial CO ₂ rebreathing
Scavenging method	Charcoal adsorption	Vacuum system conveys waste gas out of operating room	Charcoal adsorption	Not applicable	Not applicable
Method of increasing CO ₂	Dead space tubing placed between patient and Y-Piece.	Not applicable	Not applicable	Valve on absorber allows expired gas containing CO ₂ to bypass absorber	Dead space tubing placed between patient and Y-Piece
Rebreathing Volume	431 ml max.	Not applicable	Not applicable	Depends on the volume of the breathing circuit	> 400 ml
Intended Population	Surgical Patients receiving inhaled anesthetics	Dental patients receiving inhaled anesthetics	Surgical Patients receiving inhaled anesthetics Same	Surgical Patients receiving inhaled anesthetics	Surgical Patients receiving inhaled anesthetics
Environment of Use	Operating Room	Operating Room, Dental Suite	Operating Room	Operating Room	Operating Room, ICU
Placement in circuit	Between endotracheal tube and Y-Piece	Gas evacuation port	Gas evacuation port	Between inspired and expired one-way valves	Between endotracheal tube and Y-piece
Materials	Activated charcoal, polypropylene housing and rebreathing hose	Not applicable	Activated Charcoal	Not applicable	Polypropylene housing and rebreathing hose

Differences between Other Legally Marketed Predicate Devices

There are no significant differences between the intended device and the identified predicates.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

January 24, 2013

Mr. Joseph Orr
President
Axon Medical, Incorporated
2355 South 1070 West, Suite D
SALT LAKE CITY UT 84119

Re: K033028
Trade/Device Name: Anefin 100
Regulation Number: 21 CFR 868.5430
Regulation Name: Gas-Scavenging Apparatus
Regulatory Class: II
Product Code: CBN, BYW
Dated: April 29, 2005
Received: May 2, 2005

Dear Mr. Orr:

This letter corrects our substantially equivalent letter of July 22, 2005.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number: K033028 (To be assigned)

Device Name: AneFin 100

Indications for Use:

The AneFin 100 is intended to speed emergence from the effects of volatile inhaled anesthetics by removing unwanted anesthetic gases and increasing spontaneous breathing through partial rebreathing.

It is intended for use with only Isoflurane, Sevoflurane and Desflurane.

Prescription Use XX or Over-the-counter use _____
(Per CFR 801.109)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: K033028