

K042042

510(k): Ultrasound Capture System (UCS)

SEP 21 2004

This is a summary of 510(k) safety and effectiveness information is being submitted in accordance with the SMDA 1990 and 21 CFR 807.92.

**DATE:**

27 July 2004

**SUBMITTER:**

Heartlab Inc.  
One Crosswind Road  
Westerly, RI 02891  
Phone: (401) 596-0592  
Fax: (401) 596-8562

**CONTACT PERSON:**

Richard Petrocelli  
Tel No: (401) 596-0592

**IDENTIFICATION OF THE PRODUCT**

**TRADE NAME:** Ultrasound Capture System™ (UCS)  
**COMMON NAME:** Ultrasound Capture System / EcoCapture  
**CLASSIFICATION NAME:** Image Processing System, LLZ

**DEVICE DESCRIPTION / INTENDED USE:**

The Ultrasound Capture System™ (UCS) system captures analog video and audio from any ultrasound with a composite or S-video output and converts the captured video to DICOM image and audio files. UCS software runs on a standard "off-the-shelf" personal computer running the Windows XP operating system and a high-resolution video acquisition interface.

1. Allows demographic (Patient Name, ID etc.) input consistent with the DICOM standard for image storage.
2. Performs automatic image acquisition triggered by the cine-pedal.
3. Performs network transmission of images in DICOM format to a DICOM compatible storage system.

**SUBSTANTIAL EQUIVALENCE INFORMATION:**

Ultrasound Capture System (UCS) is considered comparable and substantially equivalent to the following predicate devices currently in commercial distribution:

| <b>Model</b>                      | <b>Manufacturer</b> |
|-----------------------------------|---------------------|
| CineCapture System (k990857)      | Heartlab, Inc.      |
| Echocardiography System (k992259) | Camtronics, Ltd.    |

Predicate device specifications comparison:

|                  | Principal Device –<br>Ultrasound Capture System   | Predicate Device – <u>Heartlab</u><br>CineCapture System | Predicate Device – <u>Camtronics</u><br>Echocardiography System     |
|------------------|---|--|---|
| Image Format     | Lossy JPEG, DICOM 3.0   | JPEG, DICOM 3.0  | Lossy JPEG, DICOM 3.0   |
| Compression      | Variable, 10-12:1<br>recommended compression  | Lossless, 512x512, 2:1, 8 bit                            | Up to 30:1  |
| Video Source     | S-video, or composite   | High-line signal   | RGB, YC, or composite   |
| Display          | 1024x600 on small laptop,<br>1024x768 on standard<br>laptop and desktop<br>configurations | 15” flat panel with speaker,<br>up to 1024x768           | 17”, 19”, or 21” color<br>monitors, up to 1024x1024 24<br>bit depth |
| Operating System | Windows XP  | Windows 2000   | Windows 2000  |
| User Interface   | Keyboard, mouse, and foot<br>pedal  | Keyboard, mouse, and foot<br>pedal                       | User multifunction remote<br>control, and mini-remote               |
| Network          | Fast Ethernet   | Fast Ethernet  | Fast Ethernet   |
| DICOM            | Yes   | Yes  | Yes   |

## **STANDARDS:**

Ultrasound Capture System (UCS) is designed in accordance with product safety and performance requirements set forth in the following standards:

1. Digital Imaging and Communications in Medicine (DICOM)
2. 21 CFR 1020.10 Video Monitor Performance Requirements
3. 21 CFR 1040.10 Fiberoptic communications Performance
4. Society of Motion Picture and Television Engineers (SMPTE)
5. ACR/NEMA Data Compression Standard
6. Underwriters Laboratories (U.L.) Standard No. 544 for Medical and Dental Equipment
7. ISO/IEC 10918-1 Digital Compression and Coding Continuous-Tone Still Images (JPEG)
8. Underwriters Laboratories (U.L.) Standard No. 1950 Safety Standards

## **SUMMARY OF DESIGN CONTROL ACTIVITIES:**

The software utilized was designed, developed, tested and validated according to written Design Control procedures. These procedures specify individuals within the organization responsible for developing and approving product specifications, coding, testing, validating and maintenance. Potential hazards have been studied and controlled by a Risk Management Plan.

The following quality assurance design control measures were applied to the development of the Ultrasound Capture System™ product:

1. Risk Analysis
2. Requirement Reviews
3. Design Reviews
4. Testing on unit level (Module verification)
5. Integration testing (System verification)
6. Final acceptance testing (Validation)
7. Performance testing



**CONCLUSIONS:**

The principles of operation of the Ultrasound Capture System are substantially equivalent to the currently marketed products. This system poses no added risk to safety.

This concludes this 510(k) Summary.

Richard Petrocelli,  
President  
Heartlab, Inc

Attachments: Appendix A – Indications For Use  
Appendix B – Ultrasound Capture System (UCS) Datasheet  
Appendix C – Predicate Device Promotional Material – Heartlab –  
CineCapture System  
Appendix D – Predicate Device Promotional Material – Camtronics – Echocardiography  
System



SEP 21 2004

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr Richard Petrocelli  
Official Correspondent  
Heartlab, Inc.  
One Crosswind Road  
WESTERLY RI 02891

Re: K042042  
Trade/Device Name: Ultrasound Capture System (UCS)  
Regulation Number: 21 CFR 892.2030  
Regulation Name: Medical image digitizer  
Regulatory Class: II  
Product Code: 90 LMA  
Dated: July 27, 2004  
Received: August 3, 2004

Dear Mr. Petrocelli:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.


This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of the letter:

|                                  |                |
|----------------------------------|----------------|
| 8xx.1xxx                         | (301) 594-4591 |
| 876.2xxx, 3xxx, 4xxx, 5xxx       | (301) 594-4616 |
| 884.2xxx, 3xxx, 4xxx, 5xxx, 6xxx | (301) 594-4616 |
| 892.2xxx, 3xxx, 4xxx, 5xxx       | (301) 594-4654 |
| Other                            | (301) 594-4692 |

Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97) you may obtain. Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Nancy C. Brogdon  
Director, Division of Reproductive,  
Abdominal and Radiological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

K042042

~~1003502~~

Abbreviated 510(k) Encompass



**Attachment**

**Device Name:** Ultrasound Capture System™ (UCS)

**Indications For Use:**

The Ultrasound Capture System™ (UCS) system captures analog video and audio from any ultrasound with a composite or S-video output and converts the captured video to DICOM image and audio files. UCS software runs on a standard "off-the-shelf" personal computer running the Windows XP operating system and a high-resolution video acquisition interface.

*Prescription Use*

*David G. Seymour*  
\_\_\_\_\_  
(Division Sign-Off)  
Division of Reproductive, Abdominal,  
and Radiological Devices  
510(k) Number K042042