



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room - WO66-G609  
Silver Spring, MD 20993-0002

Ms. Michelle Grossman  
Principal Regulatory Affairs Associate  
Abbott Vascular  
3200 Lakeside Drive  
Santa Clara, CA 95054

MAY 6 2011

Re: P040012/S34  
RX Acculink Carotid Stent System  
Filed: October 1, 2010  
Amended: November 26, 2010, December 22, 2010, and April 7, 2011  
Procode: NIM

Dear Ms. Grossman:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) supplement for the RX Acculink Carotid Stent System. This device is indicated for high surgical risk and standard surgical risk patients as follows:

High Surgical Risk

The RX Acculink Carotid Stent System, used in conjunction with Abbott Vascular's Accunet or Emboshield family of Embolic Protection Systems (EPS), is indicated for the treatment of patients at high risk for adverse events from carotid endarterectomy who require carotid revascularization and meet the criteria outlined below:

1. Patients with neurological symptoms and  $\geq 50\%$  stenosis of the common or internal carotid artery by ultrasound or angiogram OR patients without neurological symptoms and  $\geq 80\%$  stenosis of the common or internal carotid artery by ultrasound or angiogram, AND
2. Patients must have a reference vessel diameter within the range of 4.0 mm and 9.0 mm at the target lesion.

Standard Surgical Risk

The RX Acculink Carotid Stent System, used in conjunction with the Accunet Embolic Protection System (EPS), is indicated for the treatment of patients at standard risk for adverse events from carotid endarterectomy who require carotid revascularization and meet the criteria outlined below:

1. Patients with neurological symptoms and  $\geq 70\%$  stenosis of the common or internal carotid artery by ultrasound or  $\geq 50\%$  stenosis of the common or internal carotid artery by

angiogram OR patients without neurological symptoms and  $\geq 70\%$  stenosis of the common or internal carotid artery by ultrasound or  $\geq 60\%$  stenosis of the common or internal carotid artery by angiogram, AND

2. Patients must have a reference vessel diameter within the range of 4.0 mm and 9.0 mm at the target lesion.

We are pleased to inform you that the PMA supplement is approved. You may begin commercial distribution of the device as modified in accordance with the conditions of approval described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). The device is further restricted under section 515(d)(1)(B)(ii) of the act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 2 years. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" (please use this title even if the specified interval is more frequent than one year) and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

In addition to the Annual Report requirements, you must provide the following data in a separate post-approval study (PAS) report:

1. You have agreed to conduct a non-randomized, multi-center study of the RX Acculink Carotid Stent System used in conjunction with Abbott Vascular's Accunet Embolic Protection System when used by a broad group of physicians in the population at standard

risk for adverse events from carotid endarterectomy. The study will be conducted per the final protocol dated April 6, 2011 submitted to FDA under P040012/S034/A003. You have agreed to conduct the CANOPY trial, a post approval study that will include a minimum of 1,200 newly and sequentially enrolled subjects at up to 350 sites. The primary endpoint, which is the proportion of patients with a composite peri-procedural (within 30 days of the procedure) death and stroke, plus ipsilateral stroke between day 31 and 1 year (365 days), will be compared to a performance goal of 8.4%. Clinical follow-up for all subjects will be performed at 24 hours post-procedure, 30 days, 1 year, and annually for a total of 3 years. The secondary endpoints include the composite rate of death and stroke at 30 days post-procedure, ipsilateral stroke at 2 and 3 years post-procedure, and annual rates of clinically driven target lesion revascularization through 3 years post-procedure. Additional analyses include:

- a. a comparison of the peri-procedural death and stroke rates for symptomatic subjects and asymptomatic subjects to predefined performance goals for each group;
  - b. a descriptive analysis of the peri-procedural death and stroke rate plus ipsilateral stroke at 1, 2, and 3 years for octogenarians;
  - c. a learning curve analysis based on information collected on operators' experience level.
2. You have agreed to provide a clinical update to physician users at least annually until the last patient in your post-approval study has reached their final endpoint. Please provide copies of these updates as part of your annual reports to FDA. At a minimum, this update will include a summary of the number of patients for whom data are available, with composite death and stroke rate at 30 days, and ipsilateral stroke at 31 days to 365 days, and annually to 3 years, and rates for freedom from target lesion revascularization and device or procedure-related events.

Be advised that the failure to conduct any such study in compliance with the good clinical laboratory practices in 21 CFR part 58 (if a non-clinical study subject to part 58) or the institutional review board regulations in 21 CFR part 56 and the informed consent regulations in 21 CFR part 50 (if a clinical study involving human subjects) may be grounds for FDA withdrawal of approval of the PMA.

FDA would like to remind you that you are required to submit PAS Progress Reports every six months. The PAS Progress Reports should be submitted separately from the Annual Reports. Two copies, identified as "PMA Post-Approval Study Report" and bearing the applicable PMA reference number, should be submitted to the address below. For more information on post-approval studies, see the FDA guidance document entitled, "Procedures for Handling Post-Approval Studies Imposed by PMA Order"

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070974.htm>

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39.

All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process" ([www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm)).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at [www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm](http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm).

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at [www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm](http://www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm).

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA CDRH Internet HomePage located at [www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm](http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm). Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in triplicate, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing. One of those three copies may be an electronic copy (eCopy), in an electronic format that FDA can process, review and archive (general information:


<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/ucm134508.htm>; clinical and statistical data:

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/ucm136377.htm>)

U.S. Food and Drug Administration  
Center for Devices and Radiological Health  
PMA Document Mail Center – WO66-G609  
10903 New Hampshire Avenue  
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Ms. Sadaf Toor at (301) 796-6381.

Sincerely yours,



Christy Foreman  
Office Director  
Office of Device Evaluation  
Center for Devices and Radiological Health