

NOV - 7 2005

K052437

510 (k) SUMMARY OF SAFETY AND EFFECTIVENESS
SUBSTANTIAL EQUIVALENCY

Submitter: Surgical Specialties Corporation
Address: 100 Dennis Drive
Reading, PA 19606

Telephone: 610 404 1000, ext. 2231
Contact Person: Elizabeth Lazaro
Regulatory Affairs Specialist

Common / Classification Name: Surgical Suture
Absorbable, Synthetic

Name of Device: Modified Monoderm™ Monofilament, Synthetic Absorbable, Dyed
and/or Undyed sutures.

Indications for Use: Indicated for use in general soft tissue approximation and/or ligation, but
not for use in cardiovascular or neurological tissues, microsurgery or
ophthalmic surgery.

Predicate Devices: Ethicon's Monocryl (Poliglecaprone 25) E caprolactone/Glycolide
510 (k) K930772
Surgical Specialties Corporation's (Polyglytone 72) Glycolide, and e-
caprolactone.
510 (k) K040477

Device Description: Monoderm™ suture is a monofilament synthetic absorbable surgical
suture prepared from a copolymer of glycolide and e-caprolactone.

Performance Data: Physical testing was performed on Monoderm™ Synthetic, Absorbable
sutures to USP 27, including <861> Suture Diameter, <871> Suture
Needle Attachment, <881> Tensile Strength. Animal testing was
performed for conformance to ISO 10993 for biocompatibility and
implant studies to demonstrate rates of tensile strength and mass loss.

Modified Monoderm™ Synthetic, Monofilament Absorbable Sutures
Surgical Specialties Corporation

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS (continued)

Technological Characteristics

Surgical Specialties Corporation's Monoderm™ Synthetic, absorbable suture is made from a mixture of Glycolide/e-caprolactone. The material has been well characterized through absorption studies and biocompatibility studies. The product is similar to the predicate devices, Monoderm (Polyglactone 72) and Monocryl in that they are both made from a polymer blend of synthetic absorbable materials.

Equivalency:

Absorbable, Synthetic, Monofilament

Surgical Specialties Corporation's Monoderm sutures are made of a synthetic absorbable suture material, which will dissolve essentially in 91 days.

Ethicon's Monocryl sutures are made of synthetic absorbable suture material which will dissolve essentially between 91 and 119 days.



NOV - 7 2005

Elizabeth Lazaro
Regulatory Affairs Specialist
Surgical Specialties Corporation
100 Dennis Drive
Reading, Pennsylvania 19606

Re: K052437
Trade/Device Name: Monoderm™ Synthetic Absorbable Surgical Suture
Regulation Number: 21 CFR 878.4493
Regulation Name: Absorbable poly(glycolide/L-lactide) surgical suture
Regulatory Class: II
Product Code: GAM
Dated: September 2, 2005
Received: September 6, 2005

Dear Ms. Lazaro:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>

Sincerely yours,



Mark N. Melkerson
Acting Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K052437

Device Name: **Monoderm™** Synthetic Absorbable Surgical Suture

Indications for Use:

Monoderm™ sutures are indicated for use in general soft tissue approximation and/or ligation, but not for use in cardiovascular or neurological tissues, microsurgery or ophthalmic surgery.

Prescription Use
 (Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
 (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Barbara Friedman for MCM

(Division Sign-Off)

Division of General, Restorative,
and Neurological Devices

Page 1 of 1

510(k) Number K052437