

MAR 24 2006

K060066 page 1 of 2

510(k) Summary
for
Albert Browne Ltd., Verify Gravity Flash Integrator

1. SPONSOR

Albert Browne Ltd., subsidiary of STERIS Corporation
Chancery House
190 Waterside Road
Hamilton Industrial Park
Leicester LE5 1QZ
United Kingdom

Contact: Richard Bancroft
Telephone: 0116 276 8636

Date Prepared: March 13, 2006

2. DEVICE NAME

Proprietary Name: Albert Browne, Ltd. Verify Gravity Flash Integrator
Common/Usual Name: Chemical indicator
Classification Name: Physical/chemical sterilization process indicator

3. PREDICATE DEVICE

- Verify Flash Integrator (K051101)
- TST Control Integrator for Steam Sterilizers (K002937)

4. DEVICE DESCRIPTION

The proposed Albert Browne, Ltd. Verify Gravity Flash Integrator is a 22 mm x 143 mm strip with two 12 mm circular chemical indicator ink spots, one located on either side of a reference circle exhibiting the endpoint color. The proposed integrator can be used to monitor 3 min. and 10 min. 132°C gravity flash steam sterilization cycles.

5. INTENDED USE

The Albert Browne Ltd. Verify Gravity Flash Integrator is an integrating integrator that changes color from yellow to blue/purple when exposed to the following conditions:

- 270°F (132°C), 3 min. flash gravity steam sterilization
- 270°F (132°C), 10 min. flash gravity steam sterilization

6. TECHNOLOGICAL CHARACTERISTICS AND SUBSTANTIAL EQUIVALENCE

The proposed Albert Browne, Ltd. Verify Gravity Flash Integrator (Verify Gravity Flash Integrator) is printed with two chemical indicator inks with stated values of 3 min. and 10 min. at 132°C. The 10 min. integrator ink printed on the proposed Verify Gravity Flash Integrator is identical to the 10 min. ink printed on the predicate Verify Flash Integrator. The 3 min. integrator ink printed on the proposed Verify Gravity Flash Integrator is identical to the indicator ink printed on the predicate TST Control Integrator. The differences between the proposed and predicate devices are limited to minor differences in substrates and configuration that do not raise any new issues of safety and effectiveness.

7. PERFORMANCE TESTING

The performance of both the 3 min. and 10 min. indicator inks used in the proposed Verify Gravity Flash Integrator was evaluated in the 510(k) premarket notifications for the predicate devices (K002937 and K051101). However, the substrates are slightly different for the proposed Verify Gravity Flash Integrator and predicate TST Control Integrator. Therefore, the performance testing for the 3 min. indicator ink was repeated using the Verify Gravity Flash Integrator. Performance testing was conducted to verify that the 3 min. indicator ink printed on the polypropylene substrate meets the requirements for Class 5 integrating indicators as defined in clauses 9.1 and 9.3 of ANSI/AAMI ST60-1996 "Sterilization of health care products – Chemical indicators – Part 1: General Requirements". Simulated use testing submitted in K060066 confirms that the Verify Gravity Flash Integrator performed as expected in 132°C 3 min. gravity flash steam sterilization cycles.



MAR 24 2006

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Albert Browne Limited
C/O Dr. Cynthia J. M. Nolte
Senior Regulatory Consultant
Medical Device Consultants, Incorporated
49 Plain Street
North Attleboro, Massachusetts 02760

Re: K060066
Trade/Device Name: Albert Browne Ltd., Verify Gravity Flash Integrator
Regulation Number: 880.2800
Regulation Name: Sterilization Process Indicator
Regulatory Class: II
Product Code: JOJ
Dated: March 13, 2006
Received: March 15, 2006

Dear Dr. Nolte:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number: K060066

Device Name: Albert Browne Ltd., Verify Gravity Flash Integrator

Indications for Use:

The Albert Browne Ltd. Verify Gravity Flash Integrator is an integrating integrator that changes color from yellow to blue/purple when exposed to the following conditions:

- 270°F (132°C), 3 min. flash gravity steam sterilization
- 270°F (132°C), 10 min. flash gravity steam sterilization


Prescription Use _____
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use x
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

 3/23/06
Shula A. Murphy, M.D.
Medical Director, General Hospital,
1000 W. 10th Street, Denver, CO 80202
510(k) - K060066