

K061177

Summary of Safety and Effectiveness

MAY 16 2006

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR § 807.93

Submitter	Implant Innovations, Inc. 4555 Riverside Palm Beach Gardens, FL 33410
Contact	Jim Banic Regulatory Affairs Specialist Implant Innovations, Inc. 4555 Riverside Palm Beach Gardens, FL 33410 Tel. 561-776-6932 Fax. 561-514 6316 Email: jbanic@3implant.com
Date Prepared	25 April 2006
Device Name	Provide Temporary Cylinders
Classification Name	Endosseous Dental Implant Abutment
Device Classification	Class II Dental Devices Panel 21 CFR § 872.3630
Predicate Devices	PreFormance Posts -> K053170
Performance	Performance standards have not been established by the FDA under Section 514 of the Federal Food, Drug and Cosmetic Act.
Device Description	The Provide Temporary Cylinders will be made of the same material (PEEK) as the PreFormance Posts. The cylinder is cemented onto an abutment. An acrylic resin provisional restoration can be built onto the cylinder. The height of these cylinders will be 6.5mm and the platform diameters are 4.8 and 6.5mm. Furthermore, they will be used as single use to support a single-unit restoration and multi use where two or more cylinders are splinted together to support a provisional multi-unit restoration.

Implant Innovations, Inc.

Special 510(k) Premarket Notification – Provide™ Temporary Cylinders: Device Modification

Indications for Use

The Provide Temporary Cylinders are intended for use as an accessory to endosseous dental implants to support a prosthetic device in a partially or fully edentulous patient. They are intended for use to support single and multiple unit prostheses in the mandible or maxilla for up to 180 days during endosseous and gingival healing, and are for non occlusal loading of single and multiple unit provisional restorations. The prostheses can be cement retained to the abutment.

Technological Characteristics

The Provide Temporary Cylinders are made of the same material and contain features and functions which are similar to the currently available PreFormance Posts.

Conclusion

The Provide Temporary Cylinders are substantially equivalent to the legally marketed PreFormance Posts.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAY 16 2006

Mr. Jim Banic
Regulatory Affairs Specialist
3i Implant Innovations, Incorporated
4555 Riverside Drive
Palm Beach Gardens, Florida 33410

Re: K061177
Trade/Device Name: Provide Temporary Cylinders
Regulation Number: 872.3630
Regulation Name: Endosseous Dental Implant Abutment
Regulatory Class: II
Product Code: NHA
Dated: April 25, 2006
Received: April 27, 2006

Dear Mr. Banic:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K0601177

Device Name: Provide Temporary Cylinders

Indications for Use:

The Provide Temporary Cylinders are intended for use as an accessory to endosseous dental implants to support a prosthetic device in a partially or fully edentulous patient. They are intended for use to support single and multiple unit prostheses in the mandible or maxilla for up to 180 days during endosseous and gingival healing, and are for non occlusal loading of single and multiple unit provisional restorations. The prostheses can be cement retained to the abutment.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use: OR Over the Counter Use:
(Per 21 CFR 801.109)


Susan Runna
Department of Anesthesiology, General Hospital,
Device Control, Dental Devices
510(k) Number: K0601177