



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB - 6 2007

Shanghai Korea Economic and Trade Company
c/o Mr. Zhao J. Yu
Rome. 902, No. 15, Lane 28
Qingjiang, China 200233

Re: K062893

Trade Name: Blood Pressure Meter Model: XJ-2001AC, XJ-2002AS and XJ-2003A
Regulation Number: 21 CFR 870.1130
Regulation Name: Noninvasive blood pressure measurement system
Regulatory Class: Class II
Product Code: DXN
Dated: November 16, 2006
Received: January 11, 2007

Dear Mr. Yu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must

comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in cursive script, appearing to read "Bram D. Zuckerman for".

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Shanghai Kodea Economic and Trade Development Limited Company.

510(k) Submission Application

Indications for Use

510(k) Number (if known): K062893

Device Name: Blood Pressure Meter Model: XJ-2001AC.

Blood Pressure Meter Model: XJ-2002AS and

Blood Pressure Meter Model: XJ-2003A

Indications for Use:

The blood pressure meter Models: XJ-2001AC, XJ-2002AS and XJ-2003A are non-invasive blood pressure measurement systems which are intended to measure the diastolic and systolic blood pressure, pulse rate of the adult individual through using inflatable cuff at home. The cuff circumference is limited to 8.6614 inches to 12.60 inches.

Prescription Use _____
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, office of Device Evaluation (ODE)

B. J. [Signature]
Division Sign-Off
Division of Cardiovascular Devices
510(k) Number K062893

Page 1 of 1

A4