



Päivi Sormunen
Thermo Electron Oy
Ratastie 2
P.O. Box 100
FIN-01621 Vantaa
Finland

MAR 21 2007

Re: k063209
Trade Name: HbA1c Test System, Calibrator, Control normal, and Control abnormal
Regulation Number: 21 CFR 864.7470
Regulation Name: Glycosylated hemoglobin assay
Regulatory Class: Class II
Product Code: LCP, KRZ, JJX
Dated: February 07, 2007
Received: March 09, 2007

Dear Mr. Sormunen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0490. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address at <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,


Jean M. Cooper, M.S., D.V.M.

Director
Division of Chemistry and Toxicology
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): k063209

Device Name: HbA1c Test System, HbA1c Calibrators, HbA1c Control Normal, HbA1c Control Abnormal

Indications For Use:

The hemoglobin A1c (HbA1c) test system with associated calibrators and controls is intended for quantitative in-vitro diagnostic determination of the hemoglobin A1c (HbA1c) concentration as a percentage of total hemoglobin in human whole blood using T60 Clinical Chemistry Analyzers. Measurement of percent HbA1c is effective in monitoring long-term glucose control in individuals with diabetes mellitus.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)


Division Sign-Off

Office of In Vitro Diagnostic Device
Evaluation and Safety

Page 1 of 1

510(k) k063209