



Food and Drug Administration  
2098 Gaither Road  
Rockville MD 20850

Pi-Shiou Li  
Management Representative  
TaiDoc Technology Corporation  
4F, No. 88, Sec. 1, Kwang-Fu Rd., San-Chung  
Taipei County, Taiwan 241

NOV 22 2006

Re: k063212  
Trade/Device Name: Clever Chek TD-4231 Blood Glucose Monitoring System  
Clever Chek TD-4223 Blood Glucose Monitoring System  
Regulation Number: 21 CFR 862.1345  
Regulation Name: Glucose test system  
Regulatory Class: Class II  
Product Code: NBW, CGA  
Dated: October 18, 2006  
Received: October 23, 2006

Dear Ms. Li:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

Page 2 --

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0484. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Alberto Gutierrez, Ph.D.  
Director  
Division of Chemistry and Toxicology  
Office of In Vitro Diagnostic Device  
Evaluation and Safety  
Center for Devices and  
Radiological Health

Enclosure

**Attachment 2**

**Indications for Use**

510(k) Number: K063212

Device Name: Clever Chek TD-4231 Blood Glucose Monitoring System  
Clever Chek TD-4223 Blood Glucose Monitoring System

Indications for Use:

The Clever Chek TD-4231/ Clever Chek TD-4223 Blood Glucose Monitoring System is intended for use in the quantitative measurement of glucose in fresh capillary whole blood from the finger and the following alternative sites: the palm, the forearm, the upper-arm, the calf and the thigh. It is intended for use by healthcare professionals and people with diabetes mellitus at home as an aid in monitoring the effectiveness of diabetes control program. It is not intended for the diagnosis of or screening for diabetes mellitus, and is not intended for use on neonates.

The alternative site testing in the Clever Chek TD-4231/ Clever Chek TD-4223 Blood Glucose Monitoring System can be used only during steady-state blood glucose conditions.

Prescription Use X AND/OR Over-The-Counter Use X  
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Carol Benson  
Division Sign-Off

Page 1 of 1

Office of In Vitro Diagnostic Device  
Evaluation and Safety

510(k) K063212

A2