

K070560

510(K) SUMMARY

LifeCell Corporation's LTM Surgical Mesh

JUN 11 2007

Submitter's Name, Address, Telephone Number, Contact Person and Date Prepared

LifeCell Corporation
One Millennium Way
Branchburg, NJ 08869

Phone: (908) 947-1115

Facsimile: (908) 947-1095

Contact Person: Lorraine T. Montemurro, R.N., R.A.C.

Date Prepared: June 6, 2007

Name of Device and Name/Address of Sponsor

LTM Surgical Mesh

LifeCell Corporation
One Millennium Way
Branchburg, NJ 08869

Common or Usual Name

Surgical Mesh

Classification Name

Surgical Mesh

Predicate Devices

Permacol® Implant (K050355)

Surgisis® (K034039)

Intended Use / Indications for Use

LTM Surgical Mesh (LTM) is intended for use as a soft tissue patch to reinforce soft tissue where weakness exists and for the surgical repair of damaged or ruptured soft tissue membranes. Indications for use include the repair of hernias and /or body wall defects which require the use of reinforcing or bridging material to obtain the desired surgical outcome.

LTM is intended for single patient use only.

Technological Characteristics

The LTM is a surgical mesh that is derived from porcine skin. The LTM device consists of a terminally sterilized sheet of the processed porcine dermal matrix provided in prescribed geometric configurations and thicknesses, and packaged in double pouch configuration.

Performance Data

The LTM has undergone extensive biocompatibility testing, animal testing, viral inactivation testing and biomechanical testing. The data indicate that the device is biocompatible and that the manufacturing process is capable of inactivating any viral components that may come with the starting material. The biomechanical data show that the LTM matrix possesses sufficient strength and suture retention for the intended use.

Substantial Equivalence

LTM is substantially equivalent to legally marketed predicate devices, such as the Permacol® Implant (K050355) and Surgisis® (K034039) surgical mesh devices. LTM has the same intended uses and the same or similar indications, technological characteristics, and principles of operation as these predicate devices. Performance data demonstrate that LTM functions equivalently to the predicate devices. Thus, LTM is substantially equivalent.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

LifeCell Corporation
% Ms. Lorraine T. Montemurro, R.N., RAC
Manager, Regulatory Affairs
One Millennium Way
Branchburg, New Jersey 08876

JUN 11 2007

Re: K070560
Trade/Device Name: LTM Surgical Mesh
Regulation Number: 21 CFR 878.3300
Regulation Name: Surgical mesh
Regulatory Class: II
Product Code: FTM
Dated: May 7, 2007
Received: May 8, 2007

Dear Ms. Montemurro:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Ms. Lorraine T. Montemurro, R.N., RAC

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Mark N. Melkerson', is written over a faint, illegible typed name.

Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K070560

Device Name: LTM Surgical Mesh

Indications for Use:

LTM Surgical Mesh is intended for use as a soft tissue patch to reinforce soft tissue where weakness exists and for the surgical repair of damaged or ruptured soft tissue membranes. Indications for use include the repair of hernias and /or body wall defects which require the use of reinforcing or bridging material to obtain the desired surgical outcome.

LTM is intended for single patient use only.

Prescription Use X
(Part 21 C.F.R. 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 C.F.R. 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE -- CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page _____ of _____



(Division Sign-Off)
Division of General, Restorative,
and Neurological Devices

510(k) Number K070560