

MAY - 2 2008

5. 510(k) Summary

Contact: Mr. Justin Eggleton
Musculoskeletal Clinical & Regulatory Advisers, LLC
1331 H Street NW, 12th Floor
Washington, DC 20005
202.552.5800

Device Trade Name: DSS™ Stabilization System

Manufacturer: Paradigm Spine, LLC
505 Park Ave. 14th Floor
New York, NY 10022

Classification: 21 CFR §888.3070, Pedicle screw spinal system

Class: II

Product Code: NQP

Indications For Use:

The DSS™ Stabilization System is intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative spondylolisthesis with objective evidence of neurologic impairment, kyphosis, and failed previous fusion (pseudarthrosis).

In addition, the DSS™ Stabilization System is indicated for use in patients:

- Who are receiving fusions with autogenous graft only;
- Who are having the device fixed or attached to the lumbar or sacral spine;
- Who are having the device removed after the development of a solid fusion mass.

Device Description:

The DSS™ Stabilization System is comprised of a variety of pedicle screws sizes, and couplers that act as longitudinal spacers that are uniquely fitted for each individual case. The pedicle screws and couplers are manufactured from medical grade titanium alloy (Ti6Al4V).

Predicate Device(s):

DSS™ Stabilization System was shown to be substantially equivalent to previously cleared devices and has the same indications for use, design, function, and materials used.

Performance Standards:

Testing performed indicate that the DSS™ is as mechanically sound as predicate devices.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Musculoskeletal Clinical Regulatory Advisers, LLC
% Mr. Justin Eggleton
1331 H Street NW, 12th Floor
Washington, DC 20005

MAY - 2 2008

Re: K080963
Trade/Device Name: DSS™ Stabilization System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class II
Product Code: NQP
Dated: April 3, 2008
Received: April 4, 2008

Dear Mr. Eggleton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act and the limitations described below. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

The Office of Device Evaluation has determined that there is a reasonable likelihood that this device will be used for an intended use not identified in the proposed labeling and that such use could cause harm. Therefore, in accordance with Section 513(i)(1)(E) of the Act, the following limitation must appear in the Warnings section of the device's labeling:

The safety and effectiveness of this device for the indication of spinal stabilization without fusion have not been established.

Please note that the above labeling limitations are required by Section 513(i)(1)(E) of the Act. Therefore, a new 510(k) is required before these limitations are modified in any way or removed from the device's labeling.

Page 2 – Mr. Justin Eggleton

The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and permits your device to proceed to the market. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification if the limitation statement described above is added to your labeling.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at toll-free number (800) 638-2041 or (240) 276-3150 or the Internet address <<http://www.fda.gov/cdrh/industry/support/index.html>>.

Sincerely yours,

for 
Donna-Bea Tillman, Ph.D., M.P.A.
Director
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Indications for Use

510(k) Number (if known): _____

Device Name: DSS™ Stabilization System

The DSS™ Stabilization System is intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative spondylolisthesis with objective evidence of neurologic impairment, kyphosis, and failed previous fusion (pseudarthrosis).

In addition, the DSS™ Stabilization System is indicated for use in patients:

- Who are receiving fusions with autogenous graft only;
- Who are having the device fixed or attached to the lumbar or sacral spine;
- Who are having the device removed after the development of a solid fusion mass.


Prescription Use ✓
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

510(k) Number K080963