

K081272
Pg 1 of 2

JUL 31 2008

SECTION 16: SUMMARY OF SAFETY AND EFFECTIVENESS

This 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and CFR 807.92.

16.1 SUBMITTER INFORMATION

- a. Company Name: Brennen Medical, LLC
- b. Company Address: 1290 Hammond Road
St. Paul, MN 55110
- c. Company Phone: (651) 429-7413
Company Facsimile: (651) 429-8020
- d. Contact Person: Kenneth B. Herland
V.P. Regulatory Affairs/QA
- e. Date Summary Prepared: July 28, 2008

16.2. DEVICE IDENTIFICATION

- a. Trade/Proprietary Name: Porcine Dermal Matrix (TBD)
- b. Regulation Number : 21 CFR 878.3300
- c. Regulation Name : Surgical Mesh
- d. Device Class : II
- e. Product Code : FTM

16.3 IDENTIFICATION OF PREDICATE DEVICES

<u>Company</u>	<u>Device</u>	<u>510(k) No.</u>	<u>Date Cleared</u>
Advanced Uroscience	Porcine Surgical Mesh	K993459	11/05/1999
Brennen Medical	Porcine Surgical Mesh	K030460	03/07/2003
Brennen Medical	DermMatrix	K021160	06/12/2002
Tissue Sciences Labs	Permacol/Pelvicol	K013625	01/17/2002
TEI Biosciences	SurgiMend	K071807	08/06/2007

16.4 DEVICE DESCRIPTION

Porcine Dermal Matrix is a prescription, sterile, pyrogen free, single use, porcine skin

that has both the epidermal and subdermal sides removed. The product is available in several sizes.

16.5 SUBSTANTIAL EQUIVALENCE

Porcine Dermal Matrix is substantially equivalent to DermMatrix, Permacol/Pelvicol, and SurgiMend surgical meshes. Porcine Dermal Matrix is equivalent in intended use, mode of action, and design to the predicate devices. The introduction of this product does not raise any new issues of safety or effectiveness.

16.6 INTENDED USE

Intended for implantation to reinforce soft tissue where weakness exists and for surgical repair of damaged or ruptured soft tissue membranes. Specifically indicated for: Plastic and reconstructive surgery; Muscle flap reinforcement; Hernia repair including abdominal, inguinal, femoral, diaphragmatic, scrotal, umbilical, and incisional hernias; Suture-line reinforcement; Reinforcement of the rotator cuff, patellar, Achilles, biceps, quadriceps, or other tendons. Porcine Dermal Matrix is not intended to replace normal body structure or provide the full mechanical strength to support tendon repair of the rotator cuff, patellar Achilles, biceps, quadriceps, or other tendons.

16.7 TECHNOLOGICAL CHARACTERISTICS

The technological characteristics are identical to the predicate devices (Porcine Surgical Mesh and DermMatrix). Biocompatibility testing conducted per ISO 10993-1, bench testing and numerous clinical experiences have demonstrated that the device is safe and effective for its intended use, and that its performance is substantially equivalent to the predicate devices.

16.8 CONCLUSIONS

This notification contains all information required by 21 CFR 807.87. A completed copy of the Premarket Notification 510(k) Reviewer's Checklist is provided in this submission. Test evaluations of Porcine Dermal Matrix show that the device performs as intended and substantially equivalent to the predicate devices.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Brennen Medical, LLC
% Mr. Kenneth B. Herland
VP Regulatory Affairs/QA
1290 Hammond Road
Saint Paul, Minnesota 55110

JUL 31 2008

Re: K081272

Trade/Device Name: Porcine Dermal Matrix Surgical Mesh
Regulation Number: 21 CFR 878.3300
Regulation Name: Surgical mesh
Regulatory Class: II
Product Code: FTM
Dated: May 1, 2008
Received: May 5, 2008

Dear Mr. Herland:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Kenneth B. Herland

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K081272


Indications for Use

510(k) Number (if known): K081272

Device Name: Porcine Dermal Matrix Surgical Mesh

Indications for Use:

Intended for implantation to reinforce soft tissue where weakness exists and for surgical repair of damaged or ruptured soft tissue, including but not limited to: Plastic and reconstructive surgery; Muscle flap reinforcement; Hernia repair including abdominal, inguinal, femoral, diaphragmatic, scrotal, umbilical, and incisional hernias; Suture-line reinforcement; Reinforcement of the rotator cuff, patellar, Achilles, biceps, quadriceps, or other tendons. Porcine Dermal Matrix is not intended to replace normal body structure or provide the full mechanical strength to support tendon repair of the rotator cuff, patellar Achilles, biceps, quadriceps, or other tendons. Sutures used to repair the tear and sutures or bone anchors used to attach the tissue to the bone provide biomechanical strength for the tendon repair.



(Division Sign-Off)
Division of General, Restorative,
and Neurological Devices

510(k) Number K081272

Prescription Use (X)
(Part 21 CFR 801 Subpart D)
(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page ___ of ___
(Posted November 13, 2003)