

N0000
K082880

510 (K) Summary

SUMMARY OF SAFETY AND EFFECTIVENESS FOR
EZvue (Ocufilecon D) Soft (hydrophilic) Contact Lens for Daily Wear JUN 16 2009

Submitter Information:

Company: **I-SEE VISION TECHNOLOGY INC.**
1F, No. 45, Lane 2, Section 2, Kuang-Fu Rd., Hisnchu, Taiwan,
30071
Contact Person: Mr. Tom Lin,
Special Assistant of President Office
Phone: 886-3-5752822
FAX: 886-3-5752823
E-Mail: tomlin@iseevision.com

Identification of Device:

Classification Name: Lens, Contact, (Disposable)
Trade Name: EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens
(or EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact
Lens for daily wear)
Common or usual Name: Soft (hydrophilic) Contact lens (daily wear)
FDA Classification: Class II

Predicate Device:

Hydron Biomedics 55 (Ocufilecon D) Visibility Tint Molded daily wear soft contact lens
cleared via K972303 Ocular Sciences / American Hydron Inc.

Indications for Use

EZvue(Ocufilecon D)1-day Soft (hydrophilic) Contact Lens for daily wear is indicated
for daily wear for the correction of refractive ametropia (myopia and hyperopia) in
aphakic and/or not-aphakic persons with non-diseased eyes who may exhibited
refractive astigmatism up to 2.00 diopters that does not interfere with visual acuity..

Eyecare practitioners may prescribe the contact lens for single-use disposable daily
wear. The EZvue(Ocufilecon D)1-day Soft (hydrophilic) Contact Lenses are not
intended to be disinfected and should be discarded after a single use.

Description of Device

EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear is available as
spherical lenses manufactured by cast molded method. The lens is made from a lens
material that is approximately 55% water. The hydrogel lens' material is a copolymer

composed of 2-hydroxyethyl methacrylate (HEMA) and methacrylic acid (MAA), which was cross-linked with Ethylene Glycol Dimethacrylate (EGDMA) via UV photopolymerization. The EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens with visible tint is tinted blue using Reactive Blue Dye #19 to make the lens more visible for handling. Lenses are supplied sterile in sealed blister packers containing sterile isotonic phosphate buffered saline.

Summary of Clinical Study:

The EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lenses for Daily Wear were tested over 60 human eyes separately within 6 months. Nearly a hundred percent of the participants' vision was corrected and nearly all were satisfied with the lens wearing and care of lenses. In general, these products are good and safe for customers.

Nonclinical Studies:

A series of nonclinical performance tests were performed to demonstrate the safety and effectiveness of the EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear, and establish substantial equivalence to predicate lenses - Hydron Biomedics 55 (Ocufilecon D) Visibility Tint Molded daily wear soft contact lens (K972303). All testing was conducted in accordance with the May 1994 FDA guideline titled *Premarket Notification 510(K) Guidance Document for Class IV Contact lenses*, and in conformance to applicable device regulations. The evidence of substantial equivalent to the predicate lens described as follow:

a) Technological characteristics studies

EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear designs in the following parameter ranges:

Diameter range; 13.90~14.30mm

Power range: +8.00D~-12.00D

Center thickness: varies with power (0.080~0.150mm)

Lenses have the following properties:

Refractive index: 1.410 (hydrated)

Light transmittance: >93%

Water content: 53 to 57 %

Oxygen permeability (edged corrected): $16 \times 10^{-11} [(cm^2/sec)(mlO_2/ml-mmHg)] @ 35^\circ C$

There characterizations of EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lenses are equivalent and comparable to those of predicate lenses.

b) Biocompatibility

In accordance with the May 1994 Guidance Document for daily wear contact lenses,

toxicity studies have been conducted on the model: EZvue (Ocufileon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear. The Irritation test in the rabbit eye and Systemic toxicity studies indicate the extracts would be considered as non-toxic and nor irritated. The Cytotoxicity testing demonstrates the lens is not cytotoxic under the conditions of the study.

c) Microbiology

Steam sterilization process has been validated to deliver a minimum SAL of 10^{-6} , thereby complying with the requirement of FDA Group IV. There is shelf-life stability data supporting that the lens remains sterile through the expiration date claimed for the product.

d) Leachability

Studies were conducted to determine the leachable materials from the finished lens. The results show that, at the levels of the detection reported, there are no leachable monomers and addictive residues.

Substantial equivalence Statement:

Testing performed on the EZvue (Ocufileon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear indicated that it can support the efficiency and security use as well as the predicate devices- Hydron Biomedics 55 (Ocufileon D) Visibility Tint Molded daily wear soft contact lens (K972303), when used in accordance with the instructions for use. It is due to the facts that The risks and benefits of the subject device are the same as soft contact lenses for to the daily wear.

In conclusion, it is I-SEE's conviction that data submitted in this 510(k) to validate the claim of substantial equivalency, substantiates our ability to manufacture a soft contact lens, the EZvue (Ocufileon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear, with the same established safety profile and effectiveness as the predicate device-- Hydron Biomedics 55 (Ocufileon D) Visibility Tint Molded daily wear soft contact lens (K972303).



DEPARTMENT OF HEALTH & HUMAN SERVICES

SEP 9 2009

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G609
Silver Spring, MD 20993-0002

I-See Vision Technology Inc.
c/o Ms. Jennifer Reich
Senior Consultant
Harvest Consulting Corp.
2904 N. Boldt Drive
Flagstaff, AZ 86001

Re: K082880

Trade/Device Name: EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear

Regulation Number: 21 CFR 886.5925

Regulation Name: Soft (hydrophilic) contact lenses

Regulatory Class: Class II

Product Code: LPL, MVN

Dated: June 9, 2009

Received: June 11, 2009

Dear Ms. Reich:

This letter corrects our substantially equivalent letter of June 16, 2009.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other

requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Malvina B. Eydelman, M.D.

Director

Division of Ophthalmic, Neurological and Ear,

Nose and Throat Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K082880

Device Name: EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear
I-SEE Vision Technology, Inc.

Indications For Use:

EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lens for Daily Wear is indicated for daily wear for the correction of refractive ametropia (myopia and hyperopia) in aphakic and/or not-aphakic persons with non-diseased eyes who may exhibited refractive astigmatism up to 2.00 diopters that does not interfere with visual acuity.

Eyecare practitioners may prescribe the contact lens for single-use disposable daily wear. The EZvue (Ocufilecon D) 1-Day Soft (hydrophilic) Contact Lenses are not intended to be disinfected and should be discarded after a single use.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Division of Ophthalmic, Neurological and Ear,
Nose and Throat Devices

510(k) Number K082880