

K092101

510(k) Summary

Submitter: Nonin Medical, Inc.

Contact Person: Lori M. Roth
Clinical/Regulatory Specialist
Nonin Medical, Inc.
13700 1st Ave. North
Plymouth, MN 55441-5443

Date Prepared: July 9, 2009

Trade Name: Oximeter Sensor

**Classification Name:
and Number:** Class II, 21 CFR 870.2700

Product Code: 74 DQA

Predicate Device(s): Nonin's 8000AA sensor as cleared in the following 510(K) submissions: Model 7500 (K080255 cleared on May 23, 2008), Model LS1-9R LifeSense (K063752 cleared on May 4, 2007), Model 9600 (K040589 cleared on June 3, 2004), Model 4000 (K052669 cleared on December 23, 2005), Model 2500A (K050056 cleared on June 21, 2005), Model 2500 (K002690 cleared on October 11, 2000), Model 2120 (K031487 cleared on March 22, 2004).

Envitec's SoftTip Reusable Sensor (Small, Medium and Large), cleared under K992215, 29 March 2000.

Device Description: The 8000SX is a fingertip transmittance sensor comprised of a molded silicone boot that contains identical sensor optics and cable as the currently marketed Model 8000AA. The 8000SX sensor comes in three sizes, Model 8000SL (Large) for finger heights between 0.5 inches and 1.0 inch, Model 8000SM (Medium) for finger heights between 0.4 inches and 0.75 inches, and Model 8000SS (Small) for finger heights between 0.3 inches and 0.5 inch. The sensors are compatible with all Nonin branded pulse oximeters.

OCT 21 2009

Intended Use:

Nonin's Model 8000SX Reusable Sensors are indicated for non-invasive spot-checking and/or continuous monitoring of adult and pediatric patients who are well or poorly perfused. It is intended for use in environments including operating room, surgical recovery, critical care, emergency room, long-term care, home use and mobile environments

Functional and Safety Testing:

Nonin's Model 8000SX sensor have successfully undergone both bench and clinical testing in order to demonstrate that it meets the requirements of ISO 9919:2005 Clause 50 Accuracy of Operating Data, Clause 102 section 102.2 Labeling, and IEC 60601-1:1998 (ISO 10993-1:2003) Clause 48 Biocompatibility.

Conclusion:

The Model 8000SX sensor is substantially equivalent to Nonin's Model 8000AA and Envitec's SoftTip sensors when used with Nonin branded Pulse Oximeters monitors.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G609
Silver Spring, MD 20993-0002

Ms. Lori Roth
Clinical/Regulatory Specialist
Nonin Medical, Incorporated
13700 1st Avenue North
Plymouth, Minnesota 55441

OCT 21 2009

Re: K092101
Trade/Device Name: Nonin Medical, Inc. Model 8000SX Sensor
Regulation Number: 21 CFR 870.2700
Regulation Name: Oximeter
Regulatory Class: II
Product Code: DQA
Dated: September 18, 2009
Received: September 21, 2009

Dear Ms. Roth:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucml115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Susan Runner, D.D.S., M.A.
Acting Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k)
Number
(if known)

K092101

Device Name Nonin Medical, Inc. Model 8000SX Sensor Series

Indications
for Use


Nonin's Model 8000SX Reusable Sensors are indicated for non-invasive spot-checking and/or continuous monitoring of adult and pediatric patients who are well or poorly perfused. It is intended for use in environments including operating room, surgical recovery, critical care, emergency room, long-term care, home use and mobile environments.

Prescription Use X
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K092101