

**ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits
510(k) Summary
21 CFR 807.92**

As required by the Safe Medical Devices Act of 1990, coded under Section 513, Part (l)(3)(A) of the Food, Drug and Cosmetic Act, a summary of the information upon which substantial equivalence determination is based as follows:

Submitter Information:

JAN 14 2010

Applicant: Bard Peripheral Vascular, Inc
1625 West 3rd Street
PO Box 1740
Tempe, Arizona 85280
Phone: 480-638-2906
Fax: 480-449-2546
Contact: Joni Creal
Regulatory Affairs Associate

Subject Device:

Device Trade Name: ECLIPSE™ Filter System – Femoral Delivery Kit and
ECLIPSE™ Filter System – Jugular/Subclavian
Delivery Kit
Common or Usual Name: Filter, Intravascular, Cardiovascular
Classification: Class II
Classification Panel: Cardiovascular

Predicate Devices:

G2 EXPRESS® Filter System – Femoral Delivery Kits (K082305)

G2 EXPRESS® Filter System –Jugular/Subclavian Delivery Kits (K082305)

Summary of Change:

The primary modification from the predicate device, the G2® EXPRESS Filter System – Femoral and Jugular/Subclavian Delivery System, to the subject device, the ECLIPSE™ Femoral Filter System and Jugular/Subclavian Delivery System, was an improvement of the surface finish of the filter raw material wire by electropolishing the wire prior to forming the filter. A cosmetic colorant modification to some of the molded components of the delivery kits was also made.

Device Description:

The ECLIPSE™ Filter consists of twelve electropolished shape-memory nitinol wires emanating from a central electropolished nitinol filter hook. These 12 wires form two levels of embolic filtration: the legs provide the lower level of filtration and the arms provide the upper level of filtration. The ECLIPSE™ Filter is intended to be used in the inferior vena cava with diameters less than or equal to 28 mm.

The ECLIPSE™ Filter System – Femoral Delivery Kit consists of a 7 French inner diameter (I.D.) introducer catheter and dilator set and a storage tube preloaded with the ECLIPSE™ Filter and pusher system. The dilator is fitted with 2 radiopaque marker bands spaced 28 mm apart for caval sizing. The introducer sheath contains a radiopaque tip and hemostasis valve with a side port for injecting contrast medium via a syringe. The storage tube and pusher system attach to the introducer and allow for delivery and deployment of the ECLIPSE™ Filter.

The ECLIPSE™ Filter System – Jugular/Subclavian Delivery Kit consists of a 10 French I.D. introducer catheter and dilator set and a delivery device preloaded with the ECLIPSE™ Filter. The dilator is fitted with 2 radiopaque marker bands spaced 28 mm apart for caval sizing. The introducer sheath contains a radiopaque tip and hemostasis valve with a side port for injecting contrast medium via a syringe. The delivery device fits within the introducer sheath and has a delivery mechanism to deploy the ECLIPSE™ Filter.

Indications for Use of Device:

The ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits are indicated for use in the prevention of recurrent pulmonary embolism via permanent placement in the vena cava in the following situations:

- Pulmonary thromboembolism when anticoagulants are contraindicated.
- Failure of anticoagulant therapy for thromboembolic disease.
- Emergency treatment following massive pulmonary embolism where anticipated benefits of conventional therapy are reduced.
- Chronic, recurrent pulmonary embolism where anticoagulant therapy has failed or is contraindicated.
- ECLIPSE™ Filter may be removed according to the instructions supplied under the section labeled: Optional Procedure for Filter Removal.

Technological Comparison to Predicate Device:

The technological characteristics of the subject device, the ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits, are substantially equivalent to those of the predicate devices, the G2® EXPRESS Filter System – Femoral and Jugular/Subclavian Delivery System (K082305), in terms of intended use, indications for use, application, user population, same operating principle, basic design, fundamental scientific technology, performance, packaging configuration and sterilization method.

Conclusions:

The ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits met all predetermined acceptance criteria of design verification and validation as specified by applicable standards, guidance, test protocols and/or customer inputs. The ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits is substantially equivalent to the legally marketed predicate device, the G2® EXPRESS Filter System – Femoral and Jugular/Subclavian Delivery System Kits (K082305 – cleared October 31, 2008).



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

JAN 14 2010

Bard Peripheral Vascular, Inc.
c/o Ms. Joni Creal
Regulatory Affairs Associate
1625 West Third Street
Tempe, AZ 85280-1749

Re: K093659

Trade/Device Name: ECLIPSE Filter System, Femoral and Jugular/Subclavian Delivery Kits
Regulation Number: 21 CFR 870.3375
Regulation Name: Cardiovascular intravascular filter
Regulatory Class: Class II (two)
Product Code: DTK
Dated: December 17, 2009
Received: December 18, 2009

Dear Ms. Creal:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



 Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K093659

Device Name: ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits

Indications for Use:

The ECLIPSE™ Filter System – Femoral and Jugular/Subclavian Delivery Kits are indicated for use in the prevention of recurrent pulmonary embolism via permanent placement in the vena cava in the following situations:

- Pulmonary thromboembolism when anticoagulants are contraindicated.
- Failure of anticoagulant therapy for thromboembolic disease.
- Emergency treatment following massive pulmonary embolism where anticipated benefits of conventional therapy are reduced.
- Chronic, recurrent pulmonary embolism where anticoagulant therapy has failed or is contraindicated.
- ECLIPSE™ Filter may be removed according to the instructions supplied under the section labeled: Optional Procedure for Filter Removal.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Diana R. Kohnen
(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K093659