

510(k) Summary of Safety and Effectiveness (in accordance to 21 CFR 807.87(h))**Device Name****Proprietary Device Name:** 3Di**Establishment Name and Registration Number of Submitter**

JAN 19 2010

Name: Shina Systems Ltd.**Corresponding Official:** Dan Laor**Sireni 6, Haifa 32972, Israel TEL:** 972-4-8246632**Device Classification****Product Code:** LLZ**CFR section:** 892.2050**Panel Identification:** Radiology**Device Description:** Picture archiving and communications system**Classification:** II**Reason for 510(k) Submission**

Traditional 510(k) Submission

Identification of Legally Marketed Predicate Devices

K070226 CardioCT

K082269 Visage PACS/CS

K090462 Philips Brilliance

Device Description

3Di is a PACS device which enables users to access medical images over a network and to utilize 3Di's image visualization tools to review the images. It provides the following functions: Web server, patient browser, PACS capabilities, multi-modality viewing, CT Cardiac and Colonoscopy clinical applications.

Intended use and indications for Use

3Di is a software package of PACS workstation for handling multimodality (CT, XA, MR, PET, SPECT & Ultrasound) images, which are using DICOM protocol. It includes volume rendering, Multi-planar reconstruction (MPR) and viewing of the inner and outer surfaces of organs as well as within their walls. 3Di is intended for use as an interactive tool for assisting professional Radiologists, Cardiologists and specialists to reach their own diagnosis, by providing tools of communication, clinics networking, WEB Serving, image viewing, image manipulation, 2D/3D image visualization, image processing, reporting and archiving. This product is not intended for use with or for diagnostic interpretation of Mammography images. The 3Di indications for use are processing of Cardiac CT angiography studies, including coronaries analysis, cardiac functional assessment and CT colonoscopy

Safety & Effectiveness

The device has been designed, verified and validated complying to 21CFR 820.30 regulations. The device has been designed to meet the requirements of ISO 14971 Safety standard. Its performance has been validated by comparison to the performance of the Philips Brilliance predicate device. The following functions have been compared: General functionality of the images reformatting for the different imaging modalities, Reliability of orientation annotations displayed over the images, correctness of measurements, Image quality, Cardiac analysis Graphs and Results and Colon analysis results.

The comparison demonstrates that the results of the two devices are very similar and substantial equivalent in terms of performance.

Substantial Equivalency

It is Shina System opinion that the 3Di is substantially equivalent in terms of safety and effectiveness to the predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room – WO66-G609
Silver Spring, MD 20993-0002

Shina Systems, Ltd.
% Mr. Dan Laor
Quality & Regulatory Advisor
6 Sireni, Haifa, 32972
ISRAEL

JAN 19 2010

Re: K093703
Trade/Device Name: 3Di
Regulation Number: 21 CFR 892.2050
Regulation Name: Picture archiving and communications system
Regulatory Class: II
Product Code: LLZ
Dated: November 24, 2009
Received: December 1, 2009

Dear Mr. Laor:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Donald J. St. Pierre
Acting Director
Division of Radiological Devices
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K093703

Device Name: 3Di

Indications For Use:

3Di is a software package of PACS workstation for handling multimodality (CT, XA, MR, PET, SPECT & Ultrasound) images, which are using DICOM protocol. It includes volume rendering, Multi-planar reconstruction (MPR) and viewing of the inner and outer surfaces of organs as well as within their walls.

3Di is intended for use as an interactive tool for assisting professional Radiologists, Cardiologists and specialists to reach their own diagnosis, by providing tools of communication, clinics networking, WEB Serving, image viewing, image manipulation, 2D/3D image visualization, image processing, reporting and archiving. This product is not intended for use with or for diagnostic interpretation of Mammography images.

The 3Di indications for use are processing of Cardiac CT angiography studies, including coronaries analysis, cardiac functional assessment and CT colonoscopy.


Prescription Use AND/OR
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, ~~Office of Device Evaluation (ODE)~~

(OIVD)



(Division Sign-Off)
Division of Radiological Devices

Page 1 of 1

510(k) Number K093703