

VII. 510(k) Summary

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular §807.92, the following summary of information is provided:

A. Submitted by:

Sheila Bruschi
 Regulatory Affairs Associate
 NuVasive, Incorporated
 7475 Lusk Blvd.
 San Diego, California 92121
 Telephone: (858) 320-4515
 Fax: (858) 320-4615

FEB 19 2010**B. Device Name**

Trade or Proprietary Name:	<i>NuVasive HELIX Revolution ACP System</i>
Common or Usual Name:	Cervical Plate and Screw System
Classification Name:	Spinal Intervertebral Body Fixation Orthosis
Device Class:	Class II
Classification:	§888.3060
Product Code:	KWQ

C. Predicate Devices

The subject *HELIX Revolution ACP System* is substantially equivalent to the *Helix ACP System* currently distributed commercially in the U.S. by NuVasive.

D. Device Description

The NuVasive HELIX Revolution ACP System consists of a variety of plates and screws designed to provide stabilization as an adjunct to cervical fusion.

E. Intended Use

The NuVasive HELIX Revolution ACP System is intended for anterior screw fixation of the cervical spine. These implants have been designed to provide stabilization as an adjunct to cervical fusion. Indications for the use of this implant system include degenerative disc disease defined as neck pain of discogenic origin with the degeneration of the disc confirmed by history and radiographic studies, spondylolisthesis, trauma, spinal stenosis, deformity, tumor, pseudarthrosis or failed previous fusion.

WARNING: The NuVasive HELIX Revolution ACP System is not intended for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic or lumbar spine.

F. Comparison to Predicate Devices

The subject device has indications for use identical to those of its predicate, and employs the same principles of operation.

G. Summary of Non-Clinical Tests

Mechanical testing was presented.

H. Summary of Clinical Tests

(Not Applicable).



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

FEB 19 2010

NuVasive, Inc
% Ms. Sheila Bruschi
Regulatory Affairs Associate
7475 Lusk Boulevard
San Diego, California 92121

Re: K093804

Trade/Device Name: NuVasive Helix Revolution ACP System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal intervertebral body fixation orthosis
Regulatory Class: Class II
Product Code: KWQ
Dated: January 21, 2010
Received: January 22, 2010

Dear Ms. Bruschi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set


Page 2 – Ms. Sheila Bruschi

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucml15809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson
Director,
Division of Surgical, Orthopedic
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K093804

Device Name: HELIX Revolution Anterior Cervical Plating System

Indications For Use:

The NuVasive HELIX Revolution Anterior Cervical Plating System is intended for anterior screw fixation of the cervical spine. These implants have been designed to provide stabilization as an adjunct to cervical fusion. Indications for the use of this implant system include degenerative disc disease defined as neck pain of discogenic origin with the degeneration of the disc confirmed by history and radiographic studies, spondylolisthesis, trauma, spinal stenosis, deformity, tumor, pseudarthrosis or failed previous fusion.

WARNING: The NuVasive HELIX Revolution Anterior Cervical Plating System is not intended for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic or lumbar spine.


Prescription Use
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

Page 1 of 1

510(k) Number K093804