

Section 5
Premarket 510k Summary

K094086

Submitter Information: Augustine Biomedical & Design, LLC
6581 City West Parkway
Eden Prairie, MN 55344
952.465.3500

Contact: James D. Ecklein, Director RA/QA JUN - 3 2010

Date Prepared: November 20, 2009

Trade Name: Hot Dog Multifunction Controller
Model Number: WC52

Product Code: DWJ (21 CFR Part 870.5900)

Common Name: Thermal Regulating System

Predicate Device: Hot Dog Patient Warming System K052392
Hot Dog Patient Warming Mattress System K092807

Device Description: The Hot Dog Multifunction Controller consists of a temperature control unit that monitors and controls the temperature of patient warming blankets and mattresses.

Intended Use: The Hot Dog Patient Warming System is intended to prevent or treat hypothermia and to provide warmth to patients. The Hot Dog Patient Warming System should be used in circumstances in which patients may not maintain a state of normothermia.

The System is intended primarily for use in hospitals and surgical centers including without limitation operating, recovery and emergency rooms and on medical/surgical floors.

Technological Characteristics: A comparison between the new and predicate devices shows that the technological characteristics and indications for use are equivalent. The products have similar designs, materials, components and dimensions.

Section 5 Premarket 510k Summary

Non Clinical Data Bench testing was performed to demonstrate that the proposed controller is substantially equivalent to the predicate devices. Temperature characteristics and safety systems were compared and found to be comparable.

The controller is designed to meet the following performance standards:

IEC 60601-1 Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance, edition: 2

IEC 60601-1-2, Medical Electrical Equipment - Part 1-2: General Requirements for Safety - Collateral standard: Electromagnetic Compatibility - Requirements and Tests, edition: 2.1

IEC 60601-1-4:2000, Medical electrical equipment - Part 1: General requirements for safety - 4 - Collateral standard: Programmable electrical medical systems, edition 1.1.

Clinical Data Not required

Conclusion The Hot Dog Multifunction Controller was found to be equivalent to the predicate device in technological characteristics, safety and indications for use.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

Regulatory Technology Services LLC
c/o Mr. Mark Job
Responsible Third Party Official
1394 25th Street NW
Buffalo, MN 55313

JUN - 3 2010

Re: K094056
Augustine Biomedical + Design LLC Hotdog Multi-Functional Controller
Regulation Number: 21 CFR 870.5900
Regulation Name: Thermal Regulating System
Regulatory Class: II
Product Code: DWJ
Dated: May 17, 2010
Received: May 18, 2010

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2 – Mr. Mark Job

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

