510(k) SUBSTANTIAL EQUIVALENCE DETERMINATION DECISION SUMMARY ASSAY ONLY TEMPLATE

A. 510(k) Number:

k112075

B. Purpose for Submission:

The addition of Ampicillin to the VITEK ® 2 and VITEK®2 Compact Systems Antimicrobial Susceptibility Test (AST) Systems

C. Measurand

Ampicillin concentrations 0.5, 1, 4 and 8 $\mu g/mL$ (the MIC range is \leq 0.25- \geq 16 $\mu g/mL)$

D. Type of Test:

Quantitative growth based detection algorithm using optics light detection

E. Applicant:

bioMérieux, Inc.

F. Proprietary and Established Names:

Vitek[®]2 Streptococcus Ampicillin ($\leq 0.25 - \geq 16 \ \mu g/mL$)

G. Regulatory Information:

1. <u>Regulation section:</u>

866.1645 Short-Term Antimicrobial Susceptibility Test System

2. Classification:

II

3. <u>Product Code:</u>

LON System, Test, Automated, Antimicrobial Susceptibility, Short Incubation

4. Panel:

83 Microbiology

H. Intended Use:

1. Intended use(s):

VITEK[®] 2 *Streptococcus* Ampicillin is designed for antimicrobial susceptibility testing of Streptococcus species. VITEK[®] 2 *Streptococcus* Ampicillin is a quantitative test intended for use with the VITEK[®] 2 and the VITEK[®] 2 Compact Systems as a laboratory aid in the determination of *in vitro* susceptibility to antimicrobial agents. Ampicillin has been shown to be activity against the microorganisms listed below: Beta-hemolytic group *Streptococcus* species Viridans group *Streptococcus* species

The VITEK[®] 2 Antimicrobial Susceptibility Test (AST) is intended to be used with the VITEK[®] 2 System for the automated quantitative or qualitative susceptibility testing of isolated colonies for the most clinically significant aerobic gram-negative bacilli, *Staphylococcus spp.*, *Enterococcus spp.*, *Streptococcus* spp. and clinically significant yeast.

2. Indication(s) for use:

VITEK[®] 2 *Streptococcus* Ampicillin is designed for antimicrobial susceptibility testing of Streptococcus species. VITEK[®] 2 *Streptococcus* Ampicillin is a quantitative test intended for use with the VITEK[®] 2 and the VITEK[®] 2 Compact Systems as a laboratory aid in the determination of *in vitro* susceptibility to antimicrobial agents. Ampicillin has been shown to be activity against the microorganisms listed below: Beta-hemolytic group *Streptococcus* species Viridans group *Streptococcus* species

The VITEK[®] 2 Antimicrobial Susceptibility Test (AST) is intended to be used with the VITEK[®] 2 System for the automated quantitative or qualitative susceptibility testing of isolated colonies for the most clinically significant aerobic gram-negative bacilli, *Staphylococcus spp.*, *Enterococcus spp.*, *Streptococcus* spp. and clinically significant yeast.

3. <u>Special condition for use statement(s):</u>

Prescription Use Only

4. Special instrument Requirements:

VITEK[®] 2 and the VITEK[®] 2 Compact Systems

I. Device Description:

Each VITEK® 2 test card contains 64 micro-wells. A control well which contains only microbiological culture media is resident on all cards. The remaining wells contain premeasured portions of a specific antibiotic combined with culture media. The bacterial or yeast isolate to be tested is diluted to a standardized concentration with 0.45 - 0.5% saline before being used to rehydrate the antimicrobial medium within the card. The VITEK 2 System automatically fills, seals and places the card into the incubator/reader. The VITEK 2 Compact has a manual filling, sealing and loading operation. The VITEK 2 Systems monitor the growth of each well in the card over a defined period of time. At the completion of the incubation cycle, a report is generated that contains the MIC value along with the interpretive category result for each antibiotic contained on the card.

The VITEK 2 AST-ST Ampicillin has the following concentrations in the card: 0.5, 1, 4 and μ /mL (equivalent standard method concentration by efficacy in μ /mL). The MIC result range for the VITEK 2 card is $\leq 0.25 - \geq 16\mu$ /mL.

J. Substantial Equivalence Information:

1. Predicate device name(s)

VITEK® 2 Gram Positive Amoxicillin for Streptococcus pneumoniae

2. Predicate K number(s):

k063597

3. Comparison with predicate

Similarities									
Item	Predicate								
Intended Use	Determine antimicrobial susceptibility to antimicrobial agents	Same							
Test Card	VITEK® 2 card format with base broth	same							
Instrument	VITEK® 2 and VITEK ®2 Compact System	same							

	Differences	
Item	Device	Predicate
Antibiotic	Ampicillin	Amoxicillin
Reading algorithm	Unique for Ampicillin	Unique for amoxicillin
Test organism	Beta-hemolytic group Streptococcus species Viridans group Streptococcus species	S. pneumoniae

K. Standard/Guidance Document Referenced (if applicable):

Class II Special Controls Guidance Document: "Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA"

CLSI M7-A8 "Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically; Approved Standard"

CLSI M100-S19 "Performance Standards for Antimicrobial Susceptibility; Twenty-First Information Supplement"

L. Test Principle:

Each VITEK® 2 test card contains 64 micro-wells. A control well which contains only microbiological culture media is resident on all cards. The remaining wells contain premeasured portions of a specific antibiotic combined with culture media. The bacterial or yeast isolate to be tested is diluted to a standardized concentration with 0.45 - 0.5% saline before being used to rehydrate the antimicrobial medium within the card. The VITEK 2 System automatically fills, seals and places the card into the incubator/reader. The VITEK 2 Compact has a manual filling, sealing and loading operation. The VITEK 2 Systems monitor the growth of each well in the card over a defined period of time. At the completion of the incubation cycle, a report is generated that contains the MIC value along with the interpretive category result for each antibiotic contained on the card.

The VITEK 2 AST-ST Ampicillin has the following concentrations in the card:, 0.5, 1, 4, and 8µg/mL (equivalent standard method concentration by efficacy in µg/mL). The MIC result range for the VITEK 2 card is $\leq 0.25 - \geq 16\mu$ g/mL.

In addition to the automatic dilution, there is also a manual inoculation dilution procedure described in the package insert.

M. Performance Characteristics (if/when applicable):

- 1. <u>Analytical performance:</u>
 - a. Precision/Reproducibility:

Reproducibility was demonstrated using 10 isolates at 3 sites on 3 separate days in triplicates. The study included the Auto-dilution and the Manual dilution with the VITEK®2, and the Manual dilution with the VITEK®2 Compact. All results were >95% reproducible.

b. Linearity/assay reportable range:

Not Applicable

c. Traceability (controls, calibrators, or method):

The recommended QC isolates were tested on every test occasion with the reference method and the VITEK®2. The reference method QC results were in range for every day tested. The VITEK®2 was tested a sufficient number of times to demonstrate that the system can produce QC results in the recommended range.

Quality Control was performed during the studies using both the autodilution and the manual method of diluting the organisms. Results demonstrated that methods were comparable with the same mode.

Organism	Conc in	Auto-c	lilution	Manual dilution			
	µg/ml						
S. pneumoniae		Ref.	Test	Ref.	Test		
ATCC 49619	≤0.03	2		2			
Expected Range	0.06	114		111			
0.06- 0.25 µg/ml	0.12	66		66			
	0.25		182		179		

Quality Control Summary (VITEK®2, Auto and Manual dilution)

An additional QC study was performed with the VITEK®2 Compact, the secondary option, at three sites, with the following results.

Quality Control Summary (VITEK®2 Compact, Manual dilution)

Organism	Conc in ug/ml	Manual-dilution			
S. pneumoniae	P.8	Ref.	Test		
ATCC 49619	≤0.03				
Expected Range	0.06	61			
0.06- 0.25 µg/ml	0.12	13			
	0.25		74		

Inoculum density control was monitored using the DensiChek2 instrument. This was standardized weekly with all results recorded and in the expected range. Verification was performed during internal testing.

d. Detection limit:

Not Applicable

e. Analytical specificity:

Not Applicable

f. Assay cut-off:

Not Applicable

2. <u>Comparison studies:</u>

The reference method follows CLSI approved broth microdilution testing conditions:

- Medium: Mueller-Hinton broth supplemented with lysed blood
- Inoculum: Direct colony suspension
- Incubation : 35°C, ambient air, 20- 24 hours
- a. Method comparison with predicate device:

Clinical study was performed at four external sites using the VITEK2 *Streptococcus* Ampicillin and broth microdilution panels containing Ampicillin. A total of 1124 clinical isolates were tested at four external sites by auto inoculation. The no growth rate was 0.8% (9/1124) in the VITEK®2 AST-ST Ampicillin Test. Three hundred and eight (27.4%: 308/1124) were from stock isolates.

				()								
	total	EA	%EA	Eval EA	Eval	Eval	CA	%CA	#R	min	maj	vmj
				Total	EA	%EA						
Clinical	1115	1104	99.0	60	54	90.0	1078	96.7	14	37	0	0
Challenge	157	157	100	6	6	100	156	99.4	10	1	0	0
Combined	1272	1261	99.1	66	60	90.9	1234	97.0	24	38	0	0

Performance Summary Table (VITEK 2, Auto Dilution)

EA-Essential Agreement CA-Category Agreement R-resistant isolates

maj-major discrepancies **vmj**-very major discrepancies **min-** minor discrepancies Manual Dilution:

The challenge set of 157 organisms was also tested at one site using the manual method of inoculation for VITEK2 and VITEK2 Compact with the following performance.

Comparison Challenge Data - VITEK 2, and VITEK2 Compact (Manual dilution)

									/			
	total	EA	%EA	Eval EA	Eval	Eval	CA	%CA	# R	min	maj	vmj
				Total	EA	%EA						
VITEK2	157	156	99.4	6	6	100	156	99.4	10	1	0	0
VITEK2												
Compact	157	155	98.7	6	6	100	155	98.7	10	2	0	0

The performance of the optional VITEK®2 Compact was evaluated in the QC, challenge, and reproducibility studies.

b. Matrix comparison:

Not Applicable

- 3. <u>Clinical studies:</u>
 - a. Clinical sensitivity:

Not Applicable

b. Clinical specificity:

Not Applicable

- c. Other clinical supportive data (when a and b are not applicable):
- 4. Clinical cut-off:

Not Applicable

5. Expected values/Reference range

Beta *Streptococcus* \leq 0.25 (S), 0.5- \geq 16 (NS)

Viridans Group *Streptococcus* \leq 0.25 (S), 0.5-4(I), 8 - \geq 16(R)

N. Proposed Labeling:

The labeling is sufficient and satisfies the requirement of 21 CFR Part 809.10.

O. Conclusion:

The submitted information in this premarket notification is complete and supports a substantial equivalence decision.