

SPECIAL 510(k): Device Modification
OIVD Review Memorandum (Decision Making Document is Attached)

To: THE FILE

RE: DOCUMENT NUMBER: k121654

This 510(k) submission contains information/data on modifications made to the SUBMITTER'S own Class II, Class III or Class I devices requiring 510(k). The following items are present and acceptable for the ***i-SENS AUTO-CHEK Blood Glucose Monitoring System***:

1. The name and 510(k) number of the SUBMITTER'S previously cleared device.
i-SENS, Inc., COOL Blood Glucose Monitoring System (k103396)
2. Submitter's statement that the **INDICATION/INTENDED USE** of the modified device as described in its labeling **HAS NOT CHANGED** along with the proposed labeling which includes instructions for use, package labeling, and, if available, advertisements or promotional materials (labeling changes are permitted as long as they do not affect the intended use).
3. A description of the device **MODIFICATION(S)**, including clearly labeled diagrams, engineering drawings, photographs, user's and/or service manuals in sufficient detail to demonstrate that the **FUNDAMENTAL SCIENTIFIC TECHNOLOGY** of the modified device **has not changed**.
This change was for:
 - 1) The shape of meter's housing
 - 2) The electric connector pattern of the test strip
 - 3) The brand name of the meter, the test strip, the control solution and the blood glucose monitoring system.
4. **Comparison Information** (similarities and differences) to applicant's legally marketed predicate device including, labeling, intended use, physical characteristics, and cleaning and disinfection robustness study.
5. A **Design Control Activities Summary** which includes:
 - a) Identification of Risk Analysis method(s) used to assess the impact of the modification on the device and its components, and the results of the analysis
 - b) Based on the Risk Analysis, an identification of the verification and/or validation activities required, including methods or tests used and acceptance criteria to be applied
 - c) A declaration of conformity with design controls. The declaration of conformity should include:
 - i) A statement signed by the individual responsible, that, as required by the risk analysis, all verification and validation activities were performed by the designated individual(s) and the results demonstrated that the predetermined acceptance criteria were met, and
 - ii) A statement signed by the individual responsible, that the manufacturing facility is in conformance with design control procedure requirements as specified in 21 CFR 820.30 and the records are available for review.

6. A Truthful and Accurate Statement, a 510(k) Summary or Statement and the Indications for Use Enclosure (and Class III Summary for Class III devices).

The labeling for this modified subject device has been reviewed to verify that the indication/intended use for the device is unaffected by the modification. In addition, the submitter's description of the particular modification(s) and the comparative information between the modified and unmodified devices demonstrate that the fundamental scientific technology has not changed. The submitter has provided the design control information as specified in The New 510(k) Paradigm and on this basis, I recommend the device be determined substantially equivalent to the previously cleared (or their preamendment) device.

The i-SENS Auto-Chek Blood Glucose Monitoring System is intended for single patient home use. Disinfection efficacy studies were performed on the materials comprising the meter and lancing device by an outside commercial testing laboratory demonstrating complete inactivation of hepatitis B virus (HBV) with Clorox® Germicidal Wipes (EPA Reg. No: 67619-12). The sponsor has demonstrated that there is no change in performance or in the external materials of the meter and lancing device after 520 pre-cleaning/disinfection cycles (each pre-cleaning/disinfection cycle includes 1 pre-cleaning and 1 disinfection wipe, adding up to a total of 520 wipes) designed to simulate 5 years of device use. Labeling has been reviewed for adequate instructions for validated cleaning and disinfection procedures.