

**CENTER FOR DRUG EVALUATION AND
RESEARCH**

APPLICATION NUMBER:
202153Orig1s000

CHEMISTRY REVIEW(S)

Chemistry, Manufacturing and Controls (CMC) Labeling Memo – User Manual

NDA 202-153

Ruby-Fill[®] (Rubidium Rb 82 Generator)

Jubilant DraxImage, Inc.

by
Chemistry Reviewer: Anne Marie Russell, Ph.D.
Office of New Drug Products
Division of New Drug Products 2 (Branch VI)
for
Division of Medical Imaging Products (DMIP)

1. NDA 202-153
2. Labeling Memo – “The Ruby Rubidium Elution System User Manual”
3. REVIEW DATE: 20-Sept-2016
4. REVIEWER: Anne Marie Russell, Ph.D.
5. PREVIOUS DOCUMENTS:

| Document | Document Date (Panorama) |
|---|--------------------------|
| Chemistry Review #2 Complete Response (Anne Marie Russell Ph.D., CMC reviewer) | 20-SEP-2016 |

6. SUBMISSION(S) BEING REVIEWED:

| Document | Document Receipt Date | DARRTS SDN | Contents |
|-------------------|-----------------------|------------|--|
| Quality amendment | 12-Sep-2016 | email | Response to Information Request #3 – volume expiry (30L) |
| Quality amendment | 25-Sep-2016 | email | Response to User Manual Information Request |

7. NAME & ADDRESS OF APPLICANT:

| | |
|-----------------|--|
| Name: | Jubilant DraxImage |
| Address: | 16751 Trans-Canada Highway Kirkland, Quebec Canada H9H 4J4 |
| Representative: | Susan P. Spooner, Ph.D. INC Research, LLC, 4800 Falls of Neuse Road Suite 600 Raleigh, NC 27609 phone 919-745-2492 |
| Telephone: | (514) 630–7087 |

8. DRUG PRODUCT NAME/CODE/TYPE:

- a) Proprietary Name: Ruby-Fill[®] ¹
- b) Non-Proprietary Name: Rubidium Rb–82 Chloride for Injection

9. LEGAL BASIS FOR SUBMISSION: 505(b)(2).

The reference listed drug (RLD) is Cardiogen 82 (N019414)

10. PHARMACOL. CATEGORY: Cardiac Positron Emission Tomography

¹ The Ruby-fill ⁸²Rb generator is operated by the Ruby-Fill[®] Elution System (RbES) (b) (4)

11. DOSAGE FORM: Sterile solution for injection.
12. STRENGTH/POTENCY/PACKAGING: Variable strength eluent (mCi/mL) depending on generator release activity level, generator age, elution system operation mode and time between elutions. Column is loaded with (b) (4) $^{82}\text{SrCl}$ at calibration (adsorbed onto SnO_2). Dose is 10 – 60 mCi ^{82}Rb .
13. ROUTE OF ADMINISTRATION: IV
14. Rx/OTC DISPENSED: Rx OTC
15. SPOTS (SPECIAL PRODUCTS ON-LINE TRACKING SYSTEM):
 SPOTS product – Form Completed Not a SPOTS product
16. CHEMICAL NAME, STRUCTURAL FORMULA, MOLECULAR FORMULA, MOLECULAR WEIGHT:
- Chemical Name(s): Rubidium Rb–82 Chloride for Injection
IUPAC name: Rubidium Rb–82 Chloride for Injection
CAS Registry No. [132486-03-4]
Molecular Formula: $^{82}\text{RbCl}$
Molecular Weight: 117.5 daltons
17. RELATED/SUPPORTING DOCUMENTS:
- A. DMFs: none.
B. Other Documents: none.

Chemistry Memo

Ruby Rubidium Elution System User Manual

In CMC Review #2 (Complete Response), two open issues (D1 and D2 below) with the User Manual were still under negotiation with the Applicant at the time of document completion, so this follow-up memo reviews the User Manual and documents the outcome.

(b) (4)

7 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page

4

Evaluation: Acceptable

Overall Evaluation of the User Manual: Acceptable as revised.

E. Label - PI (Prescribing Information)

This CMC reviewer provided input to the following sections of the PI throughout labeling negotiations with the Applicant: Highlights, Section 2.4 Elution System, Section 2.6 ^{(b) (4)} Eluate Testing Protocol, 2.8 RUBY-FILL Dose Delivery Limit, Section 3 Dosage Forms and Strengths, Section 11 Description and Section 16 How Supplied and Storage/Handling. Negotiations with the Applicant were ongoing when this memo was finalized - see approval letter for the final version of the PI

Signatures:

Primary Reviewer

Anne Marie Russell, Ph.D. CMC reviewer, ONDP, Division II, Branch VI

Secondary Reviewer:

Danae Christodoulou, Ph.D. Acting Branch Chief, ONDP, Division II, Branch VI

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

/s/

ANNE M RUSSELL

09/28/2016

DANAE D CHRISTODOULOU

09/28/2016



NDA 202153 Resubmission

OPQ N202153 Integrated Quality Assessment

Review Date: 09/23/2016

| | |
|--------------------------------|---|
| Drug Name/Dosage Form | Rubyfill [®] Rubidium Rb 82 Generator/Intravenous Infusion |
| Strength | (b) (4) not to exceed a total of 60 mL |
| Route of Administration | Intravenous infusion |
| Rx/OTC Dispensed | Rx |
| Applicant | Jubilant DraxImage Inc. (JDI), 16751 Trans-Canada Highway, Kirkland, Quebec, Canada H9H 414 |
| US agent, if applicable | Susan P. Spooner, Ph.D., INC Research, LLC (4800 Falls of Neuse Road, Suite 600, Raleigh, NC 27609; phone 919-745-2492) |

Quality Review Data Sheet

1. LEGAL BASIS FOR SUBMISSION: 505(b)(2) – RLD is **Cardiogen-82 (NDA 19414)**
2. RELATED/SUPPORTING DOCUMENTS:
 - A. DMFs:

Table 1 Drug Master Files (DMFs)

| DMF # | TYPE | HOLDER | ITEM REFERENCED | STATUS ¹ | DATE REVIEW COMPLETED | REVIEWER |
|---------|------|---------|-----------------|---------------------|--------------------------|-------------------------|
| (b) (4) | II | (b) (4) | (b) (4) | 3 | 01/17/2012 (adequate) | Milagros Salazar, Ph.D. |
| | II | | | 3 | 01/18/2012 (adequate) | Milagros Salazar, Ph.D. |

¹The DMF

²Adequate, Adequate with Information Request, Deficient, or N/A (There is enough data in the application, therefore the DMF did not need to be reviewed)

³Reviewed previously and no revision since last review

- B. Other Documents: *IND, RLD, or sister applications*
N/A

3. CONSULTS:

| DISCIPLINE | RECOMMENDATION | DATE | REVIEWER |
|------------|----------------|------|------------------|
| CDRH | | | Robert Meyer, MS |

Quality Review Team

| DISCIPLINE | REVIEWER | BRANCH/DIVISION |
|-------------------------------|-------------------------------|--------------------------------|
| Drug Substance | Anne Marie Russell, Ph.D. | ONDP/Branch VI/Division II |
| Drug Product | Anne Marie Russell, Ph.D. | ONDP/Branch VI/Division II |
| Process | Anne Marie Russell, Ph.D. | ONDP/Branch VI/Division II |
| Microbiology | Yeissa ChabrierRosello, Ph.D. | OPQ/OPF/Microbiology |
| Facility | Michael Klupal | OPQ/OPF/DBP/BI |
| Biopharmaceuticals | N/A | N/A |
| Project Manager (R.Ph.) | Thao Vu, R.Ph. | OMPT/CDE/OPQ/OPRO/DP MI/RBPMBI |
| Application Technical Lead | Eldon E. Leutzinger, Ph.D. | ONDP/Branch VI/Division II |
| Laboratory (OTR) | N/A | N/A |
| ORA Lead | N/A | N/A |
| Environmental Assessment (EA) | N/A | N/A |

Table 2 Documents Reviewed

| DOCUMENT | RECEIPT DATE | DESCRIPTION | Section/reviewer |
|--|--|--|--|
| Complete Response (Resubmission-Class 2) | 12/30/2015 | Submission in response to Complete Response Letter | Anne Marie Russell, Ph.D., ONDP/Branch VII/Division II |
| Quality amendments | 06/01/2016 06/29/2016 08/30/2016 09/12/2016 | Response to IR's | Anne Marie Russell, Ph.D., ONDP/Branch VII/Division II |

Executive Summary

I. Recommendations

A. Recommendation and Conclusion on Approvability

NDA 202153 for Rubyfill is recommended for approval, from the standpoint of Chemistry, Manufacturing and Controls, pending conclusions by the CDRH review (not yet final as of the date of this Integrated Executive Summary). Both Microbiology and Manufacturing Facilities are recommending approval.

1. Summary of Complete Response issues & Subsequent IR's, Facilities

| ISSUE | STATUS |
|---|--|
| CMC From Complete Clinical Response Letter (12/18/2016) | Resolved – clinical use simulation found acceptable. Post-approval testing protocol (b) (4) acceptable. Clarifications of Elution System instructions acceptable – response of 06/01/2016 |
| CMC | Resolved – (b) (4) post- |

| | | |
|--|---------|--|
| (Continue from Complete Response Issues) | (b) (4) | approval stability protocol resolved Resolved – (b) (4) stability testing – response of 06/01/2016 |
|--|---------|--|

| ISSUE | | STATUS |
|---|---|--|
| CMC (uncertainty in 10 – 60 mCi dose) | CMC IR of 05/16/2016 – (1) assessment of uncertainty in the dose (10 – 60 mCi) at the maximum and (2) minimum range of the system operation (flow rates, ⁸² Rb concentration, elution volume). | Resolved - issues (1) and (2) – response of 06/01/2016 |
| CMC (new low dose limit of 10 mCi) | CMC IR of 5/16/2016 – (1) capability of dose calibrator to detect new alert limits (0.004 µCi ⁸² Sr/mCi ⁸² Rb, 0.04 µCi ⁸⁵ Sr/mCi ⁸² Rb) in the new dose of 10 mCi, (2) DL for strontium and assessment of uncertainty of measurement at lowest level (~ 0.01 µCi), (3) calculations to determine the reported capability, (4) study reports (data, analysis) referenced in document “RUBY-FILLRubidium Rb 82 Generators. | Resolved – issues (1), (2), (3) and (4) – response of 06/01/2016 |
| CMC (System Performance; Capability of Delivering Patient Dose) | CMC IR of 5/16/2016 – (1) assessment of delivered dose volume, strength and rate of delivery over lifetime of generator (release, mid-life, expiry) for minimum (10 mCi) and maximum (60 mCi). (2) assessment of uncertainty of values in (1) basis of assessment. (3) Rationale for (b) (4) delivery rate proposed in label | Resolved - issues (1) and (2) – response of 06/01/2016. Resolved – issue (3) - response of 09/12/2016 |
| Microbiology | Lack of dye ingress validation (1) testing (b) (4) and (2) the limitations of the dye ingress test to simulate possible microbial ingress into the system | Resolved – issues (1) and (2) – responses of 8/17/2016 |

| ISSUE | | STATUS |
|------------|--|--|
| Biopharm | N/A | N/A |
| Facilities | ⁸² Sr – manufacture (b) (4) & established manufacturer of ⁸² Sr. Drug Product (<i>Jubilant Draximage</i> – drug manufacturer, <i>Jubilant</i> <i>Hollisterstier</i> – release & stability tester, (b) (4) – release & stability tester. (u) (4) – release & stability tester | Resolved (b) (4) – Corrections involving validation of test methods for ⁸² Sr completed (b) (4) acceptable by profile Facilities in manufacture of drug product found acceptable on basis of profile and inspectional history |
| CDRH | Review of the Dose Delivery System involves several aspects, ranging from software to the physical system (b) (4) From the standpoint of those issues involving the physical system, the only mechanical issue that remained after review in CDRH (b) (4) | Conclusions on this issue, software and any other issues are pending as of the date of this integrated executive summary |

B. Recommendation on Phase 4 (Post-Marketing) Commitments, Agreements, and/or Risk Management Steps, if Approvable

None

II. Summary of Quality Assessments

INTRODUCTION:

The product from Rubyfill (Rubidium Rb 82 Generator) is ⁸²RbCl in saline, without excipients. It is produced in a “radionuclide generator” with the long/short-lived radionuclide pair (⁸²Sr/⁸²Rb). Due to the short physical half-life of ⁸²Rb (75 seconds), ⁸²RbCl is administered directly to a patient through an infusion delivery system connected to the generator. Based on the governing principles of a radionuclide generator, ⁸²Sr²⁺ is expected to remain stationary on a short chromatography column of hydrous stannic oxide, allowing for ⁸²Rb⁺ to elute from the column, thus effecting separation of ⁸²Rb⁺ from ⁸²Sr²⁺ (parent radionuclide). Hydrous stannic oxide is the stationary phase (“matrix”), whereas 0.9% Sodium Chloride (USP) is the mobile phase.

A. Drug Substance [USAN Name] Quality Summary

Chemically, the drug substance is ⁸²RbCl and the USAN is Rubidium Chloride Rb 82. Rubidium (Rb) is Element 37 belonging to Group 1 of the Periodic Table, commonly referred to as the alkali metals, with electronic configuration of [Kr]5s¹. By virtue of the very large size of the 2nd ionization potential, Rb ion exists solely in the +1 oxidation state, and the chemistry of Rb and all

its isotopes is that of Rb⁺. Rubidium possesses 32 isotopes, of which only 2 are naturally occurring (⁸⁵Rb, 72.2% natural abundance; ⁸⁷Rb, 27.8% natural abundance and radioactive with long physical half-life of 4.9 x 10¹⁰ years). The remaining isotopes, including ⁸²Rb, are radioactive and are not found in nature. The product of the decay of ⁸²Rb is stable Kr [⁸²⁽³⁷⁺⁴⁵⁾₃₇Rb → ⁸²⁽³⁶⁺⁴⁶⁾₃₆Kr + β⁺ + ν], in which a proton is converted to a neutron [p⁺ → n + β⁺ + ν (neutrino)], resulting in a change in Z from 37 to 36. In this process, two particles (β⁺ and ν) carry away the energy of the nuclear transition, and the energy spectrum of the positrons is a continuous distribution, as opposed to a emission of a discrete energy peak. Once β⁺ particles are formed, they have a finite, but very short lifetime. On collision with electrons, the β⁺ particles annihilate forming two 511 KeV γ-rays at approximately 180° apart, the basis of PET imaging with ⁸²Rb.

⁸²Rb is obtained from ⁸²Sr, and the overall process characterizing the nuclear transformations is as follows: ⁸²⁽³⁸⁺⁴⁴⁾₃₈Sr → ⁸²⁽³⁷⁺⁴⁵⁾₃₇Rb → ⁸²⁽³⁶⁺⁴⁶⁾₃₆Kr + β⁺ + ν. ⁸²Sr (absorbed as ⁸²Sr²⁺ to the column matrix) decays by orbital electron capture (EC) in which the **nucleus absorbs one of the atom's orbital electrons, reacting with a proton, neutralizing it with formation of a neutron and a neutrino** [e⁻ + p⁺ → n + ν (neutrino)]. Overall, there is a change in Z from 38 to 36. This is seen in the Periodic Table, with ⁸²Sr going from Group 2 to ⁸²Rb in Group 1, then wrapping around (left-wise) in the Periodic Table to stable ⁸²Kr of the Inert Gasses (Group 18).

The conversion of ⁸²Sr to ⁸²Rb occurs on the generator column [⁸²Sr → ⁸²Rb + ν], since the ⁸²Sr (as ⁸²Sr²⁺) stays put (in principle), although some relatively small amounts of ⁸²Sr²⁺ leaks out, by virtue of the imperfect chemistry of absorption to stannic oxide matrix.

QUALITY SUMMARY – the radionuclidic identity of the drug substance (due to ⁸²Rb) is defined by well-established physics (⁸²Sr → ⁸²Rb + ν), and radiochemical identity (⁸²RbCl) by the exchange process that occurs on the generator column matrix (b) (4) releasing ⁸²Rb⁺ (with Cl⁻) with elution by saline. There are no radionuclide impurities arising from the nuclear transformation itself. The only issue pertinent to the quality of the ⁸²RbCl is that of (b) (4)

B. Drug Product [Established Name] Quality Summary

Rubyfill [Rubidium Rb 82 Generator]. The generator eluate (containing ⁸²RbCl) is administered directly to patients, and has a stand-alone indication (*PET imaging of the myocardium under rest or pharmacologic stress conditions to evaluate regional myocardial perfusion in adult patients with suspected or existing coronary artery disease*). Hence, **by 21 CFR 310.3(n), Rubyfill (Rubidium Rb 82 Generator) is considered a drug, and furthermore defined as a PET drug and regulated under 21 CFR 212**. All of the CMC information pertaining to the generator, its manufacture and controls is in the NDA.

Rubyfill (Rubidium Rb 82 Generator) is a radionuclide generator that contains (b) (4) ⁸²SrCl₂ adsorbed onto hydrous (b) (4) stannic oxide in a column (b) (4). The System front view and Schematic showing the internal network of functional parts is reproduced from the NDA, as follows:

8 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page

Dose of ⁸²RbCl to Patient

(b) (4) JDI is proposing to extend the dose range to 10 – 60 mCi, introducing a **new low dose limit of 10 mCi**.

➤ **New Low Dose (10 mCi)**

Several issues arose regarding this new dose limit, namely (IR, 5/13/2016) affecting risk to the patient, namely (a) uncertainty of the dose delivered, and (b) uncertainty in the detectability of strontium breakthrough.

ISSUE – **New Low Dose #1 (5/13/2016)** – IR #1 - **assessment of uncertainty in the in dose range (10 – 60 mCi) – RESOLVED.** (b) (4)

ISSUE – **New Low Dose #2 (5/13/2016)** – IR #1 - **limit of detection – RESOLVED.** The limit of detection of the dose calibrator is determined (b) (4)

Summaries of Rubyfill elution system performance testing is provided.

➤ **System Performance**

The clinical team (DMIP) had requested an explanation of the slower maximum infusion rate (30 mL/min). That for Cardiogen-82 is 50 mL/min.

ISSUE – **System Performance** (5/13/2016) – IR #2 – absence of demonstrated capability of the Rubyfill system to deliver minimum and maximum dose (10 mCi and 60 mCi) at the beginning of generator life to its expiration, test method and controls - **RESOLVED**. In a

tcon with JDI (7/13/2016), additional data was requested. Data was provided by Draximage (b) (4)



ISSUE – **System Performance** (9/06/2016) – IR #3 – absence of data use to support proposed Rubyfill labeled elution volume of 30 L at expiry – **RESOLVED**. Data provided (9/12/2016)

are not primary stability data, because of differences between test conditions and commercial operation, and between development and commercial generators. As a consequence, these data are of a secondary nature. Yet, it is pertinent that at the 30 L expiry point, the release acceptance criteria were not exceeded. Some data from Canadian generators (b) (4)

(b) (4) were provided. These generators were operated to simulate clinical use, although not using the commercial elution system. The impact of differences do not adversely affect the breakthrough performance (shown by JDI). The data from the Canadian generators is primary and supportive of the 30 L expiry. The sum total of all the data assures that breakthrough performance of the commercial generators will be met at 30 L expiry. **In light of these considerations, it is to be noted that the stability data provided does not include data from 3 commercial generators operated with the commercial elution system and with commercial (b) (4). Hence, the post-approval stability program will be very important in confirming the expected performance of the commercial product.**

Microbiology

After the initial review (06/03/2016) and responses to address multiple deficiencies, the issues that remained were (1) lack of dye ingress validation testing (b) (4) and (2) the limitations of the dye ingress test to simulate possible microbial ingress into the system. Based on the information to address the issues in (1), their results demonstrate that the risk of cross-contamination (b) (4) is well controlled by the safeguards put in place (b) (4). Regards (2), the firm's response is twofold. (b) (4)



These assessments and rationale were deemed to be acceptable by the microbiologist

reviewer (Yeissa ChabrierRossello, Ph.D.) for a final determination that all microbiology deficiencies identified in the application are resolved.

Dose Delivery System

Review of the Dose Delivery System involves several aspects, ranging from software to the physical system (b) (4). From the standpoint of those issues involving the physical system, the only mechanical issue that remained after review in CDRH (Robert Meyer, M.S.) (b) (4). The final review conclusion on this issue, including an assessment by DMIP, and of the software and any other issues is pending as of the date of this integrated executive summary.

Rubyfill User Manual

(b) (4)

(b) (4)

These issues relating to improvements in the User Manual are currently under negotiation with JDI.

Labeling

Some final labeling changes (Michele Fedowitz, M.D., DMIP; Anne Marie Russell, Ph.D., ONDP) are proposed to the Outer Label (main label, assay label) and will be communicated to JDI. There are no other outstanding labeling issues, other than the continuing negotiations with JDI on the User Manual.

C. Summary of Drug Product Intended Use

| | |
|---|---|
| Proprietary Name of the Drug Product | Rubyfill |
| Non Proprietary Name of the Drug Product | Rubidium Rb 82 Generator |
| Non Proprietary Name of the Drug Substance | Rubidium Chloride Rb 82 (USAN) |
| Proposed Indication(s) including Intended Patient Population | Imaging of the myocardium under rest or pharmacologic stress in patients with suspected or existing coronary artery disease |
| Duration of Treatment | N/A |
| Maximum Daily Dose | 60 mCi |
| Alternative Methods of Administration | N/A |

D. Biopharmaceutics Considerations

N/A

E. Novel Approaches

N/A

F. Any Special Product Quality Labeling Recommendations

None

G. Process/Facility Quality Summary (see Attachment A)

See I.A. Recommendations and Conclusion on Approvability (Summary of Complete Response Issues & Subsequent IT's, Facilities)

H. Life Cycle Knowledge Information (see Attachment B)

N/A

Risk Assessment - Drug Product (Rubidium Rb 82 Generator)

| From Initial Risk Identification | | | Review Assessment | | |
|----------------------------------|------------------------------------|-----------------------------|-----------------------------|---------------------------------------|--|
| Attribute/ CQA | Factors that can impact the CQA | Initial Risk Ranking* | Risk Mitigation Approach | Final Risk Evaluation ⁶ | Lifecycle Considerations/ Comments** |
| Radionuclidic Identity/purity | | (b) (4) | N/A | (b) (4) | N/A |
| Radiochemical identity | | N/A | N/A | | |
| Radiochemical purity | | N/A | N/A | | |
| Chemical Purity | | N/A | N/A | | |
| Strength (mCi/mL) | | Data provided to address | N/A | | |
| pH | | N/A | N/A | | |
| Stability | | N/A | N/A | | |

| | | | | | |
|--------------|--|---------|-------------------------------------|---------|-----|
| Microbiology | | (b) (4) | Information to resolve deficiencies | (b) (4) | N/A |
|--------------|--|---------|-------------------------------------|---------|-----|

1. Radionuclidic Identity/Purity – sources of ⁸²Sr (b) (4) previously reviewed under DMF's, and determined to be acceptable for use in the rubidium generator.
2. Radiochemical Identity/Purity – established in (b) (4) DMF's.
3. Chemical Purity – Trace Metals – established in (b) (4) DMF's.
4. Microbiology – see Microbiology Review (Yeissa ChabrierRosello, Ph.D.); RPN (after modification when applicable) x S x D.
5. Overall Risk Assessment, (b) (4) (low, based on resolution of all issues for CMC & Microbiology).

Application Technical Lead: Eldon E. Leutzinger, Ph.D., CMC Lead

**Chemistry, Manufacturing and Controls (CMC)
Review of Complete Response
Drug Product**

NDA 202-153

Ruby-Fill[®] (Rubidium Rb 82 Generator)

Jubilant DraxImage, Inc.

**by
Chemistry Reviewer: Anne Marie Russell, Ph.D.
Office of New Drug Products
Division of New Drug Products 2 (Branch VI)
for
Division of Medical Imaging Products (DMIP)**

Table of Contents

| | |
|--|-----------|
| Chemistry Review Data Sheet..... | 3 |
| Chemistry Review for NDA 202-153 Complete Response | 6 |
| I. Recommendations..... | 6 |
| II. Review of Applicant’s Response to CMC Deficiencies: | 6 |
| III. Additional issues which arose during review: | 12 |
| IV. Labeling:..... | 38 |
| V. Overall recommendation:..... | 38 |
| VI. Signatures:..... | 38 |

Chemistry Review Data Sheet

1. NDA 202-153
2. REVIEW #2 (Complete Response)
3. REVIEW DATE: 11-MAY-2016
4. REVIEWER: Anne Marie Russell, Ph.D.
5. PREVIOUS DOCUMENTS:

| Document | Document Date (DARRTS) |
|---|------------------------|
| Chemistry Review #1 (David Place Ph.D., CMC reviewer) | 19-SEP-2014 |
| CMC memo #1 (David Place Ph.D., CMC reviewer) | 11-Dec-2014 |
| CMC memo #2 (David Place Ph.D., CMC reviewer) | 11-Dec-2014 |
| FDA action letter (Complete Response (CR)) | 18-Dec-2014 |

History of the application: This is the second review cycle. N202-153 was originally filed in 2006 as an ANDA and was converted to NDA (505b2) due to clinical differences from the Reference Listed Drug (Cardiogen N019414). The original NDA 505b2 application was not approved. See CR letter, issued 18-Dec-2014.

Review Clock: The original PDUFA date for this Complete Response was 30-Jun-2016. On 29-Jun-2016, the review clock was extended to 30-Sep-2016 due to receipt of a major amendment (CDRH).

6. SUBMISSION(S) BEING REVIEWED:

| Document | Document Receipt Date | DARRTS SDN | Contents |
|--|-----------------------|------------|---|
| Complete Response (Resubmission Class 2) | 30-Dec-2015 | 30 | Complete response to CR letter. |
| Quality Amendment | 01-Jun-2016 | 36 | Response to Information Request #1 |
| Quality Amendment | 29-Jun-2016 | 43 | Response to Information Request #2 |
| Quality Amendment | 30-Aug-2016 | 47 | Response to Information Request #2 – clinical simulation test |
| Quality amendment | 12-Sep-2016 | email | Response to Information Request #3 – volume expiry (30L) |

7. NAME & ADDRESS OF APPLICANT:

| | |
|-----------------|--|
| Name: | Jubilant DraxImage |
| Address: | 16751 Trans-Canada Highway Kirkland, Quebec Canada H9H 4J4 |
| Representative: | Susan P. Spooner, Ph.D. INC Research, LLC, 4800 Falls of Neuse Road Suite 600 Raleigh, NC 27609 phone 919-745-2492 |
| Telephone: | (514) 630-7087 |

8. DRUG PRODUCT NAME/CODE/TYPE:

- a) Proprietary Name: Ruby-Fill^{® 1}
 b) Non-Proprietary Name: Rubidium Rb-82 Chloride for Injection

9. LEGAL BASIS FOR SUBMISSION: 505(b)(2).

The reference listed drug (RLD) is Cardiogen 82 (N019414)

10. PHARMACOL. CATEGORY: Cardiac Positron Emission Tomography

11. DOSAGE FORM: Sterile solution for injection.

12. STRENGTH/POTENCY/PACKAGING: Variable strength eluent (mCi/mL) depending on generator release activity level, generator age, elution system operation mode and time between elutions. Column is loaded with (b) (4) ⁸²SrCl at calibration (adsorbed onto SnO₂). Dose is 10 – 60 mCi ⁸²Rb.

13. ROUTE OF ADMINISTRATION: IV

14. Rx/OTC DISPENSED: Rx OTC15. SPOTS (SPECIAL PRODUCTS ON-LINE TRACKING SYSTEM):

SPOTS product – Form Completed Not a SPOTS product

16. CHEMICAL NAME, STRUCTURAL FORMULA, MOLECULAR FORMULA, MOLECULAR WEIGHT:

Chemical Name(s): Rubidium Rb-82 Chloride for Injection

IUPAC name: Rubidium Rb-82 Chloride for Injection

CAS Registry No. [132486-03-4]

Molecular Formula: ⁸²RbCl

Molecular Weight: 117.5 daltons

¹ The Ruby-fill ⁸²Rb generator is operated by the Ruby-Fill[®] Elution System (RbES) (b) (4)

17. RELATED/SUPPORTING DOCUMENTS:

- A. **DMFs:** N/A. DMFs have been reviewed in previous review cycles. No new DMFs in this submission.
- B. **Other Documents:** none.

18. STATUS:

| CONSULTS/ CMC RELATED REVIEWS | RECOMMENDATION | DATE | REVIEWER |
|---------------------------------|----------------|--------------|--------------------------------|
| Facility Inspection | NA | | |
| Pharm/Tox | NA | | |
| Biopharm | NA | | |
| Labeling Nomenclature Committee | NA | | |
| Methods Validation | NA | | |
| DMEPA | | | |
| Environmental Assessment | NA | | |
| Microbiology | acceptable | 15-Sept-2016 | Yeissa Chabrier-Roselló, Ph.D. |
| CDRH | pending | | Robert Meyer, M.S. |

Chemistry Review for NDA 202-153 Complete Response

I. Recommendations

- A. Recommendation and Conclusion on Approvability:
From a Chemistry, Manufacturing and Controls standpoint, this New Drug Application is recommended for approval, pending acceptable findings by the CDRH review, which is not yet final as of this writing. The proposed volume expiry (30L) and time expiry (60 days) for the generator is granted, when operated using the commercial elution system (b) (4)
- B. Recommendation on Phase 4 (Post-Marketing) Commitments, Agreements, and/or Risk Management Steps, if Approvable: None

II. Review of Applicant's Response to CMC Deficiencies:

Two CMC deficiencies were identified in the 18-Dec-2014 Complete Response letter to the Applicant under the heading PRODUCT QUALITY items #3 and #4.

The format of this review is as follows: the deficiency from the Complete Response letter is in normal font, *the Applicant's response provided in this submission is in italics* and **the reviewer's evaluation is in bold font**.

Deficiency #3

(b) (4)

Applicant's response (received 30-Dec-2015):

In this Complete Response submission, the Applicant provided a revised post-approval stability protocol (b) (4)

III. Additional issues which arose during review:

Background: This reviewer compiled Table A which compares Cardiogen generator (RLD) with the Ruby-fill proposed commercial generator:

| Table A Rubidium-82 generator: Comparison to RLD | | |
|--|---|--|
| Characteristic | Cardiogen generator (RLD) | Ruby-Fill generator |
| Dose | 30 – 60 mCi Rb 82 | 10 – 60 mCi Rb 82 |
| Flow rate | 50 mL/min | 15 – 30 mL/min |
| Elution volume per dose | < 100 mL | < 60 mL |
| Elution time per dose | Not specified | (b) (4) |
| Radioactivity delivered by generator* | | |
| • at calibration (Day 0/ 1.00) | 90 – 150 mCi Rb 82 | 85 – 115 mCi Rb 82 |
| • at first clinical use (Day 11/ 0.737) | (b) (4) mCi Rb 82 | (b) (4) mCi Rb 82 |
| • at end of expiry [^] | (b) (4) mCi Rb 82 ([^] Day 42/ 0.312) | (b) (4) mCi Rb 82 ([^] Day 60/ 0.189) |
| How dose is delivered | User manually operates Cardiogen system control panel | User input to Ruby-fill system software interface, software controls infusion. |
| Operation modes | (b) (4) | |
| Expiry - time | 42 days | 60 days |
| Expiry - volume | 17L of eluent | 30 L of eluent |
| Calibration dose | N/A | 35mL at 20 mL/min |

*Calculated from fraction remaining in Table 6 Cardiogen 82 and Ruby-fill labels

A. New low dose (10mCi):

During the review cycle, Jubilant DraxImage (JDI) proposed to extend the dose range (b) (4) to 10 – 60 mCi, which introduced a new low dose limit (10mCi). The review issues from a CMC standpoint for the new low dose are two fold – the risk to the patient regarding uncertainty in the dose administered and the detectability of strontium breakthrough.

Dose uncertainty: Any measurement has an inherent uncertainty due to the limitation of the equipment and the measuring method. (b) (4)

(b) (4)
The Complete Response submission and the original NDA submission did not provide an assessment of the uncertainty of these radioactivity measurements – see comment below.

Strontium breakthrough detectability: The levels of strontium (⁸²Sr and ⁸⁵Sr) in the patient dose are monitored daily in the calibration dose eluted during the Daily Quality Control (Section 7 in the User Manual). The eluent is allowed to decay out the ⁸²Rb and the residual radioactivity is measured by the dose calibrator. The system then calculates the levels of strontium using the measured residual radioactivity, equations and limits specified in the Ruby-fill label. It is unknown if the strontium breakthrough alert limit (⁸²Sr 0.004 per μCi/mCi of ⁸²Rb,

^{85}Sr 0.04 per $\mu\text{Ci}/\text{mCi}$ of ^{82}Rb) in the new low dose (10 mCi) is below the detectability limit of the system. An evaluation of the Limit of Detection (LOD) of strontium breakthrough in the system is needed - see comment below.

The following three comments were sent to the Applicant on 13-May-2016 in CMC Information Request #1, response received 1-Jun-2016:

1. Provide an assessment of the uncertainty in the dose (10 – 60 mCi) administered to the patient at the maximum and minimum range of the system operation (e.g. flow rates, ^{82}Rb concentration, elution volume/time). Explain the basis for each assessment. Tabulate the data where possible.

Response: Applicant provided dose error data from Clinical Simulation Test 4 (Protocol ##3000069-P/ Appendix 8-2 Clinical Simulation Study Report).

[Redacted text block]

[Redacted text block]

| Dose Range | Flow-Rate | Average Dose Error |
|--------------------------|-----------|--------------------|
| [Redacted table content] | | |

Table 6

(b) (4)



Review:

(b) (4)



(b) (4)

| Table B | | | | |
|---|--------------------|----------------|-----|-----|
| Dose Accuracy in Clinical Simulation Test 4 | | | | |
| Dose (mCi) | Flow rate (mL/min) | Dose Error (%) | | |
| | | avg | min | max |
| (b) (4) | | | | |

(b) (4)



(b) (4)

As per Michele Fedowitz, M.D., clinical reviewer, (via email on 8-June-2016) the dose error is not clinically significant and is acceptable.

Evaluation: Acceptable

2. Limit of detection for strontium in new low dose (10 mCi):

- a. Discuss the capability of the dose calibrator to detect strontium at alert levels (^{82}Sr 0.004 per $\mu\text{Ci}/\text{mCi}$, ^{85}Sr 0.04 per $\mu\text{Ci}/\text{mCi}$) in the new low dose of 10mCi. Include an assessment of the limit of detection (LOD) for strontium using the supplied dose calibrator unit and the uncertainty in that dose calibrator measurement [REDACTED] (b) (4). Provide calculations used to determine the reported capability.**

Response: The Applicant explained that breakthrough levels are reported as the amount of strontium per the amount rubidium (e.g. ^{82}Sr 0.004 $\mu\text{Ci}/\text{mCi}$ ^{82}Rb and ^{85}Sr 0.04 $\mu\text{Ci}/\text{mCi}$ ^{82}Rb) and are assessed during the daily calibration of the system. [REDACTED] (b) (4)

Per the recently submitted report (RES.RBY.SDY.033 Ruby-Fill Elution System (RbES) Breakthrough Testing), the Limit of Detection (LOD) of the dose calibrator is experimentally determined [REDACTED] (b) (4)

Evaluation: Acceptable. [REDACTED]

(b) (4)

[REDACTED] The alert limit is essentially at or below the limit of detection of the dose calibrator when the generator is at expiry, but not earlier. This is acceptable because of the extensive margin built into the alert limits to assure patient safety.

- b. Provide the study reports (data, analysis) referenced in the document “RUBY-FILL® Rubidium Rb 82 Generators - Evaluation of Strontium Isotope Breakthrough” (Appendix 3-1) which summarized test results for:
- i. RES.RBY.SDY.034 Volume Limits and Strontium Breakthrough
 - ii. RES.RBY.SDY.033 Ruby-Fill Elution System (RbES) Breakthrough Testing
 - iii. RES.RBY.SDY.042 Interim Report: Summary of RbES [REDACTED] Performance Testing
 - iv. RES.RBY.SDY.054 REPORT: Summary of Ruby-Fill Elution System Performance Testing

- v. RES.RBY.SDY.070 REPORT: Summary of Ruby-Fill Elution System Performance Testing-Low Usage
- vi. RES.RBY.SDY.072 REPORT: Summary of Ruby-Fill Elution System Performance Testing-Extreme Usage

Response: Reports provided.

Evaluation: Acceptable. See Maximum volume at expiry (30L) in Section C below.

B. System performance/ Flow rate of eluent:

During label review, the clinical team requested an explanation of the slower maximum infusion rate recommended for Ruby-fill (30 mL/min) compared to the reference listed drug Cardiogen (50 mL/min). Additional data regarding infusion times and full system performance are needed to evaluate the clinical impact.

The following comment was sent on 13-May-2016 in CMC Information Request #1, response received 1-Jun-2016.

- 3. **For the maximum (60mCi) and minimum (10mCi) dose range, provide an assessment of the delivered volume, concentration (mCi/mL) and rate of delivery (mL/min) for the lifetime of the generator (release, mid-life and expiry). Include an assessment of the uncertainty in the values and explain the basis for each assessment. Describe the rationale for the 30 mL/min maximum rate proposed in the draft label. Tabulate the information, where possible.**

Response: Data provided for minimum, intermediate and maximum dose at generator release, mid-point and expiry, in Tables 11, 12 and 13 below.

| Time Point | Dose (mCi) | Volume (mL) | Flow-Rate (mL/min) | Elution Time (sec) |
|------------|------------|-------------|--------------------|--------------------|
| (b) (4) | | | | |



(b) (4)

The maximum flow of the Ruby-Fill system is 30 mL/min

(b) (4)

Initial Evaluation: Not Acceptable. (b) (4) Data were not provided for a minimum dose (10 mCi) delivered from a generator at release and for a maximum dose (60mCi) delivered from a generator at mid-point and expiry. These conditions represent the limits of the system. (b) (4)



(b) (4)

The following deficiency was sent to the Applicant on 20-Jun-2016 in CMC Information Request #3:

CMC Deficiency:

(b) (4)

Information needed to resolve the deficiency:

1. Provide data to demonstrate that the commercial system is capable of delivering the maximum and minimum dose (10mCi and 60mCi) at the beginning and end of generator lifetime within the proposed operating range of the system (b) (4)

Include raw data, averages, n value, uncertainty and range of values (minimum and maximum).

First Response: received 29-Jun-2016 (SDN #43):

No new clinical simulation test data provided.

(b) (4)

(b) (4)




Initial evaluation: Not acceptable. In this response, JDI summarized previously submitted clinical simulation data, which had already been reviewed and found to be inadequate because data at system limits were missing. In the 13-Jul-2016 tcon, FDA advised that additional data were needed and JDI agreed to conduct additional testing and submit.

(b) (4)



Second Response: received 30-Aug-2016 (SDN #47) Clinical Simulation Study Report – Dosing Evaluation. The submitted Executive Summary is copied below:



Final Evaluation: Acceptable. Data were provided to demonstrate that the minimum dose (10 mCi) is delivered throughout the generators' lifetime (60 days) at the minimum and maximum operating conditions (flow, time, elution volume) in all three elution modes. (b) (4)



(b) (4)

(b) (4)

These clinical performance results were provided to the clinical review team to assess the suitability of the clinical simulation test, who found them acceptable.

2. Provide the test method used to produce the data and summarized descriptions of how the reported data (volumes, flow and times) are determined. Include data for dose, flow, time (infusion, elution), system mode and volume (infusion, elution, total). Briefly describe the fluid path of the system as it delivers the entire infusion to the patient, including the radioactive dose and any non-radioactive saline – for each mode of operation, if different.

Response: received 29-Jun-2016 (SDN #43): The submission described the test method, including how reported data are determined (see summary below, Table C, compiled by this reviewer) and the method used to determine accuracy of that reported data in the dose error calculations. The fluid paths were also described in diagrams, see Appendix.

Evaluation: Acceptable. The test method and calculation of dose error are acceptable. Dose, volume and flow rate accuracy were determined against the “true” value and precision was determined from repeat (n=2) measurements.

| Table C. Summary of how submitted data are determined in clinical simulation tests | | |
|--|------------------|--------------|
| Data | System component | “True” value |
| (b) (4) | | |

*This true value is a best estimate based on use of a calibrated dose calibrator

(b) (4)

3. Describe the controls in place which prevent the system from operating when an undeliverable dose is requested by the user. This may include for example - software lockout controls, user manual instructions and labeling language.

Response received 29-Jun-2016 (SDN #43): JDI described the software controls in place and the constraints on the deliverable dose for all parameters (Table 6 below):

| Parameter | Minimal value | Maximal value |
|-----------|---------------|---------------|
| (b) (4) | | |

Table 6- generator activity level constraints

Evaluation: Acceptable.

(b) (4)

(b) (4)

C. Maximum volume at expiry (30L):

The Ruby-fill label (Section 2.7), in keeping with the RLD Cardiogen label, specifies three attributes for expiry: volume (30L), time (60 days) and strontium breakthrough level (Sr 82 and Sr 85) as follows:

2.7 RUBY-FILL® Expiration

Stop use of the RUBY-FILL® Rubidium Rb 82 Generator once any one of the following Expiration Limits is reached:

- A total elution volume of 30 L has passed through the generator column, or
- Expiration date of the generator (60 days post-manufacturing), or
- An eluate Sr 82 level of 0.01 $\mu\text{Ci/mCi}$ (kBq/MBq) Rb 82, or
- An eluate Sr 85 level of 0.1 $\mu\text{Ci/mCi}$ (kBq/MBq) Rb 82

The acceptance criteria for two of these attributes, time and breakthrough level, have been found acceptable in the previous review cycle. Acceptable stability data were provided in the original submission to support the 60 day expiry (see CMC review#1 David Place, Ph.D. 19-SEP-2014). The strontium breakthrough levels are based on the (b) (4)

The third attribute, volume (30L), was not discussed in CMC review #1. In this review cycle, the report "Investigation of Volume Limits and Strontium Breakthrough" RES.RBY.SDY.034 dated April 2014 to support the expiry volume was submitted on 01-Jun-2016 in Quality Amendment (DARRTS SDN#36). Figure 2 in the report (below) provides Sr-82 breakthrough values to show that generator delivered 30L of doses with strontium levels below the breakthrough acceptance criteria (b) (4)

(b) (4)

Initial Evaluation: Not Acceptable. The submission provides a summary report (b) (4)

The following comment was sent to the Applicant in Information Request #3 on 6-Sept-2016:

- 1. Provide data to support the proposed Ruby-fill generator labeled elution volume expiry of 30L. Describe the test procedure used to collect the data and differences from the proposed commercial product (e.g. generator, elution system, operating conditions).**

Response: Received (via email) 12-Sep-2016.

The Applicant provided four pieces of information to support their proposed 30L expiry data:

1. Developmental and validation generators – stability data in original NDA (30-Jun-2010):

Response: At the time of completing this review, this User Manual issue is still under negotiation with the Applicant. See subsequent CMC labeling memo.

Final Overall Evaluation of the Application:

Pending acceptable findings by CDRH review, which is not yet final as of this writing, the NDA is recommended for approval from a CMC standpoint, with the proposed volume expiry (30L) and time expiry (60 days) when operated using the commercial elution system (b) (4)

Pertinent communications with the Applicant during the Review Cycle:

| Communication type | Date sent to Applicant | CMC comments & deficiencies |
|----------------------------|------------------------|--|
| tcon | 11-May-2016 | See minutes in DARRTS. CMC discussed dose and calibration. |
| CMC Information Request #1 | 13-May-2016 | <p>1. Post-approval stability protocol:</p> <div data-bbox="688 467 1906 831" style="background-color: #cccccc; height: 224px; width: 580px; margin-left: 20px;"> (b) (4) </div> <p>2. Provide an assessment of the uncertainty in the dose (10 – 60 mCi) administered to the patient at the maximum and minimum range of the system operation (e.g. flow rates, ⁸²Rb concentration, elution volume/time) . Explain the basis for each assessment. Tabulate the data where possible.</p> |

N202-153 CHEMISTRY REVIEW#2

| Communication type | Date sent to Applicant | CMC comments & deficiencies |
|--------------------|------------------------|--|
| | | <p>3. Limit of detection for strontium in new low dose (10 mCi):</p> <ol style="list-style-type: none"> a. Discuss the capability of the dose calibrator to detect strontium at alert levels (^{82}Sr 0.004 per $\mu\text{Ci}/\text{mCi}$, ^{85}Sr 0.04 per $\mu\text{Ci}/\text{mCi}$) in the new low dose of 10mCi. Include an assessment of the limit of detection (LOD) for strontium using the supplied dose calibrator unit and the uncertainty in the dose calibrator measurement [REDACTED] ^{(b) (4)} Provide calculations used to determine the reported capability. b. Provide the study reports (data, analysis) referenced in the document “RUBY-FILL® Rubidium Rb 82 Generators - Evaluation of Strontium Isotope Breakthrough” (Appendix 3-1) which summarized test results for: <ol style="list-style-type: none"> i. RES.RBY.SDY.034 Volume Limits and Strontium Breakthrough ii. RES.RBY.SDY.033 Ruby-Fill Elution System (RbES) Breakthrough Testing iii. RES.RBY.SDY.042 Interim Report: Summary of RbES [REDACTED] ^{(b) (4)} Performance Testing iv. RES.RBY.SDY.054 REPORT: Summary of Ruby-Fill Elution System Performance Testing v. RES.RBY.SDY.070 REPORT: Summary of Ruby-Fill Elution System Performance Testing-Low Usage vi. RES.RBY.SDY.072 REPORT: Summary of Ruby-Fill Elution System <p>4. For the maximum (60mCi) and minimum (10mCi) dose range, provide an assessment of the delivered volume, concentration (mCi/mL) and rate of delivery (mL/min) for the lifetime of the generator (release, mid-life and expiry). Include an assessment of the uncertainty in the values and explain the basis for each assessment. Describe the rationale for the 30 mL/min maximum rate proposed in the draft label. Tabulate the information, where possible.</p> |

| Communication type | Date sent to Applicant | CMC comments & deficiencies |
|----------------------------|------------------------|--|
| CMC Information Request #2 | 20-Jun-2016 | <p>CMC Deficiency #1: Undeliverable dose. System performance has not been demonstrated over the lifetime of the generator for the full range of doses (10mCi to 60mCi) within the operating range of the system, consequently some doses may not be deliverable. The data provided from Clinical Simulation Test 4 did not provide test results for 10mCi at generator release or for 60mCi at generator expiry – conditions which represent the limits of the system as labeled.</p> <p>Information needed to resolve the deficiency:</p> <ol style="list-style-type: none"> 1. Provide data to demonstrate that the commercial system is capable of delivering the maximum and minimum dose (10mCi and 60mCi) at the beginning and end of generator lifetime within the proposed operating range of the system (b) (4) (b) (4) Include raw data, averages, n value, uncertainty and range of values (minimum and maximum). 2. Provide the test method used to produce the data and summarized descriptions of how the reported data (volumes, flow and times) are determined. Include data for dose, flow, time (infusion, elution), system mode and volume (infusion, elution, total). Briefly describe the fluid path of the system as it delivers the entire infusion to the patient, including the radioactive dose and any non-radioactive saline – for each mode of operation, if different. 3. Describe the controls in place which prevent the system from operating when an undeliverable dose is requested by the user. This may include for example - software lockout controls, user manual instructions and labeling language. <p>CMC Comment #1: Your proposal (b) (4) in your stability protocol is not acceptable. Revise the post-approval stability protocol (b) (4)</p> |
| tcon | 23-Jun-2016 | See minutes in DARRTS. CMC discussed data to support deliverable doses. |
| Clock Extension | 29-Jun-2016 | No CMC information was sent. The clinical division informed the Applicant “On June 15, 2016, we received your June 11, 2016, major amendment to this application. Therefore, we are extending the goal date by three months to provide time for a full review of the submission. The extended user fee |



N202-153 CHEMISTRY REVIEW#2



| Communication type | Date sent to Applicant | CMC comments & deficiencies |
|--|------------------------|---|
| | | goal date is September 30, 2016.” |
| Tcon | 13-Jul-2016 | See minutes in DARRTS. CMC discussed data to support deliverable doses. |
| Information Request #3 | 6-Sept-2016 | Provide data to support the proposed Ruby-fill generator labeled elution volume expiry of 30L. Describe the test procedure used to collect the data and differences from the proposed commercial product (e.g. generator, elution system, operating conditions).” |
| Information Request #4 (joint with clinical) | 15-Sept-2016 | <p>We are concerned that your user manual does not clearly explain [redacted] (b) (4)</p> <p>[redacted]</p> <p>[redacted] (b) (4)</p> |

IV. Labeling:

Labeling is ongoing by the review team at this time. See CMC labeling memo.

V. Overall recommendation: The application, as amended, is recommended for approval pending acceptable review by CDRH.

VI. Signatures:

CMC primary reviewer: Anne Marie Russell, Ph.D. I recommend approval (pending acceptable CDRH review).

CMC secondary reviewer: Danae Christodoulou, Ph.D. Branch Chief. I concur.



Danae
Christodoulou

Digitally signed by Danae Christodoulou
Date: 9/22/2016 02:58:25PM
GUID: 5050dd27000012a4c69bfc70b47660b7





Anne
Russell

Digitally signed by Anne Russell
Date: 9/22/2016 02:56:53PM
GUID: 508da7210002a03c7e3cba5e276a8027

ASSESSMENT OF THE FACILITIES

(b) (4)





OVERALL ASSESSMENT AND SIGNATURES: FACILITIES

Reviewer's Assessment and Signature:

There appears to be no significant or outstanding risks to the manufacturing process or final product based on the individual and composite evaluation of the listed facility's inspection results, inspectional history, and relevant experience. The facilities are determined acceptable to support approval of NDA202153

**Michael Klupal
4/18/16**

Secondary Review Comments and Concurrence:

I concur with Mr. Klupal's recommendations.

**Vidya Pai
4/26/2016**

Tertiary Review Comments and Concurrence:

I concur with the above recommendations.

**Krishna Ghosh
5/16/2016**

MEMORANDUM to FILE

To: NDA **202-153**

From: David A. Place, PhD Reviewing Chemist

Through: Eldon Leutzinger, PhD, Chemistry Lead

Through: Eric Duffy, PhD Director, ONDQA Division III

Subject: CMC Comparison of Labeling (Package Insert) and User Manual Documents for RubyFill

Date: December 29, 2014

Background – Jubilant DraxImage has submitted two key Amendments to NDA 202-153 that relate to the preparation and use of the Drug Product – both an updated Package Insert as well as a User Manual.

The titles, filenames, DARRTS submissions, and filing dates of these documents are as follows:

- Package Insert 1 14 1 3 Package Insert (clean) (2).doc SDN # 16 9/23/2013
- User Manual user-manual 18MAR2014.pdf SDN # 19 3/25/2014

The Table of Contents of both documents are reproduced on the following pages.

7 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

/s/

DAVID A PLACE
12/11/2014

ELDON E LEUTZINGER
12/11/2014

RAMESH RAGHAVACHARI
12/11/2014
for Eric Duffy

MEMORANDUM to FILE

To: NDA 202-153

From: David A. Place, PhD Reviewing Chemist

Through: Eldon Leutzinger, PhD, Chemistry Lead

Through: Eric Duffy, PhD Director, ONDQA Division III

Subject: Review Update

Date: December 10, 2014

Background – The following document contains several updates based on the 9/17/2014 Primary CMC review.

Note that the review cover page lists identifies the submission as an **ANDA**. After receipt, the submission was reclassified as an **NDA**. The review was done for the Division of Medical Imaging, not the Office of Generic Drugs.

H. Post-Approval Stability Protocol and Commitment – Post-approval, the sponsor proposes to carry out (b) (4) the proposed protocol below.

(b) (4)

The deletion of this sentence will improve consistency with the actions CMC will recommend to the sponsor.

Deficiencies to Communicate – (Suggested additions are in bold italic type).

The post-approval testing protocol needs to be more rigorous.

(b) (4)

(b) (4)

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

/s/

DAVID A PLACE
12/11/2014

ELDON E LEUTZINGER
12/11/2014

RAMESH RAGHAVACHARI
12/11/2014
for Eric Duffy

Review of Chemistry, Manufacturing, and Controls

NDA 202–153

Ruby-Fill[®]

Jubilant DraxImage, Inc.

by

Chemistry Reviewer: David A. Place, PhD

Division of New Drug Quality Assessment III Branch IX

for

Clinical Review Division: HFD–160

Division of Medical Imaging

and

Office of Generic Drugs

Table of Contents

| | |
|--|----|
| Table of Contents | 1 |
| Chemistry Review Data Sheet | 2 |
| Executive Summary | 5 |
| I. Recommendations | 5 |
| A. Recommendation and Conclusion on Approvability | 5 |
| B. Recommendation on Phase 4 (Post-Marketing) Commitments, Agreements, and/or Risk Management Steps, if Approvable | 5 |
| II. Summary of Chemistry Assessments | 5 |
| A. Description of the Drug Product(s) and Drug Substance(s) | 5 |
| B. Description of How the Drug Product is Intended to be Used | 5 |
| C. Basis for Approvability or Not-Approval Recommendation | 5 |
| III. Administrative | 6 |
| A. Reviewer’s Signature | 6 |
| B. Endorsement Block | 6 |
| C. CC Block | 6 |
| Chemistry Assessment | 7 |
| Review Notes | 7 |
| Drug Substance | 7 |
| Drug Product | 12 |
| Post-Approval Stability Protocol | 23 |
| Labeling | 23 |
| Environmental Assessment | 26 |
| Regional Information | 26 |
| IV. List Of Deficiencies To Be Communicated | 27 |

Chemistry Review Data Sheet

1. NDA 202–153

2. REVIEW # 1

3. REVIEW DATE: 17–SEP–2014

4. REVIEWER: David A. Place, PhD

5. PREVIOUS DOCUMENTS:

Previous Documents

N/A

Document Date

N/A

6. SUBMISSION(S) BEING REVIEWED:

Submission(s) Reviewed

Original

Amendment

Labeling Amendment (Container)

Labeling Amendment (Package Insert)

Document Date

18–JUN–2010

19–MAY–2011

20–DEC–2011

25–OCT–2012

7. NAME & ADDRESS OF APPLICANT:

Name: Jubilant DraxImage
Address: PO Box 1000, Montville, NJ 07045–1000
Representative: Philip Johnson, Deputy Director, Global Regulatory Affairs
Telephone: (973) 487–2181

8. DRUG PRODUCT NAME/CODE/TYPE:

- a) Proprietary Name: Ruby–Fill®
b) Non-Proprietary Name: Rubidium Rb–82 Chloride for Injection
c) Code Name/# (ONDQA only): NA
d) Chem. Type/Submission Priority (ONDQA only):
• Chem. Type: NA
• Submission Priority: NA

9. LEGAL BASIS FOR SUBMISSION: Not Applicable to NDAs

10. PHARMACOLOGICAL CATEGORY/INDICATION: Cardiac Positron Emission Tomography

11. DOSAGE FORM: Sterile solution for injection.

12. STRENGTH/POTENCY:

13. ROUTE OF ADMINISTRATION: IV

14. R /OTC DISPENSED: R OTC

15a. SPOTS (Special Products On–Line Tracking System)

_____ SPOTS product – Form Completed X Not a SPOTS product

15b. NANOTECHNOLOGY PRODUCTS: Not Applicable

16. CHEMICAL NAME, STRUCTURAL FORMULA, MOLECULAR FORMULA, MOLECULAR WEIGHT:

Chemical Name(s): Rubidium Rb–82 Chloride for Injection

IUPAC name: Rubidium Rb–82 Chloride for Injection

CAS Registry No. [132486-03-4]

Molecular Formula: ⁸²RbCl

Molecular Weight: 117.5 daltons

17. RELATED/SUPPORTING DOCUMENTS:

A. DMFs: NA

| DMF # | Type | Holder | Item Referenced | Code ^a | Status ^b | Date Review Completed | Comments |
|---------|------|---------|-----------------|-------------------|---------------------|-----------------------|-------------------|
| (b) (4) | II | (b) (4) | (b) (4) | 3 | Adequate | 1/18/2012 | Updated 4/17/2012 |
| | II | | | 3 | Adequate | 1/18/2012 | Updated 8/31/2012 |

a Action codes for DMF Table:

1 – DMF Reviewed.

Other codes indicate why the DMF was not reviewed, as follows:

2 –Type 1 DMF

3 – Reviewed previously and no revision since last review

4 – Sufficient information in application

5 – Authority to reference not granted

6 – DMF not available

7 – Other (explain under "Comments")

b Adequate, Inadequate, or N/A (There is enough data in the application, therefore the DMF did not need to be reviewed)

B. Other Documents:

| DOCUMENT | APPLICATION NUMBER | DESCRIPTION |
|----------|--------------------|-------------|
| N/A | | |

Patent: NA

Exclusivity: NA.

18. STATUS:

ONDC:

| CONSULTS/ CMC Related Reviews | RECOMMENDATION | DATE | REVIEWER |
|--|------------------------------------|-------------|-------------------|
| Biometrics | NA | | |
| EES | Acceptable | 1/16/2014 | OC |
| Pharm/Tox | NA | | |
| Biopharm | NA | | |
| Methods Validation | Acceptable per this CMC review | 9/17/2014 | D. Place |
| DMEPA | Acceptable | 4/2/2014 | M. Rutledge |
| EA | Categorical Exclusion – Acceptable | 9/17/2014 | D. Place |
| Microbiology | Acceptable | 2/29/2012 | D. Palmer–Ochieng |
| DMIP/safety | Deficiencies Identified | 6/27/2014 | I. Krefting |

OGD:

| CONSULTS/ CMC Related Reviews | RECOMMENDATION | DATE | REVIEWER |
|--|-----------------------|-------------|-----------------|
| Microbiology | N/A | | |
| EES | N/A | | |
| Methods Validation | N/A | | |
| Labeling | N/A | | |
| Bioequivalence | N/A | | |
| EA | N/A | | |
| Radiopharmaceutical | N/A | | |

19. ORDER OF REVIEW (OGD Only): *Not Applicable*

The application submission(s) covered by this review was taken in the date order of receipt. Yes
 No If no, explain reason(s) below:

Chemistry Review for NDA 202–153

Executive Summary

I. Recommendations

A. Recommendation and Conclusion on Approvability

NDA 202–153 is not recommended for approval from a CMC standpoint until a complete response on identified CMC deficiencies is received from the sponsor.

B. Recommendation on Phase 4 (Post-Marketing) Commitments, Agreements, and/or Risk Management Steps, if Approvable

None identified.

II. Summary of Chemistry Assessments

A. Description of the Drug Product(s) and Drug Substance(s)

The Ruby-Fill® 82Sr/82Rb generator provides an eluted solution of the drug substance, Rubidium Rb82 Chloride Injection in sterile normal saline. ^{82}Rb is produced on the generator by the radioactive decay of ^{82}Sr . ^{82}Sr remains bound to the column while ^{82}Rb is eluted from the column as RbCl with 0.9% sodium chloride.

^{82}Rb decays by positron emission with a half-life of 1.273 minutes (76.38 sec) to stable ^{82}Kr gas. Due to this very short half-life, the Ruby-Fill elution system will be located in very close proximity to directly dose the patient being imaged to allow prompt injection of the Rubidium-82 eluate. Also, due to the short half-life (as with other PET radioisotopes), proactive sterility controls must be in place.

B. Description of How the Drug Product is Intended to be Used

It is indicated as an agent for positron emission tomography (PET) imaging, specifically for the assessment of myocardial perfusion to aid in the diagnosis of coronary artery disease, as the Rubidium ions mimic the cardiac biological function of Potassium ions. As this product involves a radionuclide generator and a product delivery device (b) (4) there are two important issues: (1) how the (b) (4) works with the generator and (2) the nature of the various factors underlying user interaction with the system to assure operation of the system to produce a safe product. Accordingly, consults are requested for CDRH Device Engineering and CDRH Human Factors Assessment.

C. Basis for Approvability or Not-Approval Recommendation

A critical CMC issue has been identified in the Chemistry sections of the submission. The level of post-approval testing proposed is inadequate. (b) (4)

III. Administrative**A. Reviewer's Signature**

Chemist David A. Place, PhD _____ Date: 17-SEP-2014

B. Endorsement Block

Chemistry Lead Eldon Leutzinger, PhD _____ Date:
Division Director Eric P. Duffy, PhD _____ Date:

cc: Orig. NDA 202-153
 HFD-160

6 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page

FDA CDER EES
ESTABLISHMENT EVALUATION REQUEST
SUMMARY REPORT

| | | | |
|----------------|----------------|--|---|
| Application: | ANDA202153/000 | Sponsor: | DRAXIMAGE |
| Org. Code: | 600 | | 7361 CALHOUN PL STE 500 |
| Priority: | | | ROCKVILLE, MD 20855 |
| Stamp Date: | 30-JUN-2010 | Brand Name: | |
| PDUFA Date: | 30-APR-2011 | Estab. Name: | RUBIDIUM CHLORIDE RB 82 |
| Action Goal: | | Generic Name: | |
| District Goal: | 01-MAR-2011 | Product Number; Dosage Form; Ingredient; Strengths | 001; GENERATOR; RUBIDIUM CHLORIDE RB-82; (b) (4)mCi |

| | | | | |
|---------------|-------------|------------------------|-----------|------------|
| FDA Contacts: | R. D COSTA | Prod Qual Reviewer | (HFD-623) | 2402768407 |
| | M. GONITZKE | Product Quality PM | (HFD-600) | 2402768422 |
| | D. DOAN | Regulatory Project Mgr | (HFD-617) | 2402769336 |
| | ID = 109049 | Team Leader | | |

| | | | | | |
|-------------------------|------------|----------------|--------------|----|------------|
| Overall Recommendation: | ACCEPTABLE | on 16-JAN-2014 | by T. WILSON | () | 2404024226 |
| | PENDING | on 02-OCT-2013 | by EES_PROD | | |

| | | | |
|-------------------|--|-------------|------------|
| Establishment: | CFN: | FEI: | 3009003838 |
| | JUBILANT DRAXIMAGE INC 16751 RTE TRANS CANADA KIRKLAND, , CANADA H9h 4j4 | | |
| DMF No: | | AADA: | |
| Responsibilities: | FINISHED DOSAGE MANUFACTURER | | |
| Profile: | POSITRON EMISSION TOMOGRAPHY | OAI Status: | NONE |
| Last Milestone: | OC RECOMMENDATION | | |
| Milestone Date: | 16-JAN-2014 | | |
| Decision: | ACCEPTABLE | | |
| Reason: | DISTRICT RECOMMENDATION | | |

| | | | | |
|-------------------|------------------------------|-------------|------|---------|
| Establishment: | CFN: | (b) (4) | FEI: | (b) (4) |
| | (b) (4) | | | |
| DMF No: | | AADA: | | |
| Responsibilities: | FINISHED DOSAGE OTHER TESTER | | | |
| Profile: | CONTROL TESTING LABORATORY | OAI Status: | NONE | |
| Last Milestone: | OC RECOMMENDATION | | | |
| Milestone Date: | 26-MAR-2013 | | | |
| Decision: | ACCEPTABLE | | | |
| Reason: | BASED ON PROFILE | | | |

Establishment: CFN: (b) (4) FEI: (b) (4)
(b) (4)

DMF No: AADA:

Responsibilities: FINISHED DOSAGE OTHER TESTER

Profile: CONTROL TESTING LABORATORY OAI Status: NONE

Last Milestone: OC RECOMMENDATION

Milestone Date: 27-DEC-2013

Decision: ACCEPTABLE

Reason: DISTRICT RECOMMENDATION

Establishment: CFN: (b) (4) FEI: (b) (4)
(b) (4)

DMF No: AADA:

Responsibilities: FINISHED DOSAGE OTHER TESTER

Profile: CONTROL TESTING LABORATORY OAI Status: NONE

Last Milestone: OC RECOMMENDATION

Milestone Date: 26-MAR-2013

Decision: ACCEPTABLE

Reason: BASED ON PROFILE

Establishment: CFN: (b) (4) FEI: (b) (4)
(b) (4)

DMF No: AADA:

Responsibilities: FINISHED DOSAGE OTHER TESTER

Profile: CONTROL TESTING LABORATORY OAI Status: NONE

Last Milestone: OC RECOMMENDATION

Milestone Date: 26-MAR-2013

Decision: ACCEPTABLE

Reason: BASED ON PROFILE

Establishment: CFN: (b) (4) FEI: (b) (4)
(b) (4)

DMF No: AADA:

Responsibilities: FINISHED DOSAGE OTHER TESTER

Profile: CONTROL TESTING LABORATORY OAI Status: NONE

Last Milestone: OC RECOMMENDATION

Milestone Date: 26-MAR-2013

Decision: ACCEPTABLE

Reason: BASED ON PROFILE

8 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page



Batch Records – Executed, bilingual batch records have been submitted, fully reflecting the manufacturing process and controls.

Satisfactory.

Labeling – The PACKAGE INSERT contains the following black box warning:




Satisfactory. The Package Insert is equivalent to the reference listed drug.

Container Labels – Since Ruby-Fill is radioactive, there are no internal labels.

Outer Main Label – This label is applied to the lead “pig”. It is general in nature.



Satisfactory. Note: None of the labeling 



(b) (4)



Deficiencies to Communicate –

The post-approval testing protocol needs to be more rigorous.

(b) (4)

(b) (4)

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

/s/

DAVID A PLACE
09/17/2014

ELDON E LEUTZINGER
09/17/2014

ERIC P DUFFY
09/19/2014

Initial Quality Assessment (IQA)
For
Division of New Drug Quality Assessment III, Branch VII
Office of New Drug Quality Assessment

OND Division: OGD

ANDA: 202-153

Applicant: Draximage

16751 Autoroute Transcanadienne / Trans-Canada Highway
Kirkland(Quebec) Canada H9H 4J4

Stamp Date: 06/30/2010

Trademark: Ruby-Fill

USAN: None

INN: None

Company Code: None

Established: Rubidium Rb 82 Generator

Dosage Form: Sterile solution

Route of Administration: IV

Indication: assessment of regional myocardial perfusion (b) (4)

CMC Lead: Eldon E. Leutzinger, Ph.D., Branch VII

| | | |
|---|-----|----|
| ONDQA Fileability (N/A) (OGD function) | YES | NO |
|---|-----|----|

Comments for 74-Day Letter: N/A

Summary and Critical Issues:

A. Summary

The **Drug Product** (Ruby-Fill) is a radionuclide generator (Rubidium Chloride Rb 82 Generator) that contains at calibration (b) (4) $^{82}\text{SrCl}_2$ adsorbed onto hydrous (b) (4) stannic oxide in a column. Elution of the generator column with 0.9% Sodium Chloride Injection USP produces Rubidium Chloride Rb 82 Injection USP. It contains (b) (4) $^{82}\text{RbCl}$ activity delivered depends on the elution rate and the amount of volume eluted, based on the intended dose, as well as (of course) the amount of $^{82}\text{SrCl}_2$ adsorbed onto the column. (b) (4)

the set up is shown as follows. See the next review page.

14 Pages have been Withheld in Full as B4 (CCI/TS) immediately following this page

In this IQA, I have not tried to capture all of the considerations and issues that might be involved in this ANDA, but have tried to identify those issues I think are most important and in particular in relation to ^{82}Sr breakthrough, since it relates most severely to generator column performance.

Manufacturing Facilities:



CMC Lead: Eldon E. Leutzinger, Ph.D. Date: 11/27/2012
Division of New Drug Quality Assessment III, Branch VII

Division Director: Eric Duffy, Ph.D.
Division of New Drug Quality Assessment III

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

/s/

ELDON E LEUTZINGER
11/28/2012

ERIC P DUFFY
12/12/2012