

Attachment VI:**Summary of Safety and Effectiveness Information
[510(k) Summary]****SUBMITTER**

Synthes (USA)
1690 Russell Road
Paoli, PA 19301
(610) 647-9700

Contact: Sheri L. Musgnung

DEVICE NAME:

Synthes Locking Condylar Plate (LCP) System

**COMMON OR USUAL
NAME**

Appliance, Fixation, Nail/Blade/Plate Combination, Single
Component;
Screw, Fixation, Bone

**DEVICE
CLASSIFICATION:**

Class II, 21 CFR 888.3030 and 888.3040

PREDICATE DEVICE:

Synthes Distal Femur Plate (DFP) System, (K982222)

DESCRIPTION:

Synthes LCP System is a plate and screw system. The primary feature of the plate and screw system is that the locking screws engage with the head and shaft of the plate to form a locked, fixed angle construct. The system features a limited-contact profile, dynamic compression unit (DCU) and locking screw holes. The plate accepts 7.3 mm locking screws, 7.3 mm conical screws, 5.0 mm locking screws, 5.0 mm cannulated screws, and 4.5 mm cortex screws. A screw nut is also utilized with this system.

INTENDED USE:

Synthes LCP System is intended for buttressing multifragmentary distal femur fractures including: supracondylar, intra-articular and extra-articular condylar fractures, fractures in normal or osteopenic bone, and non-unions and malunions.

MATERIAL:

Stainless Steel



MAR 15 2000

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ms. Sheri L. Musgnung
Regulatory Affairs Specialist
Synthes (USA)
1690 Russell Road
Paoli, Pennsylvania 19301

Re: K000066
Trade Name: Synthes Locking Condylar Plate (LCP) System
Regulatory Class: Class II
Product Codes: HRS, HWC, HTY, and JDW
Dated: January 7, 2000
Received: January 10, 2000

Dear Ms. Musgnung:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general control provisions of the Act. The general control provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

Page 2 – Ms. Sheri L. Musgnung

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4595. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten", followed by a small flourish.

Celia M. Witten, Ph.D., M.D. *for*
Director
Division of General and
Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K000066

Device Name: Synthes (USA) Locking Condylar Plate (LCP) System

Indications For Use:

Synthes Locking Condylar Plate (LCP) System is intended for buttressing multifragmentary distal femur fractures including: supracondylar, intra-articular and extra-articular condylar fractures, fractures in normal or osteopenic bone, and non-unions and malunions.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use X OR Over-The-Counter Use _____
(Per 21 CFR 801.109)

NPO for CDRH
(Division Sign-Off)
Division of General Restorative Devices
510(k) Number K000066