

## A 510(k) Summary of Safety and Effectiveness

DEC 13 2001

**Submitter's name:** Smith & Nephew, Inc., Orthopaedic Division  
**Submitter's address:** 1450 Brooks Road, Memphis, TN 38116  
**Submitter's telephone number:** 901-399-6487  
**Contact person:** David Henley, Senior Clinical/Regulatory Affairs Specialist  
**Date summary prepared:** July 25, 2001  
**Trade or proprietary device name:** 3D Humeral Heads  
**Common or usual name:** Shoulder Joint Prosthesis

**Classification name:** 21 CFR 888.3660, shoulder joint metal/polymer, semi-constrained cemented prosthesis – Class II

**Substantially Equivalent  
 Legally Marketed Devices**

- Aequalis™ Shoulder System – Tornier, S.A.
- Anatomica Humeral Stems/Heads – Sulzer Orthopedics

**Device Description**

3D humeral heads are manufactured from forged cobalt chrome (CoCr) material (ASTM F799) and can be used with existing cobalt chrome humeral stem components from the Cofield<sup>2</sup> Total Shoulder System previously cleared from market. 3D taper sleeves are manufactured from titanium (Ti-6Al-4V) material conforming to the requirements of ASTM F1472.

**Device Intended Use**

The subject humeral head devices can be mated with approved humeral stem components from the Cofield<sup>2</sup> Total Shoulder System previously cleared for market under K955767. 3D Humeral Heads are indicated for use as orthopedic implants for the partial or total replacement of the human shoulder joint articulating either directly against the glenoid face or a compatible glenoid component, respectively. 3D Humeral Heads are intended for use with bone cement only (cemented fixation) and for single use only. 3D Humeral Heads are intended for the following indications:

Proximal Humeral Prosthesis – (1) complex, acute fractures or fracture-dislocations of the humeral head (e.g. trauma – three and four-part injuries in the Neer classification, or head splitting, or head impression fractures); (2) complex, chronic fractures or fracture-dislocations of the humeral head with malunion, non-union of a small osteoporotic head fragment, or chronic dislocation with loss of humeral head cartilage, or large impression fractures; (3) avascular necrosis with intact glenoid cartilage; and (4) selected patients with arthritis who do not have adequate scapular bone to support a glenoid component or must engage in moderately heavy activities.

Total Shoulder Arthroplasty (when used in conjunction with a compatible glenoid component) – severe destruction of the glenohumeral articular surfaces with intractable chronic pain in rheumatoid arthritis, osteoarthritis, traumatic arthritis, cuff tear arthroplasty, ancient septic arthritis, avascular necrosis with secondary glenoid changes, radiation necrosis, and other failed reconstructive procedures.

The assembled humeral stem component (including a 3D Humeral Head) may be used alone for hemiarthroplasty or combined with a Cofield<sup>2</sup> Total Shoulder System glenoid component for use in total shoulder arthroplasty.

**Technological Characteristics:**

3D Humeral Heads are similar to the legally marketed predicate devices listed above. All of these devices are indicated for total shoulder arthroplasty or hemiarthroplasty, are similar in design to the 3D Humeral Heads and have the same technological characteristics.

**Performance Characteristics:**

Mechanical *humeral head distraction* testing and *ion release analysis* was performed on these devices. Results indicate that the subject devices meet or exceed acceptable performance. Data indicate that 3D Humeral Heads are substantially equivalent to legally marketed devices.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. David Henley  
Senior Clinical/Regulatory Affairs Specialist  
Smith & Nephew, Inc.  
1450 Brooks Road  
Memphis, Tennessee 38116

DEC 13 2001

Re: K012378

Trade/Device Name: 3D Humeral Heads  
Regulation Number: 21 CFR §888.3660 and §888.3690  
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis;  
Shoulder joint humeral (hemi-shoulder) metallic cemented or  
uncemented prosthesis

Regulatory Class: Class II  
Product Code: KWS & HSD  
Dated: October 25, 2001  
Received: October 26, 2001

Dear Mr. Henley:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 - Mr. David Henley

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten". The signature is written in a cursive style and is positioned to the right of a small, stylized "for" symbol.

Celia M. Witten, Ph.D., M.D.  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Premarket Notification  
Indications Enclosure**

510(k) Number (if known): **K012378**

Device Name: **3D Humeral Heads**

Indications for Use:

When used with an appropriate humeral stem from the Cofield<sup>2</sup> Total Shoulder System, **3D Humeral Heads** are indicated for use as orthopedic implants for the partial or total replacement of the human shoulder joint articulating either directly against the glenoid face or a compatible glenoid component from the Cofield<sup>2</sup> or Neer III Total Shoulder Systems, respectively. When used with an appropriate humeral stem, **3D Humeral Heads** are intended for use with bone cement only (i.e. cemented fixation) and for single use only. **3D Humeral Heads** are intended for the following indications:

Proximal Humeral Prosthesis – (1) complex, acute fractures or fracture-dislocations of the humeral head (e.g. trauma – three and four-part injuries in the Neer classification, or head splitting, or head impression fractures); (2) complex, chronic fractures or fracture-dislocations of the humeral head with malunion, non-union of a small osteoporotic head fragment, or chronic dislocation with loss of humeral head cartilage, or large impression fractures; (3) avascular necrosis with intact glenoid cartilage; and (4) selected patients with arthritis who do not have adequate scapular bone to support a glenoid component or must engage in moderately heavy activities.

Total Shoulder Arthroplasty (when used in conjunction with a compatible glenoid component as described above) – severe destruction of the glenohumeral articular surfaces with intractable chronic pain in rheumatoid arthritis, osteoarthritis, traumatic arthritis, cuff tear arthroplasty, ancient septic arthritis, avascular necrosis with secondary glenoid changes, radiation necrosis, and other failed reconstructive procedures.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*for Mark N. Millerson*  
(Division Sign-Off)  
Division of General, Restorative  
and Neurological Devices

510(k) Number K012378

Prescription Use X  
(Per 21 CFR 801.109)

OR

Over-the-Counter Use \_\_\_\_\_  
(Optional Format 1-2-96)