

NOV 07 2001

K012605



ALLIANCE  
MEDICAL CORPORATION

# PART B: 510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

10232 South 51st Street  
Phoenix, Arizona 85044

TEL 480.763.5300  
FAX 480.763.5310  
Toll Free 888.888.3433  
www.alliance-medical.com

**Submitter:** Alliance Medical Corporation  
10232 South 51<sup>st</sup> Street  
Phoenix, Arizona 85044

**Contact:** Don Selvey  
Vice President, Regulatory Affairs and Quality Assurance  
(480) 763-5300

**Date of preparation:** August 11, 2001

**Name of device:** Trade/Proprietary Name: Reprocessed Arthroscopics Burs  
Common or Usual Name: Arthroscopic Bur  
Classification Name: Arthroscope

### Reprocessed devices:

Manufacturer	Description	Model
Stryker®	6 Flute Round Bur	275-846
Stryker®	Barrel Bur 8 Flute	275-849-000
Stryker®	Standard Round Bur	275-851-000
Stryker®	Egg Bur 8 Flute Standard	275-852-000
Stryker®	Bur Pear 8 Flute	275-853
Stryker®	Aggressive Egg Bur 6 Flute	275-857-000
Stryker®	Bur Pear Aggressive 6 Flute	275-858
Stryker®	Barrel Bur 8 Flute	275-859
Stryker®	Round Bur 6 Flute	275-866
Stryker®	Round Bur Hollow 6 Flute	275-940-000
Stryker®	Aggressive Barrel Bur 6 Flute	275-949
Stryker®	Aggressive Round Bur Hollow	275-950-000
Stryker®	Round Bur Hollow	275-950-012
Stryker®	Barrel Bur Aggressive 6 Flute	275-951-000
Stryker®	Barrel Bur Hollow	275-951-012
Stryker®	Ultra Aggressive Barrel Bur 4	275-955
Stryker®	Aggressive Barrel Bur 6 Flute	275-959
Stryker®	Ultra Aggressive Bur 4 Flute	275-965
Stryker®	6 Flute Barrel Bur	275-969-000

**Predicate device(s):** **K963332** Stryker® Thermo-Plastic Shaver Blades  
**K973195** Stryker® Total Performance System Shaver Handpiece  
**K982375** Stryker®, Stryker Hip Arthroscopy Set



**ALLIANCE**  
MEDICAL CORPORATION

**Device description:** Arthroscopic Burs can be used to abrade, cut and excise tissue and bone; remove loose fragments; and shave away debris in arthroscopic surgeries, as well as surgeries of the jaw and sinuses.

The Arthroscopic Bur components reprocessed by Alliance Medical Corporation include a bur or blade at the end of a long rod that rotates within a long hollow stainless steel housing. The housing has a window cut out on one side of the distal end, allowing the bur to cut one structure while the adjacent one is still protected by the housing on the opposite side of the bur or blade. This system attaches to a motorized handpiece that drives the internal bur or blade inside the outer housing and provides suction to pull the cut tissue away from the surgical site.

**Intended use:** Reprocessed Arthroscopic Burs are intended for resecting tissue and bone found in articular body cavities during orthopedic, maxillofacial, hand, foot and plastic surgery in patients requiring arthroscopic or orthopedic surgery.

**Indications statement:** Reprocessed Arthroscopic Burs are indicated for use in orthopedic surgical procedures of the joints, jaw or sinuses where the cutting and removal of soft and hard tissue or bone is needed in patients requiring orthopedic surgery.

**Technological characteristics:** The design, materials, and intended use of the Reprocessed Arthroscopic Burs are identical to the predicate devices. The mechanism of action of the Reprocessed Arthroscopic Bur is identical to the predicate devices in that the same standard mechanical design, materials, shapes and sizes are utilized. There are no changes to the claims, intended use, clinical applications, patient population, performance specifications, or method of operation.

**Performance data:** Bench and laboratory testing was conducted to demonstrate performance (safety and effectiveness) of the Reprocessed Arthroscopic Burs.

- Biocompatibility
- Validation of reprocessing
- Function test(s)

Performance testing demonstrates that Reprocessed Arthroscopic Burs perform as originally intended.

**Conclusion:** In accordance with the Federal Food, Drug and Cosmetic Act 21 CFR Part 807 and based on the information provided in this premarket notification, Alliance Medical Corporation concludes that the modified device (the Reprocessed Arthroscopic Bur) is safe, effective and substantially equivalent to the predicate devices as described herein.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

NOV 07 2001

Mr. Don Selvey  
Vice President, Regulatory Affairs  
and Quality Assurance  
Alliance Medical Corporation  
10232 South 51<sup>st</sup> Street  
Phoenix, Arizona 85044

Re: K012605

Trade/Device Name: Alliance Medical Corporation Reprocessed Arthroscopic Burs  
Regulation Number: 888.1100  
Regulation Name: Arthroscope  
Regulatory Class: II  
Product Code: HRX  
Dated: August 10, 2001  
Received: August 13, 2001

Dear Mr. Selvey:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Don Selvey

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



 Celia M. Witten, Ph.D., M.D.  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

NOV 07 2001

K012605

## II. Indications for Use Statement

510(k) Number (if known): K012605

Device Name: Alliance Medical Corporation Reprocessed Arthroscopic Burs

**Indications for Use:** Reprocessed Arthroscopic Burs are indicated for use in orthopedic surgical procedures of the joints, jaw or sinuses where the cutting and removal of soft and hard tissue or bone is needed in patients requiring orthopedic surgery.

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use   
(per 21 CFR 801.109)

or

  
Over-the-Counter Use   
**(Division Sign-Off)**  
Division of General, Restorative  
and Neurological Devices

510(k) Number K012605

CONFIDENTIAL

Alliance Medical Corporation  
Reprocessed Arthroscopic Burs  
Traditional 510(k)

12