



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAY 08 2002

Norav Medical Ltd.
c/o Mr. Steve Springrose
Springrose Partners, LLP
12505 58TH Avenue N
Plymouth, MN 55442

Re: K012712

Trade Name: Norav Holter Analysis System, Model NH-300, V1.07
Regulation Number: 21 CFR 870.2800
Regulation Name: Medical Magnetic Tape Recorder
Regulatory Class: Class II (two)
Product Code: MLO
Dated: February 6, 2002
Received: February 7, 2002

Dear Mr. Springrose:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 - Mr. Steve Springrose

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4646. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Donna-Bea Tillman, Ph.D.
Acting Director
Division of Cardiovascular
and Respiratory Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510K Number: K012712

Device Name: Norav Holter Analysis System

Indications for Use:

The Norav Holter Analysis System is intended for patients requiring ambulatory (Holter) monitoring from 1 to 48 hours. Such monitoring is most frequently used for the indications below:

- 1) Evaluation of symptoms suggesting arrhythmia or myocardial ischemia.
- 2) Evaluation of ECG documenting therapeutic interventions in individual patients or groups of patients.
- 3) Evaluation of patients for ST segment changes.
- 4) Evaluation of a patient's response after resuming occupational or recreational activities (e.g. after M.I. or cardiac surgery.)
- 5) Clinical and epidemiological Research Studies
- 6) Norav Holter Systems containing Heart Rate Variability (HRV) software have been tested to measure Heart Rate Variability within 4 ms tolerance. The clinical significance of Heart Rate Variability measures should be determined by a physician.
- 7) Evaluation of patients with pacemakers.

(PLEASE DO NOT WRITE BELOW THIS LINE- CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ☒
 (Per 21 CFR 801.109)

OR

Over-the-Counter Use
 (Optional Format)


Division of Cardiovascular & Respiratory Devices
510(k) Number K012712