

Submitter: Card Guard Scientific Survival Ltd.,
2 Pekeris St. P.O.B. 527
Rehovot 76100, Israel

K013879
P.1/2

Tel: 972-8-9484600

FEB 01 2002

Fax: 972-8-9484605

Contact Person: Alex Gonorovsky,
Deputy Chief Engineer, Regulatory Affairs

Tel: 972-8-9484624

E-mail: alexanderg@cardguard.com

Date Prepared: August 12, 2001

1. Definition and Intended Use

The Telemedicine 2000 (TM2000) Transtelephonic Receiving Center is a software system designed and intended for supporting a remote (transtelephonic) monitoring ECG, Spirometric and Fetal/Maternal patient parameters.

The enhanced TM2000 enables receiving, storing, displaying, measuring, updating, printing and re-transmitting of patient ECG, Spirometric and Fetal/Maternal parameters and other patient related data, (such as demographics, doctors, medical history and status, diagnoses, etc.), and business information (e.g. billing, inventory sales etc.).

2. Device Class

The TM2000 Transtelephonic Receiving Center system is classified as Class II medical device (21 C.F.R. Par. 870.2920 (1992)).

3. Applicable Regulatory Documents and Card Guard Procedures

No performance standards have been developed under Section 514 of the Federal Food, Drug and Cosmetic Act for telephone ECG and Spirometric transmitter devices.

The complete list of the Applicable Documents referenced and/or incorporated in this project is provided in the FDA submission package: *Chapter 19, Applicable Normative Documents: Standards, Regulations, Guidances, Procedures; Publications*

Chapter 19 includes the following sections:

- Applicable FDA Documents
- American National Standards Institute (ANSI), Association for the Advancement of Medical Instrumentation (AAMI)
- International Electrotechnical Commission (IEC/CEI), International Organization for Standardization (ISO), European Norm (EN)
- Code of Federal Regulations
- International Special Committee On Radio Interference (CISPR)
- Institute of Electrical and Electronics Engineers (IEEE)
- Card Guard Ltd: Applicable Procedures
- Card Guard Ltd: Product Definitive Reference
- Articles, Publications



4. Operational Characteristics, Features and Functions

K013879
P. 2/2

1. Runs on any MS Windows operating system
2. Maximum database file size of 2 GB,
Storage of up to 10 - 40 thousand transmissions (depending on transmitter type)
3. Accessing and updating the receiving center DB.
Manual entry of patient and physician detail
4. Receiving and processing patients' transmitted signals and medical data
5. Graphic representation of ECG, Spirometric and FM signals
6. Intuitive, user friendly HMI
7. Availability for signal/data transmission over Web
8. Localization
9. Reports generation
10. Reports previewing and printing
11. Manageable security, fail-safe protection against unauthorized access.
Discretionary access control (restrictions based on privileges)

5. User Interface

The TM2000 Graphic User Interface (GUI) features pull down menus and dialog boxes for representation and updating of data. Generation of GUI and implementation of the related DB connectivity mechanism shall utilize JVM and the C/C++ native methods in conjunction with the Windows API functions.

6. Substantial Equivalence

Card Guard hereby claims that the TM2000 is substantially equivalent to Telemedicine 2000, the Transtelephonic Receiving Center, K992164

The proof of substantial equivalence in all that concerns the intended use, principles of operation, features and technological characteristics is provided in Chapter 7. *Substantial Equivalence to Cleared Devices.*

7. Design Controls and Hazard Analysis

The Card Guard's product design procedure, and QA and QC policy, formalize the design and production process and assure that all the respective requirements are met. In the framework of the Design Controls the testing was conducted to verify the system compliance with all its design specifications.

The device Level of Concern criteria were evaluated and the system was determined to be *a moderate level of concern system.*

The rigorous design evaluation and the System Safety and Risk analysis expose potential failures or possible system flaws which could directly or indirectly effect the patient.

8. Conclusions

The system constitutes a safe and reliable means for receiving, storing, displaying, analyzing, updating, printing and re-transmitting of patient ECG and Spirometric parameters and other patient related data.

Its operation present no adverse health effect or safety risks to patients when used as intended.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 01 2002

Mr. Alex Gonorovsky
Deputy Chief Engineer, Regulatory Affairs
Card Guard Scientific Survival, Ltd.
2 Pekeris Street
Rehovot
ISRAEL

Re: K013879

Trade Name: TM2000 Telemedicine Receiving Center
Regulation Number: 21 CFR 870.2920
Regulation Name: Telephone Electrocardiograph Transmitters and Receivers
Regulatory Class: Class II (two)
Product Code: DXH
Dated: November 20, 2001
Received: November 23, 2001

Dear Mr. Gonorovsky:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2 - Mr. Alex Gonorovsky

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4646. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Bram D. Zuckerman, M.D.
Acting Director
Division of Cardiovascular
and Respiratory Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

