

JUN 14 2002

K013977

'510 (k) Summary'  
Summary and safety effectiveness

Submitter's Name & Address: Fischer-Zoth  
Audiologic Systems  
1257 W. Margaret View Circle  
Riverton, UT 84065

Contact person & Telephone: Mr. Michael Sorensen  
(801) 541 1123

Data summary prepared: November 8th, 2001

Device name: Classification Name – Audiometer, Neurology  
Common/Usual Name - Screening Audiometer  
Proprietary name – Echo-Screen T, TA, TD, TDA, TC

Predicate Devices:

Fischer-Zoth, model Echo-Screen: 510(k) K982642  
SLE Limited, model SABRE: 510(k) K993177  
see also: Appendix A

Device description , intended Use & Effectiveness:

The Fischer-Zoth Audiologic Systems model family 'Echo-Screen T' consisting of Echo-screen T, Echo-Screen TA, Echo-Screen TD, Echo-Screen TDA, Echo-Screen TC are based upon Otoacoustic Emissions (OAE) and Auditory Brainstem Response (ABR) technology. Similar techniques and methods are – among others - used by the Fischer-Zoth model Echo-Screen and SLE Limited model SABRE.

The device is intended to screen hearing for newborns through adults, including geriatric patients. The device does not measure hearing per se, but helps to determine whether or not a hearing loss may be present.

The 'Echo-Screen T' product family are handheld, automated OAE- and ABR based hearing screening systems which are easy to be used. The measurement flow is menu guided and the evaluation is based upon signal statistics. They are intended to be used by trained personnel in a medical or school environment. The Echo-Screen T models are not intended for fitting assistive listening devices such as hearing aids or cochlea implants.

Technological Characteristics: (also see Appendix B)

- Probe: miniature probe easy to apply even for premature infants
- Electrodes: GXY, from Medicotest Inc. 510(k) #k931430  
or other FDA-approved products
- Hardware:
  - Hand held device : size 21,5x10x5,4cm , weight 600g.
  - Battery powered : 6V/1500mAh, rechargeable
  - Stimulus Rate: 40....100Hz
  - Signal Bandwidth : 1,4KHz to 4KHz
- Software: echo-screen evaluation SW

Automated stimulus adjustment and online control  
Signal statistical criterion for 'PASS' condition  
Additional hints to measurement quality (stimulus stability and artefact rate)

- **Safety:**

Battery powered, no connection to mains is possible during measurement  
Sound Pressure level: 70-85 dB SPL maximum  
Current through electrodes limited to < 10µA maximum

Safety:

The 'Echo-Screen' Screening System is designed to provide electrical safety to the patient as well as the user. The system is designed to meet the following standards related electrical safety: IEC601-1, UL2601 and VDE750 (German MedGv).

The Echo-Screen T models are stand alone-system. They are battery powered with a 6V battery. During operation there are no connections to the mains.

During measurement there is a conductive connection to the patient via three Electrodes. The current through each single channel during the measurement is actively limited to < 5µA.

To prevent excessive tone levels within the ear, the Echo-Screen T models are continuously controlling the level of the output tone or click-burst. The max. output level is controlled via software to 85dB SPL (sound pressure level) and the electronic hardware is build in such a way, that the speakers within the probe are incapable of producing enough decibels to permanently damage the ear.

The system will also be certified according to the following standards:

- ISO 9001                      Quality Management System
- EN 60601                    Safety
- EN 46001                    Class IIa device, disposable are Class I
- EN 55011                    Group 1 (medical equipment, conducted and radiated emissions)
- IEC801-3                    Radiated Electromagnetic Field Susceptibility
- CE Mark                     Conforms the provisions of European council Directive 93/42EEC concerning medical devices.

Summary of Effectiveness:

The 'Echo-Screen T' models are portable OAE/ABR based screening systems. The aim is to detect and track hearing loss and abnormal middle ear conditions as early as possible in the person's life. Due to the automated test it easy to be used by trained personal (nurses,..) in a medical or school environment.



JUN 14 2002

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. Michael Sorensen  
President  
Fischer-Zoth Audiologic Systems, Inc.  
1257 W. Margaret View Circle  
Riverton, UT 84065-4017

Re: K013977

Trade/Device Name: Echo-Screen T, TA, TD, TDA, TC  
Regulation Number: 882.1900  
Regulation Name: Evoked response auditory stimulator  
Regulatory Class: II  
Product Code: GWJ  
Dated: March 14, 2002  
Received: March 19, 2002

Dear Mr. Sorensen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Michael Sorensen

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten". The signature is fluid and cursive, with a large initial "C" and "W".

Celia M. Witten, Ph.D., M.D.  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**K013977**

**ECHO-SCREEN T, TA, TD, TDA, TC**

**INDICATIONS FOR USE:**

The Fischer-Zoth Audiologic Systems model family "Echo-Screen T" consisting of Echo-screen T, Echo-Screen TA, Echo-Screen TD, Echo-Screen TDA and Echo-Screen TC are based upon Otoacoustic Emissions (OAE) and Auditory Brainstem Response (ABR) technology.

The device is intended to screen hearing for newborns through adults, including geriatric patients. The device does not measure hearing per se, but helps to determine whether or not a hearing loss may be present.

The "Echo-Screen T" product family consists of handheld, automated OAE and ABR based hearing screening systems which are easy to use. The measurement flow is menu guided and the evaluation is based upon signal statistics. The "Echo-Screen T" devices are intended to be used by trained personnel in a medical or school environment. The "Echo-Screen T" models are not intended for fitting assistive listening devices such as hearing aids or cochlear implants.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use  X   
(per 21 CFR 801.109)

OR

Over-The Counter Use \_\_\_\_\_

(Optional Format 1-2-96)

  
**(Division Sign-Off)**  
**Division of General, Restorative  
and Neurological Devices**

510(k) Number  K013977