

**Ethical Hearing Instruments**  
90 Fleet Street Portsmouth, NH 03801  
603-430-1024 603-436-3220 (fax)

AUG 21 1996

K962012

**Safety and Effectiveness**  
**DRAERD 510 (k) Summary/Statement Certification**  
**for Wizard Loudness Growth Tester**

The summary of information, new product to present products demonstrates substantial equivalence as required by the "Safe Medical Devices Act (SMIDA)" of 1990. The design follows traditional principles of operation for Loudness Growth Testers. The device is manufactured and assembled with equivalent manufacturing practices and performance guidelines. Performance of this instrument meets requirements specified by ANSI S3.6-1989. User controls and methods are similar to others currently being marketed.