

K962218

CARRINGTON

SEP 12 1998

Device Name:

Classification Name: Dressings, wound and burns, hydrogel

Common Name: Hydrogel Wound Dressings

Proprietary Names: Carrasyn® Hydrogel Wound Dressing
CarraGauze® Pads
CarraGauze® Strips
CarraSorb M™ Freeze-Dried Gel

Device Sponsor: Carrington Laboratories, Inc.
2001 Walnut Hill Lane
Irving, TX 75038-4404
Registration No: 162544

Regulatory Classification: Unclassified. (Product code MGQ)

Summary of Safety and Effectiveness:

This 510 (k) Premarket Notification provides for the standardization and clarification of previously cleared indications for the labeling of Carrasyn® wound dressings. Carrasyn® Hydrogel Wound Dressing was originally marketed pursuant to 510 (k) Number K902345. The additional wound dressings that are included in this labeling standardization and clarification include CarraGauze® pads (K915001), CarraGauze® strips (K925002) and CarraSorb M™ Freeze-Dried Gel (K944427).

Carrasyn® wound dressings are either smooth, nonoily clear hydrogels or freeze-dried preparations of the same. They are supplied in either a liquid or dry state and are designed to be used in conjunction with an appropriate cover dressing and are intended for the management of wounds.

Since the products are for topical use, the safety of these wound gels was established through standard biocompatibility tests. These biocompatibility tests were conducted in standard animal models with preparations of the Carrasyn® wound dressings. These tests included Primary Dermal Irritation testing and Primary Eye Irritation testing. These studies demonstrated that Carrasyn® wound dressings are not primary dermal irritants or primary eye irritants. These studies were conducted in conformance with Good Laboratory Practices (GLP) regulations of the United States Food and Drug Administration (21 CFR, Part 58).
Carrington Laboratories, Inc. • Post Office Box 168128 • Irving, TX 75016-8128 • (800) 527-5216

Corporate Office & Manufacturing:
2001 Walnut Hill Lane
Irving, Texas 75038
(214) 518-1300
(214) 518-1020 Fax

Research & Development:
1300 E. Rochelle Boulevard
Irving, Texas 75062
(214) 717-5009
(214) 717-0997 Fax

Distribution Center & Accounting:
1909 Hereford Drive
Irving, Texas 75038
(214) 518-1300
(214) 756-0108 Fax

Costa Rica
Suite SJO-2632 - Unit C-101
1601 N.W. 97th Avenue
Miami, Florida 33102-5216
011-506-666-0100

00005

Clinical experience to date has been gained during the course of two studies that involved four patients with radiation dermatitis and 30 with diabetic ulcers. Grades III or IV diabetic ulcers of the foot were treated with Carrasyn® Hydrogel Wound Dressing for 10 weeks. Acute and subacute radiation dermatitis was treated for various times with Carrasyn® Hydrogel Wound Dressing. These studies evaluated the acceptability of Carrasyn® Hydrogel Wound Dressing to both the patients and clinicians, to wound and skin appearance, and to the wound healing environment. These studies concluded that Carrasyn® wound dressings are safe and effective for their intended use.