

K963095

NOV 4 1996

SECTION 2 - SUMMARY AND CERTIFICATION

510(k) SUMMARY

Monaghan Medical Corporation
5 Latour Ave., Suite 1600
P.O. Box 2805
Plattsburgh, NY 12901-0299
Phone: 518-561-7330 Fax: 518-561-5660
Contact: James A. Cochie
QA/RA Manager

Date Prepared: July 23, 1996

Device Proprietary Name: TruPeak Peak Flow Meter

Device Classification name: Peak Flow Meter for
Spirometry, 73 BZH

Legally Marketed Device Identity: TruPeak Peak
Flow Meter, 510(k) #955234

Device Description:

The TruPeak Peak Flow Meter is a simple mechanical device that responds to and indicates the Peak Expiratory Flow Rate (PEFR) generated by the user during a forced exhalation maneuver.

The inlet of the device is placed into the user's mouth after which the maximum rate at exhalation is attempted from nearly full lungs. As this forced exhalation commences an internal piston extends a spring as reaction to an internal pressure spike. An indicator is pushed by the moving piston and remains at the maximum displacement position of the piston after the maneuver is terminated. This position corresponds to the PEFR and is quantified in liters per minute by reading a scale next to the indicator. The piston displacement function is logarithmic thereby providing greater resolution of scale at lower PEFR readings.

The TruPeak Peak Flow Meter is identical in form and function to the TruPeak Peak Flow Meter as described in its 510(k) #955234 except for changes in component dimensions necessary for the performance of the device to conform to the November 1994 ATS Standards update.