

K963953

IMIG-MRI 510(k) Summary

DEC 18 1996

*Submitter:* IMIG-MRI Systems I.I.c., 300 Vesper Executive Park, Tyngsborough MA 01879

*Company contact:* Greg Hurst, Ph.D., Manager of Imaging Science and Applications  
(508) 649-8590; FAX (508) 649-8520

*Regulatory Counsel:* Jonathon S. Kahan, Attorney at Law  
Hogan & Hartson  
555 Thirteenth Street N.W., Washington D.C. 20004-1109  
(202) 637-5794; FAX (202) 637-5910

*Date of Summary Preparation:* 18 September 1996

*Device Name:* Magnetic Resonance Diagnostic Device.

*Classification Number:* 90LNH

*Device Class:* Class II, under 21 CFR 892.1000

*Device Trade Name:* IMIG-MRI™

*Intended Uses:* Anatomy: head, limbs, spine, torso; Nuclei: H-1; Diagnostic uses: Imaging

*Predicate Device:* Siemens Magnetom P8

*Safety parameter data summary:*

Maximum static magnetic field: 0.15 Tesla  
Maximum rate of magnetic field change: 18.4 Tesla/sec  
Maximum RF power deposition: .05 W/kg  
Acoustic noise levels: 114 dB peak; 95 dB A-weighted RMS

*Performance test data summary:*

(Receive coil:)	(head)	(body)
Specification volume:	15 cm dsv	30 cm dsv
S/N:	> 32	> 31
Uniformity:	< 15 %	< 40 %
Geometric distortion:	< 2.5 %	< 3.5 %
Slice thickness :	within 10% of nominally designated value	
Slice position (gap):	within 10% of nominally designated value	
Spatial resolution:	nominally equivalent to pixel size	

General safety and effectiveness concerns: Safe and effective use of the machine is assured by associated labeling. This labeling includes: advertising brochures, Site Planning Guide, and Instructions for Use (comprised of Clinical Users Guide, User Safety Guide, User Training Guide, User Applications Guide, and User QA & Maintenance Guide)

Substantial equivalence: This device has the same intended use and basic technological characteristics as the Siemens Magnetom P8.